

**DEPARTMENT OF RADIODIAGNOSIS AND IMAGING
INSTITUTE OF MEDICAL SCIENCES
BANARAS HINDU UNIVERSITY**

Corrigendum – 05

This is in reference with the tender with the following details:

Tender Reference Number **BHU/RD&I/2018-19/014**

Tender I.D **2018_BHU_330419_1**

Titled **Supply of 3 T MRI equipment on turn-key basis in
Deptt. of Radiodiagnosis & Imaging, Institute of Medical Sciences,
BHU, Varanasi**

Published on Date **24.04.2018**

A modification in the technical bid along with certain corrections in turnkey clause as appended. Please note that the corrigendum has the acceptance of all our technical experts (official E.Mails of external experts are also appended).

The amendment in technical bid is as follows:

Ack
27/8/18

Head of Department
विभागाध्यक्ष
Radio Diagnosis & Imaging
रेडियोलॉजी एवं चित्रण विभाग
Institute of Medical Sciences
भारतीय चिकित्सा विद्यापीठ
Banaras Hindu University
काशी हिन्दू विश्वविद्यालय

Corrigendum for Advertisement No: BHU/RD&I/2018-19/014

BID DOCUMENT

(e – Procurement - CORRIGENDUM)

Supply of 3 T MRI equipment on turn-key basis
in Deptt. of Radiodiagnosis & Imaging
Institute of Medical Sciences
BHU, Varanasi



DEPTT. OF RADIODIAGNOSIS & IMAGING
INSTITUTE OF MEDICAL SCIENCES
BANARAS HINDU UNIVERSITY
VARANASI-221005, INDIA

This document is subject to copyright.

This document may be used and reproduced for non-commercial purposes only. Any commercial use, including reselling, charging to access, redistribution, or for derivative works such as unofficial translations based on these documents is not allowed.

Page 2 of 4

Neeraj Sharma
27-8-18

Ashish Verma
27-8-18
Ashish
27/8/18

CORRIGENDUM

Clause no.	Tender specification	Amendment
5 (e)	Body phased array coils 32 channels of more (single or in combination) for a coverage of 45cm without repositioning.	Body phased array coils 32 channels of more (single or in combination) for a coverage of 45cm without repositioning. In case a vendor does not have a 32 channel coil then a 28 channel coil may be considered with an undertaking certifying that they do not have a 32 channel coil.
6	Table/Table top should be detachable type for emergency evacuation and incorporating future advanced applications like MR Guided Focused US /RT Planning.	Table/Table top should be fixed (emergency evacuation facility should be provided) or detachable type
10 Imaging Sequences	Added	Please do quote all available software for silent MR capability to be quoted with maximum sound level in all sequences not more than 85 decibel. Please provide necessary details.
Corrigendum 2, Page 11 SOW Under turnkey works:	A total of approximately 4000 sq. feet area (split in two places, as discussed during site visits of all interested vendors) shall be included in the turn key work. This would include free of cost guideline based dis-installation and proper disposal of an old 0.2 tesla equipment occupying the provided area as the first step and an old Hitachi head scanner (the responsibility to find out any guidelines if available for this purpose and to certify that the dis-installment/disposal has been done according to same would be on the vendor. The certificate to the effect should be provided to the university after dis-installation /disposal. In	A total of approximately 3045 Sq Ft. (split in two places, as discussed during site visits of all interested vendors – i.e approx. 2012 Sq Ft area needs modification of an already constructed area and 1033 Sq Ft requiring new construction) shall be included in the turn key work. This would include free of cost guideline based dis-installation and proper disposal of an old 0.2 tesla equipment occupying the provided area as the first step and an old Hitachi head scanner (the responsibility to find out any guidelines if available for this purpose and to certify that the dis-installment/disposal has been done according to same would be on the vendor. The certificate to the effect

Weerisham
27-8-18

Ashish Kumar
27.8.19

Ashish Kumar
27/8/19

	<p>case the equipment may be sold as a junk, the university reserves the right to secure the same in the university junkyard). The vendors are free to inspect and seek for required details of this equipment.</p>	<p>should be provided to the university after dis-installation /disposal. In case the equipment may be sold as a junk, the university reserves the right to secure the same in the university junkyard). The vendors are free to inspect and seek for required details of this equipment.</p> <p>Seeking for the permissions for approval of the Structural drawings/sanctions/permission from local authorities/ regulating bodies will be the responsibility of Institute/consignee. The standard guidelines for all such aspects (i.e drawings, planning, execution etc.) should be followed by the vendor and adequate time given to the university for each such permission.</p>
<p>Corrigendum 2, Page 14</p>	<p>Furniture:- Furniture of good quality of reputed make should be provided. Office table with side: 1 no. (customized) Computer/Office chairs: 4 no. Godrej/Feather lite or equivalent make, Storage counter: 2Nos. (size 2000x450x600 mm) customized in console room.</p>	<p>Furniture:- Furniture of good quality of reputed make should be provided. Office table with side: 1 no. (customized) Computer/Office chairs: 4 no. Godrej/Feather lite or equivalent make, Storage counter: 2Nos. (size 2000x 450x600 mm) customized in console room. Patient waiting chairs (in set of 3) - 15 No. Phantom Rack: 1 no. (customized)</p>

Weerjsharma
22-8-18

Ashish Kumar
27.8.18

Ashish
27/8/18
(Tender Inviting Authority)
Head of Department
रिडिआयग्राफी व इमेजिंग
रेडिओलॉजी व इमेजिंग
Instt. of Medical Sciences
चिकित्सा विज्ञान संस्थान
Banaras Hindu University
काशी हिन्दू विश्वविद्यालय