Dated: 11-02-2021

# **BID DOCUMENT** (e - Procurement)

Project Supply of Laboratory equipments in Department of Neurology, Institute of Medical Sciences, BHU, Varanasi



Issued On: 17-02-2021

DEPARTMENT OF NEUROLOGY, INSTITUTE OF MEDICAL SCIENCES BANARAS HINDU UNIVERSITY VARANASI-221005, INDIA This document is subject to copyright.

This document may be used and reproduced for non-commercial purposes only. Any commercial use, including reselling, charging to access, redistribution, or for derivative works such as unofficial translations based on these documents is not allowed.

#### **Checklist for Bid/Tender Submission**

(The following check-list must be filled in and submitted with the bid documents)

#### Pre- Qualification Bid

Sr. No.	Particulars	Yes/No
1	Have you attached the techno commercial unpriced bid form duly filled in appropriately?	
2	Have you attached a copy of the last audited balance sheet of your firm	
3	Have you attached the details of the income tax clearance certificate, proof of manufacturing unit/ dealership letter/ general order suppliers and copy of Central / State sales tax registration certificate?	
4	Have you attached the copies of relevant work orders from Govt. Depts. / PSUs and Central Autonomous Bodies?	
5	EMD: Have you submitted EMD asked for (as specified in BDS).	
6	Have you submitted samples of all items indicated in the respective schedule of requirements at the address of tender inviting authority within due date.	
7	Have you enclosed the schedule of requirement indicating the make offered without indicating the pricing components along with the techno commercial unpriced bid?	
8	Have you submitted the bids both techno commercial unpriced and priced bid separately for each tender?	
9	Have you enclosed the statement of deviations from financial terms and conditions, if any?	
Price	Bid	
1	Have you signed and attached the priced bid form?	
2	Have you attached the schedule of requirements duly priced?	

#### **Table of Contents**

Part 1: Bide	ding Procedures	5
Section I.	Instructions for Online Bid Submission	6
Section II.	Instructions to Bidders	9
Section III.	Bid Data Sheet (BDS)	27
Section IV.	Prequalification	
Section V.	Policy of University against the Corrupt and Fraudulent Practices	32
Part 2: Sup	ply Requirements	34
Section VI.	Schedule of Requirements	35
	(1) List of Goods and Delivery Schedule	
	(2) List of Related Services and Completion Schedule	
	(3) Technical Specifications	
Part 3: Con	ıtract	49
Section VII.	General Conditions of Contract	50
Section VIII.	Special Conditions of Contract	66
Part 4: Bide	ding Forms & Contract Forms	70
Section IX.	Bidding Forms	71
Section X.	Contract Forms	79
NIT		

## Part-1 Bidding Procedures

#### **Section I: Instructions for Online Bid Submission**

Instructions to the Bidders to submit the bids online through the Central Public Procurement Portal for e Procurement at <u>https://eprocure.gov.in/eprocure/app</u>.

- 1. Possession of valid Digital Signature Certificate (DSC) and enrollment/registration of the contractors/bidders on the e-Procurement/e-tender portal are prerequisite for e-tendering.
- 2. Bidder should register for the enrollment in the e-Procurement site using the "<u>Online Bidder Enrollment</u>" option available on the home page. Portal enrollment is generally free of charge. During enrollment/registration, the bidders should provide only valid and true information including valid email id. All the correspondence shall be made directly with the contractors/bidders through email id as registered.
- 3. Bidder need to login to the site through their user ID/ password chosen during enrollment/registration.
- 4. Then the Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by SIFY/TCS/nCode/eMudra or any other Certifying Authority recognized by Controller of Certifying Authorities (CCA) India on eToken/SmartCard, should be registered.
- 5. The registered DSC only should be used by the bidder in the transactions and should ensure safety of the same.
- 6. Contractor/Bidder may go through the tenders published on the site and download the tender documents/schedules for the tenders.
- 7. After downloading / getting the tender document/schedules, the Bidder should go through them carefully and then submit the documents as required, otherwise bid will be rejected.
- 8. Any clarifications may be sought online through the tender site, through the contact details or during pre-bid meeting if any. Bidder should take into account the corrigendum if any published before submitting the bids online.
- 9. Bidder may log in to the site through the secured login by the user id/ password chosen during enrolment/registration and then by submitting the password of the e-Token/Smartcard to access DSC.
- 10. Bidder may select the tender in which he/she is interested in by using the search option and then move it to the 'my tenders' folder.
- 11. From my tender folder, he may select the tender to view all the details uploaded there.

- 12. It shall be deemed that the bidder has read and understood all the terms and conditions before submitting the offer. Bidder should go through the tender schedules carefully and upload the documents as asked; otherwise, the incomplete bid shall stand rejected.
- 13. Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document/schedule and ordinarily it shall be in PDF/xls/rar/jpg/dwf formats. If there is more than one document, all may be clubbed together and provided in the requested format. Bidders Bid documents may be scanned with 100 dpi with black and white option. It is advisable that each document to be uploaded through online for the tenders should be less than 2 MB. If any document is more than 2MB, it can be reduced through zip/rar and the same if permitted may be uploaded. The file size being less than 1 MB the transaction uploading time will be very fast.
- 14. The Bidders can update well in advance, the documents such as certificates, annual report details etc., under "My Space option" and these can be selected as per tender requirements and then send along with bid documents during bid submission. This will facilitate the bid submission process faster by reducing upload time of bids.
- 15. Bidder should submit the Tender Fee/ EMD as specified in the tender. The hard copy should be posted/couriered/given in person to the Tender Inviting Authority, within bid submission due date and time as indicated in the tender. Scanned copy of the instrument should be uploaded as part of the offer.
- 16. While submitting the bids online, the bidder shall read the terms and conditions and may accept the same to proceed further to submit the bid packets.
- 17. The bidder has to select the payment option as offline to pay the Tender FEE/ EMD as applicable and enter details of the instruments.
- 18. The details of the DD/any other accepted instrument, physically delivered, should tally with the details available in the scanned copy and the data entered during bid submission time, otherwise submitted bid shall not be acceptable or liable for rejection.
- 19. The bidder has to digitally sign and upload the required bid documents one by one as indicated. Very act of using DSC for downloading the bids and uploading their offers shall be deemed to be a confirmation that they have read, understood and agreed with all clauses of the bid document including General conditions of contract without any exception.
- 20. The bidder has to upload the relevant files required as indicated in the cover content. In case of any irrelevant files, the bid may be rejected.
- 21. If the price bid format is provided in a spread sheet file like BoQ\_XXXX.xls, the rates offered should be entered in the allotted space only and uploaded after filling the relevant columns. The Priced-bid/BOQ template shall not be modified / replaced by the bidder; else the bid submitted is liable to be rejected for the tender.

- 22. The bidders are advised to submit the bids through online e-tendering system to the Tender Inviting Authority (TIA) well before the bid submission due date and time (as per Server System Clock). The TIA shall not be held responsible for any delay or the difficulties faced during the submission of bids online by the bidders.
- 23. After the bid submission (i.e. after Clicking "Freeze Bid Submission" in the portal), the acknowledgement number indicated by the system should be printed by the bidder and kept as a record of evidence for online submission of bid for the particular tender and also be used as entry pass to participate in the bid opening.
- 24. The time settings fixed in the server side and displayed at the top of the tender site, shall remain valid for all actions of requesting, bid submission, bid opening etc., in the e-Tender system. The bidders should follow such time during bid submission.
- 25. All the data being entered by the bidders would be encrypted using Public Key Infrastructure (PKI) encryption techniques to ensure the secrecy of the data. The data entered is not retrievable by unauthorized persons during the bid submission and until the time of bid opening by any person.
- 26. Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key. Further this key is subjected to asymmetric encryption using buyers/bid openers' public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 27. The confidentiality of the bids is maintained with the use of Secured Socket Layer (SSL) 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 28. The bidder should logout of the tendering system using the normal logout option available at the top right hand corner and not by selecting the (X) exit option in the browser.
- 29. For any queries regarding e-Tendering process, the bidders may contact at address as provided in the tender document. Parallely for any further queries, the bidders are advised to contact over phone: **1-800-233-7315** or send an e-mail to <u>cppp-nic@nic.in</u>.

\*\*\*\*\*

#### Section II: Instructions to Bidders

#### **Table of Clauses**

<b>A.</b>	General	
1. 2.	Scope of Bid Corrupt and Fraudulent Practices	
2. 3.	Eligible Bidders	
5.		10
B.	Contents of Bidding Document	12
<b>в.</b> 4.	Sections of Bidding Document	
4. 5.	Clarification of Bidding Documents, Site Visit, Pre-Bid Meeting	
	-	
6.	Amendment of Bidding Document	13
C.	Preparation of Bids	14
7.	Cost of Bidding	14
8.	Language of Bid	14
9.	Documents Comprising the Bid	
10.	Tender Forms(Technical and Price) and Price Schedule(BOQ)	
11.	Alternative Bids	
12.	Bid Prices and Discounts	
13.	Currencies of Bid and Payment	
	Documents Establishing the Eligibility and Qualifications of the Bidder	
	Period of Validity of Bids	
	Bid Security	
10.	bla becanty	1)
D.	Submission and Opening of Bids	20
17.	Sealing and Marking of Bids	
	Deadline for Submission of Bids	
	Late Bids	
	Withdrawal, Substitution, and Modification of Bids	
	Bid Opening	
E.	Evaluation and Comparison of Bids	
	Confidentiality	
	Clarification of Bids	
	Determination of Responsiveness	
	Conversion to Single Currency	
	Margin of Preference	
	Evaluation of Bids	
	Comparison of Bids	
	Qualification of the Bidder	
30.	Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids	24
F.	Award of Contract	24
31.	Award Criteria	
	Purchaser's Right to Vary Quantities at Time of Award	
	Notification of Award	
	Signing of Contract	
	Performance Security	
		-0

#### Section II. Instructions to Bidders

#### A. General

1.	Scope of Bid	1.1	Scope of Work :
			a. Supply of Laboratory Equipments as per Specifications.
		1.2	Throughout these Bidding Documents unless the context otherwise requires:
			a. "in writing" means communicated in written form (e.g. by mail, e-mail, fax, telex) with proof of receipt;
			b. "singular" means "plural" and vice versa; and
			c. "day" means calendar day.
2.	Corrupt and Fraudulent	2.1	The Purchaser requires compliance with its policy in regard to corrupt and fraudulent practices as set forth in Section V.
	Practices	2.2	Further in pursuance of this policy, Bidder shall permit and cause its agents (whether declared or not), sub-contractors, sub-consultants, service providers or suppliers to provide access to purchaser to all the accounts, records and other documents relating to submission of the applicant, bid submission (in case prequalified), and contract performance (in case of award), to inspect and to have them audited by auditors appointed by the purchaser.
3.	Eligible Bidders	3.1	A Bidder may be a firm, a company, a limited liability partnership (LLP), a government-owned entity or any combination of such entities in the form of a joint venture (JV) under an existing agreement or with the intent to enter into such an agreement supported by a letter of intent.
		3.2	In the case of a joint venture, all members shall be jointly and severally liable for the execution of the contract in accordance with the Contract terms. The JV shall nominate a Representative who shall have the authority to conduct all business for and on behalf of any and all the members of the JV during the bidding process and during the contract execution in the event the JV is awarded the contract.
		3.3	A Bidder shall not have a conflict of interest. Any Bidder found to have a conflict of interest shall be disqualified. A Bidder may be considered to have a conflict of interest for the

a. directly or indirectly controls, is controlled by or is under

purpose of this bidding process, if the Bidder:

common control with another Bidder; or

- b. receives or has received any direct or indirect subsidy from another Bidder; or
- c. has the same legal representative as another Bidder; or
- d. has a relationship with another Bidder, directly or through common third parties, that puts it in a position to influence the bid of another Bidder, or influence the decisions of the Purchaser regarding this bidding process; or
- e. participates in more than one bid in this bidding process. Participation by a Bidder in more than one Bid shall result in the disqualification of all Bids in which such Bidder is involved. This, however does not limit the inclusion of the same subcontractor in more than one bid; or
- f. any of its affiliates participated as a consultant in the preparation of the design or technical specifications of the works that are the subject of the bid; or
- g. any of its affiliates has been hired (or is proposed to be hired) by the Purchaser for the contract implementation; or
- h. would be providing goods, works, or non-consulting services resulting from or directly related to consulting services for the preparation or implementation of the project that it provided or was provided by any affiliate that directly or indirectly controls, is controlled by, or is under common control with that firm; or
- i. has a close business or family relationship with a professional staff of the Purchaser (or of the project implementing agency, or of a recipient of a part of the loan) who:
  - (i) are directly or indirectly involved in the preparation of the bidding documents or specifications of the contract, and/or the bid evaluation process of such contract; or
  - (ii) would be involved in the implementation or supervision of such contract.

3.4 A foreign firm and individual may be ineligible if as a matter of law or regulations, India prohibits commercial relations with the country of bidder.

3.5 A Bidder shall provide such evidence of eligibility satisfactory to the Purchaser, as the Purchaser shall reasonably request.

#### **B.** Contents of Bidding Document

 4 Sections of Bidding Document
 4.1 The Bidding Documents consist of Parts 1, 2, 3 and 4, which include all the Sections indicated below, and should be read in conjunction with any Addenda if any, issued.

#### **Part 1: Bidding Procedures**

Section I.	Instructions for Online Bid Submission
Section II.	Instructions to Bidders
Section III.	Bid Data Sheet (BDS)
Section IV.	Prequalification
Section V.	Policy of University against the Corrupt and fraudulent Practices

#### Part 2: Supply Requirements

Section VI. Schedule of Requirements

#### Part 3: Contract

Section VII.	General Conditions of Contract
Section VIII.	Special Conditions of Contract

#### Part 4: Bidding and Contract Forms

Section IX.	Bidding Forms
Section X.	Contract Forms

- 4.2 The Invitation for Bids issued by the Purchaser is not part of the Bidding Document.
- 4.3 Unless obtained directly from the Purchaser, the Purchaser is not responsible for the completeness of the document, responses to requests for clarification, the Minutes of the pre-Bid meeting (if any), or Addenda to the Bidding Document. In case of any contradiction, documents obtained directly from the Purchaser shall prevail.
- 4.4 The Bidder is expected to examine all instructions, forms, terms, and specifications in the Bidding Documents and to furnish with its Bid all information or documentation as required by the Bidding Documents.
- **5** Clarification of 5.1 A Bidder requiring any clarification of the Bidding Document

Bidding Documents, Site Visit, Pre-Bid Meeting shall contact the Purchaser in writing at the Purchaser's address specified in the BDS or raise its enquiries during the pre-bid meeting if provided. The Purchaser will respond in writing to any request for clarification, provided that such request is received prior to the deadline for submission of bids within a period specified in the BDS. The Purchaser shall forward copies of its response to all Bidders who have acquired the Bidding Documents, including a description of the inquiry but without identifying its source. If so specified in the BDS, the Purchaser shall also promptly publish its response at the web page identified in the BDS. Should the clarification results in changes to the essential elements of the Bidding Documents, the Purchaser shall amend the Bidding Documents following the due procedure.

- 5.2 If so specified in the BDS, the Bidder is advised to visit and examine the project site and obtain for itself on its own responsibility all information that may be necessary for preparing the bid and entering into a contract for procurement of Goods. The costs of visiting shall be at the Bidder's own expense.
- 5.3 The Bidder and any of its personnel or agents shall be granted permission by the Purchaser to enter upon its premises and lands upon the express condition that the Bidder, its personnel, and agents shall indemnify the Purchaser against all liability in respect thereof.
- 5.4 If so specified in the BDS, the Bidder's designated representative is invited to attend a pre-bid meeting. The purpose of the meeting will be to clarify issues and to answer questions on any matter that may be raised at that stage.
- 5.5 The Bidder is advised to submit any questions in writing to reach the Purchaser not beyond one week preceding the meeting.
- 5.6 Minutes of the pre-bid meeting, if applicable, including the text of the questions asked by Bidders, without identifying the source, and the responses given, together with any responses prepared after the meeting, will be transmitted promptly to all Bidders who have acquired the Bidding Documents. Any modification to the Bidding Documents that may become necessary as a result of the pre-bid meeting shall be made by the Purchaser exclusively through the issue of an addendum and not through the minutes of the pre-bid meeting. Absence in the pre-bid meeting shall not be a cause for disqualification of a Bidder.
- 6 Amendment of 6.1 At any time prior to the deadline for submission of bids, the

- **Bidding Document** Purchaser may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder, modify the bidding documents by corrigendum. In case of e-procurement, corrigendum / amendment shall be published on <u>http://eprocure.gov.in/eprocure/app</u>.
  - 6.2 Any addendum issued shall be part of the Bidding Documents and shall be communicated in writing to all who have obtained the Bidding Documents from the Purchaser. The Purchaser shall also promptly publish the addendum on the Purchaser's web page.
  - 6.3 The Purchaser may, at its discretion to give prospective Bidders reasonable time in which to take an addendum into account in preparing their bids, extend the deadline for the submission of bids.

#### **C. Preparation of Bids**

- 7 Cost of Bidding 7.1 The Bidder shall bear all costs associated with the preparation and submission of its bid, and the Purchaser shall not be responsible or liable for those costs, regardless of the conduct or outcome of the bidding process.
- 8 Language of Bid
  8.1 The Bid, as well as all correspondence and documents relating to the bid exchanged by the Bidder and the Purchaser, shall be written in the language specified in the BDS. Supporting documents and printed literature that are part of the Bid may be in another language provided they are accompanied by an accurate translation of the relevant passages into the language specified in the BDS, in which case, for purposes of interpretation of the Bid, such translation shall govern.
  - 9.1 The tender/Bid shall be submitted online in two part, viz., Technical Bid and Commercial Bid.

#### 9.1.1 TECHNICAL BID

The following documents are to be scanned and uploaded as part of the Technical Bid as per the tender document:

- (a) Scanned copy of Tender Forms (Techno Commercial Un-Priced Bid) and Tender Acceptance Letter);
- (b) Scanned copy of the completed schedules,

9 Documents Comprising the Bid

- (c) Scanned copy of Bid Security or copy of proof for submission of Tender Document Fee/ Earnest Money Deposit etc.;
- (d) Scanned copy of written confirmation authorizing the signatory of the Bid to commit the Bidder;
- (e) Scanned copy of documentary evidence
  - (i) establishing the Bidder's qualifications to perform the contract if its bid is accepted and
  - (ii) the Bidder's eligibility to bid;
- (f) Scanned copy of
  - (i) documentary evidence, that the Goods and Related Services to be supplied by the Bidder are of eligible origin and
  - (ii) conform to the Bidding Documents, and
  - (iii) any other document required in the BDS;
- (g) Scanned copy of Pre-qualification Details as per Section-IV like PAN/TIN/Sales Tax / Service Tax etc.
- (h) Technical Bid.

All the original documents as well as the original payment instrument like Demand Draft/Bank Guarantee /Pay order or banker cheque of any scheduled bank against Tender Fee/EMD, samples as specified in this tender document have to be sent to the address of the Purchaser mentioned in Bid Data Sheet (BDS) by post/speed post/courier/by hand on or before bid Submission closing date & time. Beyond that the tender shall be summarily rejected without assigning any reason.

#### 9.1.2 COMMERCIAL BID

The commercial bid comprises of :

- (i) Scanned copy of Tender Form (Price Bid)
- (ii) Price bid in the form of BoQ\_XXXX.xls.
- (iii) Scanned copy of item wise breakup of price bid.

The Price bid format is provided as BoQ\_XXXX.xls along with this Tender Document at http://eprocure.gov.in/eprocure/app. Bidders are advised to download this BoQ\_XXXX.xls and quote their offer/rates in the prescribed column. Bidders can quote Basic Price in INR or CURRENCY (for other than INR) but it is mandatory to quote taxes/levies in INR only, in the prescribed column and upload the same in the commercial bid.

- 9.2 In addition to the above requirements, bids submitted by a JV shall include a copy of the Joint Venture Agreement entered into by all members. Alternatively, a letter of intent to execute a Joint Venture Agreement in the event of a successful bid shall be signed by all members and submitted with the bid, together with a copy of the proposed Agreement, there to.
- 9.3 The Bidder shall furnish in the Tender Forms information on commissions and gratuities, if any, paid or to be paid to agents or any other party relating to this Bid.
- 10 Tender Forms(Technical and Price) and Price
   10.1 Tender Forms and Price Schedules (Bill of Quantity-BOQ) shall be prepared using the relevant forms furnished in Section IX, Bidding Forms and BOQ provided. The forms must be completed without any alterations to the text, and no substitutes shall be accepted. All blank spaces shall be filled in with the information requested.
  - **Bids** 11.1 Unless otherwise specified in the BDS, alternative bids shall not be considered.
    - 1 12.1 The prices and discounts quoted by the Bidder in the Tender Forms and in the Price Schedules (BOQ) shall conform to the requirements specified as under.
      - (a) All lots (contracts) and items must be listed and priced separately in the Price Schedules (BOQ).
      - (b) The price to be quoted in the Tender Forms shall be the total price of the bid, excluding any discounts offered.
      - (c) The Bidder shall quote any discount and indicate the methodology for their application in the Tender Forms.
      - (d) Prices quoted by the Bidder shall be fixed during the Bidder's performance of the Contract and not subject to variation on any account, unless otherwise specified in the BDS A bid submitted with an adjustable price quotation shall be treated as nonresponsive and shall be rejected. However, if in accordance with the BDS, prices quoted by the Bidder shall be subject to adjustment during the performance of the Contract, a bid submitted with a fixed price quotation shall not be rejected, but the price adjustment shall be treated as zero.
      - 12.2 If so bids are being invited for individual lots (contracts) or for any combination of lots (packages). Unless otherwise specified in the BDS, prices quoted shall correspond to 100 % of the items specified for each lot and to 100% of the

- 11 Alternative Bids
- 12 Bid Prices and Discounts

quantities specified for each item of a lot. Bidders wishing to offer discounts for the award of more than one Contract shall specify in their bid the price reductions applicable to each package, or alternatively, to individual Contracts within the package. Discounts shall be submitted provided the bids for all lots (contracts) are opened at the same time.

- 12.3 Prices shall be quoted as specified in each Price Schedule (BOQ) as provided. The dis-aggregation of price components is required solely for the purpose of facilitating the comparison of bids by the Purchaser. This shall not in any way limit the Purchaser's right to contract on any of the terms offered. In quoting prices, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible country. Prices shall be entered in the following manner:
  - (a) For Goods manufactured in India:
    - (i) the price of the Goods quoted EXW (ex-works, ex-factory, ex warehouse, ex showroom, or off-the-shelf, as applicable), including all customs duties and sales and other taxes already paid or payable on the components and raw material used in the manufacture or assembly of the Goods;
    - (ii) any sales tax/VAT and other taxes payable on the Goods, if the contract is awarded to the Bidder; and
    - (iii) the price for inland transportation, insurance, and other local services required to convey the Goods to their final destination (Project Site) as specified in the BDS.
  - (b) For Goods manufactured outside India, to be imported:
    - (i) the price of the Goods quoted under Carriage and Insurance Paid (CIP) Model upto named place of destination in India as specified in the BDS;
    - (ii) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified in the BDS;
  - (c) For Goods manufactured outside India, already imported:
    - (i) the price of the Goods, including the original import value of the Goods; plus any mark-up (or rebate); plus any other related local cost, and custom duties and other import taxes already paid or to be paid on the Goods already imported.

- (ii) the custom duties and other import taxes already paid (need to be supported with documentary evidence) or to be paid on the Goods already imported;
- (iii) the price of the Goods, obtained as the difference between (i) and (ii) above;
- (iv) any sales and other taxes which will be payable on the Goods if the contract is awarded to the Bidder; and
- (v) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified in the BDS.
- (d) for Related Services, other than inland transportation and other services required to convey the Goods to their final destination, whenever such Related Services are specified in the Schedule of Requirements:
  - (i) the price of each item comprising the Related Services (inclusive of any applicable taxes).
- The currency(ies) of the bid and the currency(ies) of 13.1 payments shall be as specified in the BDS. The Bidder shall and Payment quote in Indian Rupees the portion of the bid price that corresponds to expenditures incurred in Indian Rupees, unless otherwise specified in the BDS.
  - 14.1 To establish Bidder's their eligibility, Bidders shall complete the Tender Form (Techno Commercial Un-Priced Bid & Priced Bid), included in Section-IX, Bidding Forms.
    - 14.2 The documentary evidence of the Bidder's qualifications to perform the contract if its bid is accepted shall establish to the Purchaser's satisfaction:
      - (a) that, if required in the BDS, a Bidder that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Section IX, Bidding Forms to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in India:
      - (b) that, if required in the BDS, in case of a Bidder not doing business within India, the Bidder is or will be (if awarded the contract) represented by an Agent in the country equipped and able to carry out the Supplier's maintenance. parts-stocking repair and spare obligations prescribed in the Conditions of Contract

**13** Currencies of Bid

**14** Documents Establishing the **Eligibility and Oualifications of** the Bidder

and/or Technical Specifications; and

- (c) that, the Bidder meets each of the qualification criterion.
- 15 Period of Validity of Bids
   15.1 Bids shall remain valid for the period specified in the BDS after the bid submission deadline date prescribed by the Purchaser. A bid valid for a shorter period shall be rejected by the Purchaser as non –responsive.
  - 15.2 In exceptional circumstances, the Purchaser may solicit the Bidder's consent to an extension of the period of validity. The request and the responses thereto shall be made in writing. A Bidder may refuse the request without forfeiting its Earnest Money Deposit (EMD). A Bidder acceding to the request will neither be required nor permitted to modify the bid.
  - 15.3 Bid evaluation will be based on the bid prices without taking into consideration the above modifications.
- **16 Bid Security** 16.1 The Bidder shall furnish as part of its bid, a bid security, as specified in the BDS, in original form the amount and currency as specified in the BDS.
  - 16.2 If a bid security is specified, the bid security shall be a demand guarantee in any of the following forms at the Bidder's option:
    - (a) an unconditional guarantee issued by a bank or financial institution (such as an insurance, bonding or surety company);
    - (b) an irrevocable letter of credit;
    - (c) a banker's cheque or Demand Draft ; or
    - (d) any other security as specified in the BDS,

of a reputed source from an eligible country. If the unconditional guarantee is issued by a financial institution located outside India, the issuing financial institution shall have a correspondent financial institution located in India to make it enforceable The bid security shall be valid for forty five (45) days beyond the original validity period of the bid, or beyond the extended period.

- 16.3 If a Bid Security is specified, any bid not accompanied by a substantially responsive Bid Security, shall be rejected by the Purchaser as non-responsive.
- 16.4 If a Bid Security is specified, the Bid Security of unsuccessful Bidders shall be returned as promptly as possible upon the successful Bidder's signing the contract

and furnishing the Performance Security.

- 16.5 The Bid Security of the successful Bidder shall be returned as promptly as possible once the successful Bidder has signed the contract and furnished the required performance security.
- 16.6 The Bid Security of the bidder may be forfeited or the Bid Securing Declaration executed:
  - (a) if he withdraws from the bid during the period of bid validity specified by the Bidder on the Tender Forms, or any extension thereto provided by the Bidder ; or
  - (b) if he being successful Bidder fails to:
    - (i) sign the Contract; or
    - (ii) furnish a performance security.
- 16.7 The bid security of a JV must be in the name of the JV that submits the bid. If the JV has not been legally constituted into a legally enforceable entity at the time of bidding, the bid security shall be in the names of all members as named in the letter of intent.

#### **D.** Submission and Opening of Bids

- 17 Sealing and Marking of Bids
   17.1 The Bidder shall submit the bids electronically, through the e-procurement system (<u>http://eprocure.gov.in/eprocure/app</u>). Any document submitted through any other means will not be considered as part of the Bid except for the Originals as asked for in this tender.
- 18 Deadline for Submission of Bids
   18.1 The Purchaser may, at its discretion, extend the deadline for the submission of bids by amending the Bidding Documents, in which case all rights and obligations of the Purchaser and Bidders previously subject to the deadline shall thereafter be subject to the deadline as extended.
- 19 Late Bids
  19.1 The e-Procurement system would not allow any late submission of bids after due date and time as per server system. After electronic online proposal submission, the system generates a unique identification number which is time stamped. This shall be treated as acknowledgement of the proposal submission
- 20 Withdrawal, Substitution, and Modification of Bids
   20.1 A Bidder may withdraw, substitute, or modify its bid on the e-procurement system before the date and time specified but not beyond.

- 20.2 No bid may be withdrawn, substituted, or modified in the interval between the deadline for submission of bids and the expiration of the period of bid validity specified by the Bidder on the Tender Forms or any extension thereof Modification/Withdrawal of the Bid sent through any other means shall not be considered by the Purchaser.
- The Purchaser shall open the bids as per electronic bid 21 Bid Opening 21.1 opening procedures specified in Central Public Procurement Portal (CPPP) at the date and time specified. Bidders can also view the bid opening by logging on to the eprocurement system. Specific bid opening procedures are laid down at http://eprocure.gov.in/eprocure/app under the head "Bidders Manual Kit". The tenderer/bidder will be at liberty to be present either in person or through an authorized representative at the time of opening of the Bid or they can view the bid opening event online at their remote end. Price Bids of only those tenderers shall be opened whose technical bids qualify.
  - The withdrawn bid will be available in the system therefore 21.2 will be considered. If bidder once withdraws the bid then he will not be able to participate in the respective tender again. Modification to the bid shall be opened and read out with the corresponding bid. Only bids that are opened and read out at bid opening shall be considered further.
  - The Purchaser shall prepare a record of the bid opening that 21.3 shall include; the name of the Bidder; whether there is a withdrawal, substitution, or modification; the Bid Price including any discounts and alternative bids; and the presence or absence of a bid security, if one was required. The Bidders' representatives who are present in the office of the Purchaser to witness the bid opening shall be requested to sign the record. The omission/refusal of a Bidder's signature on the record shall not invalidate the contents and effect of the record. A copy of the record shall be made available on the e-procurement system.

#### **E.** Evaluation and Comparison of Bids

- 22.1 Information relating to the evaluation of bids and recommendation of contract award shall not be disclosed to bidders or any other persons not officially concerned with the bidding process until information on Contract Award is communication to all Bidders.
  - 22.2 No Bidder shall contact the purchaser on any matter relating to its bid from the time of the bid opening to the time the contract is awarded. If the Bidder wishes to bring additional information to the notice of the Purchaser it

22 Confidentiality

should be done in writing.

- 22.3 Any effort by a Bidder to influence the purchaser in its decisions on bid evaluation, bid comparison or contract award decisions may result in rejection of the Bidder's bid.
- 23 Clarification of 23.1 To assist in the examination, evaluation, comparison of the Bids bids, and qualification of the Bidders, the Purchaser may, at its discretion, ask any Bidder for a clarification of its Bid. Any clarification submitted by a Bidder in respect to its Bid and that is not in response to a request by the Purchaser shall not be considered. The Purchaser's request for clarification and the response shall be in writing. No change, including any voluntary increase or decrease, in the prices or substance of the Bid shall be sought, offered, or permitted, except to confirm the correction of arithmetic errors discovered by the Purchaser in the Evaluation of the bids.
  - 23.2 If a Bidder does not provide clarifications of its bid by the date and time set in the Purchaser's request for clarification, its bid may be rejected.
- 24 Determination of Responsiveness
- 24.1 The Purchaser's determination of a bid's responsiveness is to be based on the contents of the bid itself.
- 24.2 A substantially responsive Bid is one that meets the requirements of the Bidding Documents without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that:
  - (a) if accepted, would
    - (i) affect in any substantial way the scope, quality, or performance of the Goods and Related Services specified in the Contract; or
    - (ii) limit in any substantial way, inconsistent with the Bidding Documents, the Purchaser's rights or the Bidder's obligations under the Contract; or
  - (b) if rectified, would unfairly affect the competitive position of other bidders presenting substantially responsive bids.
- 24.3 The Purchaser shall examine the technical aspects of the bid submitted in accordance with instructions specified in tender document, in particular, to confirm that all requirements enumerated in the 'Schedule of Requirements' Section-VI have been complied with, without any material deviation or reservation or omission.
- 24.4 If a bid is not responsive to the requirements of Bidding

Documents, it shall be rejected by the Purchaser and may not subsequently be made responsive by correction of the material deviation, reservation or omission.

- 25 Conversion to Single Currency25.1 For evaluation and comparison purposes, the currency(ies) of the Bid shall be converted in a single currency as specified in the BDS.
- 26 Margin of<br/>Preference26.1 Unless otherwise specified in the BDS, a margin of<br/>preference shall not apply.
- **27 Evaluation of Bids** 27.1 The Purchaser shall use the criteria and methodologies listed in this Clause. No other evaluation criteria or methodologies shall be permitted.
  - 27.2 To evaluate a Bid, the Purchaser shall consider the following:
    - (a) evaluation will be done for Items or Lots (contracts), as specified in the BDS; and the Bid Price
    - (b) price adjustment due to discounts offered;
    - (c) converting the amount resulting from above, if relevant, to a single currency;
    - (d) price adjustment due to quantifiable nonmaterial nonconformities in;
  - 27.3 The estimated effect of the price adjustment provisions of the Conditions of Contract, applied over the period of execution of the Contract, shall not be taken into account in bid evaluation.
  - 27.4 The Purchaser's evaluation of a bid shall exclude and not take into account:
    - (a) in the case of Goods manufactured in the India, sales and other similar taxes, which will be payable on the goods if a contract is awarded to the Bidder;
    - (b) in the case of Goods manufactured outside India, already imported or to be imported, customs duties and other import taxes levied on the imported Good, sales and other similar taxes, which will be payable on the Goods if the contract is awarded to the Bidder;
    - (c) any allowance for price adjustment during the period of execution of the contract, if provided in the bid.
  - 27.5 The Purchaser's evaluation of a bid may require the consideration of other factors, in addition to the Bid Price quoted. These factors may be related to the characteristics, performance, and terms and conditions of purchase of the Goods and Related Services. The effect of the factors

selected, if any, shall be expressed in monetary terms to facilitate comparison of bids.

- **Comparison of** The Purchaser shall compare the evaluated prices of all 28 28.1 Bids substantially responsive bids established to determine the lowest evaluated bid. The comparison shall be on the basis of CIP-Carriage and Insurance Paid to (place of destination) prices for imported goods and EXW - Ex Works (named place of delivery) prices, plus cost of inland transportation and insurance to place of destination, for goods manufactured within India, together with prices for any required installation, training, commissioning and other services. The evaluation of prices shall not take into account custom duties and other taxes levied on imported goods quoted CIP and sales and similar taxes levied in connection with the sale or delivery of goods.
- 29 Qualification of the Bidder29.1 The Purchaser shall determine to its satisfaction whether the Bidder that is selected as having submitted the lowest evaluated and substantially responsive bid meets the qualifying criteria.
  - 29.2 The determination shall be based upon an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder.
  - 29.3 An affirmative determination shall be a prerequisite for award of the Contract to the Bidder. A negative determination shall result in disqualification of the bid, in which event the Purchaser shall proceed to the next lowest evaluated bid to make a similar determination of that Bidder's qualifications to perform satisfactorily.
- 30 Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids
   30.1 The Purchaser reserves the right to accept or reject any bid, and to annul the bidding process and reject all bids at any time prior to contract award, without thereby incurring any liability to Bidders. In case of annulment, all bids submitted and specifically, bid securities, shall be promptly returned to the Bidders.

#### F. Award of Contract

**31** Award Criteria 31.1 The Purchaser shall award the Contract to the Bidder whose bid has been determined to be the lowest evaluated bid and is substantially responsive to the Bidding Documents, provided the Bidder is determined to be qualified to perform the Contract satisfactorily.

- 32 Purchaser's Right to Vary Quantities at Time of Award
   32.1 At the time the Contract is awarded, the Purchaser reserves the right to increase or decrease the quantity of Goods and Related Services originally specified in Section VI, Schedule of Requirements, provided this does not exceed the percentages specified in the BDS, and without any change in the unit prices or other terms and conditions of the bid and the Bidding Documents.
- 33 Notification of Award
  33.1 Prior to the expiration of the period of bid validity, the Purchaser shall, notify the successful Bidder, in writing, that its Bid has been accepted. The notification letter (hereinafter and in the Conditions of Contract and Contract Forms called the "Letter of Acceptance") shall specify the sum that the Purchaser will pay the Supplier in consideration of the supply of Goods (hereinafter and in the Conditions of Contract Forms called "the Contract Price"). At the same time, the Purchaser shall also notify all other Bidders of the results of the bidding.
  - 33.2 Until a formal Contract is prepared and executed, the notification of award shall constitute a binding Contract.
  - 33.3 The Purchaser shall promptly respond in writing to any unsuccessful Bidder who, after notification of award, requests in writing the grounds on which its bid was not selected.
  - 34.1 Promptly after notification, the Purchaser shall send the successful Bidder the Contract Agreement.
  - 34.2 Within twenty-eight (28) days of receipt of the Contract Agreement, the successful Bidder shall sign, date, and return it to the Purchaser.
  - 34.3 Notwithstanding anything contained in clause 34.2, in case signing of the Contract Agreement is prevented by any export restrictions attributable to the Purchaser, or to the use of the products/goods, systems or services to be supplied, where such export restrictions arise from trade regulations from a country supplying those products/goods, systems or services, the Bidder shall not be bound by its bid, always provided however, that the Bidder can demonstrate to the satisfaction of the Purchaser that signing of the Contact Agreement has not been prevented by any lack of diligence on the part of the Bidder in completing including applying anv formalities, for permits, authorizations and licenses necessary for the export of the products/goods, systems or services under the terms of the Contract.
- 34 Signing of Contract

- 35 Performance Security
  35.1 Within twenty eight (28) days of the receipt of notification of award from the Purchaser, the successful Bidder, if required, shall furnish the Performance Security in accordance with the General Conditions of Contract (GCC), using for that purpose the Performance Security Form included in Section-X, Contract Forms, or another Form acceptable to the Purchaser. If the Performance Security furnished by the successful Bidder is in the form of a bond, it shall be issued by a bonding or insurance company that has been determined by the successful Bidder to be acceptable to the Purchaser. A foreign institution providing a bond shall have a correspondent financial institution located in India.
  - 35.2 Failure of the successful Bidder to submit the abovementioned Performance Security or sign the Contract shall constitute sufficient grounds for the annulment of the award and forfeiture of the Bid Security. In that event the Purchaser may award the Contract to the next lowest evaluated Bidder, whose bid is substantially responsive and is determined by the Purchaser to be qualified to perform the Contract satisfactorily.

#### Section III. Bid Data Sheet (BDS)

The following specific data for the goods to be procured shall complement, supplement, or amend the provisions in the Instructions to Bidders (ITB). In case of inconsistency, the provisions herein shall prevail over those in ITB.

Serial No.	A. General	
1	The reference number of the Invitation for Bids is BHU/IMS/Neuro/2020-21/01	Dated: 11-02-2021
2	The Purchaser is <b>The Head, Department of Neurology,</b> <b>Sciences, Banaras Hindu University, Varanasi</b>	Institute of Medical
3	Maximum number of members in the JV shall be: 4	

Serial No.	B. C	ontents of Bidding Documents
4		d purposes only, the Purchaser's address is <b>The Head</b> , ology, Institute of Medical Sciences, Banaras Hindu
	Attention Address	<ul> <li>Prof. Rameshwar Nath Chaurasia</li> <li>Department of Neurology Institute of Medical Sciences, BHU, Varanasi.</li> </ul>
	Floor/ Room number City ZIP Code Country Telephone	: : Varanasi : 221005 : India : 91-542-230-9373
	Facsimile number Electronic mail address	91-9415353255 : : <u>email@gmail.com</u>
5 6	Web page A site visit shall <b>not</b> be	: <u>http://eprocure.gov.in/eprocure/app</u> organized by the purchaser.
7	A Pre-Bid meeting	: NA

Serial No.	C. Preparation of Bids
8	The language of the bid is: English or Hindi.
	All correspondence exchange shall be in English or Hindi language.
	Language for translation of supporting documents and printed literature is <b>English</b> or Hindi.
9	The Bidder shall submit the following additional documents in its bid: NA
10	Alternative Bids shall not be considered.
11	The prices quoted by the Bidder <b>shall not</b> be subject to adjustment during the performance of the Contract.
12	Place of Destination: Department of Neurology, Institute of Medical Sciences, Banaras Hindu University, Varanasi
13	Final destination (Project Site): <b>Department of Neurology, Institute of Medical</b> Sciences, Banaras Hindu University, Varanasi
14	The prices shall be quoted by the bidder in : Foreign Currency of Principal's
	<b>Country (Preferably in Indian Rupees)</b> The Bidder is required to quote in Indian Rupees (INR), the portion of the bid price that corresponds to expenditures incurred in Indian Rupees (INR).
15	Manufacturer's authorization is Required
16	After sales service is Required.
17	The bid validity period shall be <b>120 days</b> .
18	EMD/Bid security shall be paid @ 2% of the estimated value(s) of quoted items or <b>Rs. 50000/-</b> whichever is higher by the way of Demand Draft (DD)/Bank Guarantee (BG) in favor of the Registrar, Banaras Hindu University, Varanasi-221005 and should be valid for a period of 45 days beyond the BID validity period. All tenders received without EMD/Bank Security shall be rejected.
	TENDER FEE : NIL
19	Other types of acceptable securities: NA

Serial No.	D. S	Submission and Opening of Bids
20	For bid submission purposes only, the Purchaser's address is The Head, Department of Neurology, Institute of Medical Sciences, Banaras Hindu University, Varanasi	
	Attention	: Prof. Rameshwar Nath Chaurasia
	Street Address	: Department of Neurology, Institute of Medical
		Sciences, Banaras Hindu University, Varanasi
	City	: Varanasi
	ZIP/Postal Code	: 221005
	Country	: India
	The deadline for bid su	bmission is:
	Date	: 20March, 2021
	Time	: <b>05:00 PM</b>
	The electronic bidding opening procedures shall be as given in Section I-Instructions for Online Bid Submission.	
21	The bid opening shall t	take place at: Institute of Medical Sciences, Faculty of
		BHU, Varanasi
	Street Address	: Banaras Hindu University, Varanasi
	Floor/ Room number	: Director Office
	City	: Varanasi
	Country	: India
	Date	: 22 March, 2021
	Time	: 03:00 PM
	The electronic bidding for Online Bid Submiss	opening procedures shall be as given in Section I-Instructions sion.

Serial No.	E. Evaluation and Comparison of Bids
22	The currency that shall be used for bid evaluation and comparison purposes to convert all bid prices expressed in various currencies into a single currency is:
	Indian Rupees
	The source of exchange rate shall be: Reserve Bank of India.
	The date for the exchange rate shall be: Last day for submission of Bids.
23	A margin of domestic preference shall apply.
24	Evaluation will be done for concern equipment.
	Note: Bids will be evaluated for each item and the Contract will comprise the item(s) awarded to the successful Bidder.

Serial No.	F. Award of Contract
25	The maximum percentage by which quantities may be increased is: NA
	The maximum percentage by which quantities may be decreased is: NA

#### Section IV. Prequalification

- **1.** A notarized affidavit by the firm that it has never been black-listed must be attached along with the Bid, failing which the Bid shall be rejected.
- Profile of each Bidder and past experience in supply of the material (certificates to be enclosed), proof of manufacturing Unit/Dealership letter and general order supplier. Manufacturer's authorization certificate as prescribed in Section IX in case bidder is not manufacturer.
- **3.** List of other Govt. Departments, Public Sector units and Central Autonomous Bodies for which the bidder is supplying material or having the similar type of contracts and a certificate regarding the satisfactory performance of the contract.
- **4.** Copy of the audited balance sheet of the bidder for the previous three financial years indicating the turnover in supply of the material.
- 5. True copy of Permanent Account Number.
- 6. Details of Sales Tax / VAT along with a copy of certificate to be attached.
- 7. Service Tax No. along with copy of certificate.
- 8. TIN along with copy of certificate.
- **9.** Submission of samples if required, for all items indicated in the schedule of requirements. The make of items proposed to be supplied should be indicated in the format of the schedule of requirements and submitted along with the techno commercial unpriced bid without indicating the pricing components.
- **10.** Willingness to execute all orders which are placed to meet emergency requirement on priority basis. The Bidder shall note that standards for workmanship, material and equipment, and references to brand names designated by the Purchaser in the schedule of requirements are intended to be descriptive only and not restrictive. The Bidder may substitute alternative standards, brand names and/or catalogue numbers in his bid, provided that it demonstrates to the Purchaser's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

#### Section V. Policy of University against the Corrupt and Fraudulent Practices

University strictly adheres to its policy against corruption and requires that bidders and their agents, subagents, sub-contractors, suppliers etc. shall not indulge in any kind of corrupt practices, fraudulent practices, collusive practices, coercive practices, obstructive practices or other kinds of corruption declared as crimes under Indian law.

a) If bidder or their agents, subagents, sub-contractors, suppliers etc. are found, directly or indirectly, involved in such practices, bid or agreement and execution thereof at any stage may be rejected or cancelled as the case may be by the University and besides it, University may initiate legal actions including civil and criminal proceeding.

For the purpose of this provision the terms are defined as follows:

- (i) "Corrupt practice" is the offering, giving, receiving, or soliciting, directly or indirectly, of anything of value to influence improperly the actions of another party;<sup>1</sup>
- (ii) "Fraudulent Practices" is any act or omission including a misrepresentation which knowingly or recklessly made to mislead another party to obtain financial or other benefit or to avoid an obligation;<sup>2</sup>
- (iii) "Collusive Practice" is an arrangement between two or more parties designed to achieve an improper purpose, including to influence improperly the actions of another party;<sup>3</sup>
- (iv) "Coercive Practice" is impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;<sup>4</sup>
- (v) "Obstructive Practice" is deliberately destroying, falsifying, altering, or concealing of evidence material to the investigation or making false statements to investigators in order to materially impede an investigation into allegations of a corrupt, fraudulent, coercive or collusive practice; and/or threatening, harassing or

<sup>&</sup>lt;sup>1</sup> For the purpose of this sub-paragraph, "*another party*" refers to a public official acting in relation to the procurement process or contract execution. In this context, "*public official*" includes university staff and employees of other organizations taking or reviewing procurement decisions.

 $<sup>^{2}</sup>$  For the purpose of this sub-paragraph, "party" refers to a public official; the terms "benefit" and "obligation" relate to the procurement process or contract execution; and the "act or omission" is intended to influence the procurement process or contract execution.

<sup>&</sup>lt;sup>3</sup> For the purpose of this sub-paragraph, "party" refers to participants in the procurement process (including public officials) attempting either themselves, or through another person or entity not participating in the procurement or selection process, to simulate competition or to establish bid prices at artificial, non-competitive levels, or are privy to each other's bid prices or other conditions.

<sup>&</sup>lt;sup>4</sup> For the purpose of this sub-paragraph, "party" refers to a participant in the procurement process or contract execution.

intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation.

b) Besides actions under clause (a) University may also take action to blacklist such bidder either indefinitely or for a specified period.

### Part-2 Supply Requirements

### Section VI. Schedule of Requirements

#### Contents

1.	List of Goods and Delivery Schedule	36
2.	List of Related Services and Completion Schedule	37
3.	Technical Specifications	38

#### 1. List of Goods and Delivery Schedule

Line	Description of Goods	Quantity	Physical unit	Final (Project Site) Destination as specified in BDS	Delivery (as per Incoterms) Date		
Item N°					Earliest Delivery Date	Latest Delivery Date	Bidder's offered Delivery date [to be provided by the bidder]
[insert item No]	[insert description of Goods]	[insert quantity of item to be supplied]	[insert physical unit for the quantity]	[insert place of Delivery]	[insert the number of days following the date of effectiveness the Contract]	[insert the number of days following the date of effectiveness the Contract]	[insert the number of days following the date of effectiveness the Contract]
1.	40 channel long term Video EEG system	1	No.	As per tender documents	21	30	
2.	EMG, NCV, EP and Digital EEG Combo system (portable)	1	No.	As per tender documents	21	30	
3.	Repetitive Transcranial Magnetic stimulation (rTMS) machine	1	No.	As per tender documents	21	30	
4.	Transcranial Doppler (TCD) (Battery operated with Robotic Probe Frame)	1	No.	As per tender documents	21	30	
5.	Polysomnography (PSG) Machine with Video EEG	1	No.	As per tender documents	21	30	

## 2. List of Related Services and Completion Schedule

Service	Description of Service	Quantity <sup>1</sup>	Physical Unit	Place where Services shall be performed	Final Completion Date(s) of Services
[insert Service No]	[insert description of Related Services]	[insert quantity of items to be supplied]	[insert physical unit for the items]	[insert name of the Place]	[insert required Completion Date(s)]

1. If applicable

## **3. Technical Specifications**

### "Summary of Technical Specifications"

Item No	Name of Goods or Related Service	Technical Specifications and Standards
1.	Long term Video EEG	The System hardware should consist of :-
	System	<ul> <li>System should have at least 32 EEG channels with 8 Polygraphic channels including 2 DC channels.</li> <li>The amplifier should have USB powered and enabling the amplifier to be connected to any PC/Laptop and transfer data through same USB for portability.</li> <li>Should be supplied along with Video Camera (2MP, IR, PTZ Network Camera)</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body Or USFDA approved and meet international safety standards like IEC and Class IIb.</li> </ul>
		<ul> <li>EEG channels specification :</li> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> The System EEG software should consist of :
		<ul> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> <li>Complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have inbuilt facility in the same software of automatic spike and seizure detection software.</li> <li>Software should have facilities of Amplitude analysis, Spectral analysis, Coherent analysis and many more in 2D &amp; 3D view. And also should have multiple EEG trends windows option simultaneously with live EEG recording.</li> <li>Photic stimulator with LED's.</li> <li>A high resolution 21" color monitor for display of waveforms</li> <li>Provision for hard copy output of recordings on a laser printer of 600 dpi resolution.</li> <li>Should have facility for User definable events and preferred recording and review settings such as amplifier set-ups, event palettes and views to be saved as specific user protocols.</li> </ul>

		Must be operating on Windows 10.
		<ul> <li>Should have option to save EEG as a video clip (*.avi), ASCII-text file</li> </ul>
		(*.txt), XML file (*.xml), set of graphical files (*.bmp or *.emf) or in
		EDF+ format (*.edf).
		Should have facility to check electrode impedance from the head box
		through LED indicator (Red, Green & Yellow) of corresponding
		electrode impedance.
		DVD writer for recording the VEEG data.
		• Facility for exporting data using the software for review on any PC
		without any additional software.
		• Should have facility to prune recordings to store only sections of
		interesting records
		• There should be provision of event markers ideally for marking the
		recording either by patient or Attendant.
		Report generation to be customizable and in MS word & PDF format.
		• The Test information database software which should include
		Information of patient database, resource scheduler and it should be
		customizable according to the end user.
		Should have synchronous Long-term recording of EEG with VIDEO     (Accuracy is 1 Frame)
		<ul><li>(Accuracy is 1 Frame).</li><li>The Video Camera should have facility to record in night mode also.</li></ul>
		<ul> <li>The video camera should have acting to record in high mode also.</li> <li>Video camera should have autofocus zoom lens.</li> </ul>
		<ul> <li>Should have facilities to control the camera position and zoom from</li> </ul>
		computer (Programming of up to 6 preliminary presets of camera
		position).
		···· ,
		The system should be supplied with:
		(Desktop) Intel Core i5 processor with 4GB RAM, network card,
		multimedia speakers, mouse, minimum 1TB hard disk for storing,
		Laser printer (B/W), movable trolley.
		Operating Manuals and user manual
		EEG Cup electrodes (Gold/Silver chloride) (40 nos.)
		Ten-20 Paste (03 Nos.)
		Neuprep Skin preparing gel (03 Nos.)
		Safety and Electromagnetic Compatibility :
		carder and Elocationagnosic Companismy
		• Electromagnetic compatibility (EMC) is provided by IEC 60601-1-
		2:2007 requirements Fulfillment.
		• The device is intended for operation in electromagnetic
		environment, which special features are specified in Appendix 1. As
		for safety, each digital system satisfies IEC 60601-
		1:1988+A1:1991+A2:1995, IEC 60601-1-1:2000, IEC 60601-2-
		26:2002 and IEC 60601-2-40:1998 requirements.
		<ul> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and PE time work parts according</li> </ul>
		interface, it has double isolation and BF type work parts according to IEC 60601-1.
		System should be European CE with 4 digit notified body Or
		USFDA approved and meet international safety standards like IEC and Class IIb.
		Warranty :- 3 (Three) years
2.		The System hardware should consist of :-
4.	EMG, NCV, EP and Digital	
	EEG combo System	<ul> <li>System should be a combination of 21 EEG channels, 4 wide-range Bolygraphic channels, 1 ECG channel, 2 direct current channels</li> </ul>
	(Portable)	Polygraphic channels, 1 ECG channel, 2 direct current channels, and 1 breath channel.
		<ul> <li>The amplifier should have set of built-in stimulators for NCV, EMG,</li> </ul>
		& EP within same amplifier.
		<ul> <li>The system should be able to perform all Electrophysiological tests like EEC_EMC_NCV and EP in one amplifier</li> </ul>
		<ul><li>like EEG, EMG, NCV, and EP in one amplifier.</li><li>Equipment should be imported.</li></ul>
		<ul> <li>Equipment should be imported.</li> <li>It should not require external stimulating module (AEP/VEP/SEP).</li> </ul>
		<ul> <li>The amplifier should have USB powered and enabling the amplifier</li> </ul>

		cted to any PC/Laptop and transfer data through	same
	USB for porta	ability.	
	<ul> <li>The system</li> </ul>	should be supplied with Headphone for AEP,	LED
	Goggle simul		
		should consist of LED Photic stimulator for EEG.	
		should have wireless control unit for easy operation	
		ent shall incorporate features like full optical iso	
	and confirm	<ul> <li>to international standards (IEC) with certific</li> </ul>	cation
	authorities (E	European CE).	
	<ul> <li>Should have</li> </ul>	e facility to check electrode impedance from the	head
		LED indicator (Red, Green & Yellow) of correspo	
	electrode imp		
		shall be upgradable to PSG (Polysomnography	) with
			) with
	the same am	ipiliter.	
	1 EEC abannala	an a sification.	
	21 EEG channels s		
	<ul> <li>Common r</li> </ul>	mode input impedance >400Mohm	
	<ul> <li>Sampling</li> </ul>	Rate - 2 KHz (2000 Hz)	
	<ul> <li>A/D conve</li> </ul>	ertor - 16 bits	
	<ul> <li>Band pass</li> </ul>	s Filter - 0.16 to 200Hz	
	•		
	<ul> <li>CMRR &gt; 1</li> </ul>		
		/el (RMS)< 0.25 μV	
	<ul> <li>Sensitivity</li> </ul>	$\nu$ 0.01 – 12000 $\mu$ V/mm with user definable steps.	
		nnels EMG/NCV/EP) Channels specification:	
	<ul> <li>Common</li> </ul>	n mode input impedance > 200 Mohms	
		g Rate - 40KHz	
		vertor - 16 bits	
		s Filter from 0.02 Hz to 3 KHz	
	<ul> <li>Low pass</li> </ul>	s Filter from 10 Hz – 10 KHz	
	<ul> <li>CMRR &gt;</li> </ul>	100 dB	
	<ul> <li>Noise Let</li> </ul>	evel(RMS) < 0.5 μV	
		er; 50Hz, 60 Hz on or off	
		stimulus duration should be 0.05 to 5 mS, El	
		amplitude should be 1 to 100 mA and Electric stir	mulus
		y: 0.1 to 100 Hz.	
	frequency		
	frequency	y: 0.1 to 100 Hz. are should consist of-	
	frequency	are should consist of-	
	frequency he System softwa • <u>EEG Softwa</u>	are should consist of- are-	abase
	frequency he System softwa • <u>EEG Softwa</u>	a <mark>re should consist of-</mark> are- are- e .NET based software for comprehensive data	abase
	frequency he System softwa • EEG Softwa • Should have search & sto	a <mark>re should consist of-</mark> are- are- e .NET based software for comprehensive data prage.	
	frequency he System softwa • EEG Softwa • Should have search & sto • Should have	a <mark>re should consist of-</mark> are- e .NET based software for comprehensive data brage. e facility for up gradation to record EP signals c	during
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi	are should consist of- are- e .NET based software for comprehensive data brage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif	during ferent
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations	are should consist of- are- are- e .NET based software for comprehensive data brage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil	during ferent
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF	are should consist of- are- e .NET based software for comprehensive data brage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals.	during ferent lity to
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroo	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b	during ferent lity to ipolar
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroor derivations of	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals o ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil > signals. de can be used as a reference one, and the b can be recoded without placing any other addi	during ferent lity to ipolar
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroor derivations of	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b	during ferent lity to ipolar
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroo derivations of reference electron	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals o ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil > signals. de can be used as a reference one, and the b can be recoded without placing any other addi	during ferent lity to ipolar itional
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroo derivations of reference electron	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings	during ferent lity to ipolar itional
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroo derivations of reference ele Provision fo Keyboard/mo	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse	during ferent lity to ipolar itional using
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroor derivations of reference ele Provision fo Keyboard/mo Complete pr	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse	during ferent lity to ipolar itional using
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroor derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc.	during ferent lity to ipolar itional using isition
	frequence he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electroor derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic	during ferent lity to ipolar itional using isition
	frequence he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection.	during ferent lity to ipolar itional using isition spike
	frequence he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/ma Complete pr Sensitivity, fi Should have and seizure of Software sho	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. would have integrated facilities for Amplitude ana	during ferent lity to ipolar itional using isition spike
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/mod Complete pr Sensitivity, fi Should have and seizure of Spectral ana	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. nould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more.	during ferent lity to ipolar itional using isition spike alysis,
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Spectral ana The EEG re	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui liter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. oud have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. aport should be associated with patient data for	during ferent lity to ipolar itional using isition spike alysis,
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Spectral ana The EEG re review and a	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. ioud have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. aport should be associated with patient data for analysis in the software.	during ferent lity to ipolar itional using isition spike alysis,
	he System softwar EEG Softwar Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have Software sh Spectral ana The EEG re review and a Should have	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals or ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. iould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. e port should be associated with patient data for analysis in the software.	during ferent lity to ipolar itional using isition spike alysis,
	he System softwar EEG Softwar Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have Software sh Spectral ana The EEG re review and a Should have	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. ioud have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. aport should be associated with patient data for analysis in the software.	during ferent lity to ipolar itional using isition spike alysis,
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Spectral ana The EEG re review and a Should have Must be run	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. hould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system.	during ferent lity to ipolar itional using isition spike alysis, quick
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrod derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Software sh Spectral ana The EEG re review and a Should have Must be run Facility for ex	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. hould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any	during ferent lity to ipolar itional using isition spike alysis, quick
	frequency he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference electrood derivations of reference electrood Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Spectral ana The EEG re review and a Should have Must be run Facility for ex Should have	are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. Hould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any e facility to prune recordings to store only sectio	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwar EEG Softwar Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference electrood derivations of reference electrood Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Spectral ana The EEG re review and a Should have Must be run Facility for ex- Should have Should have	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. Hould have integrated facilities for Amplitude and alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only section ecords.	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwar EEG Softwar Should have Search & sto Should have EEG acquisis stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for ei Should have interesting re	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. oud have integrated facilities for Amplitude and alysis, Coherent analysis and many more. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only sectior ecords. d be provision of event markers ideally for marking e software in the software ideally for marking	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for ex- Should have interesting re	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. build have integrated facilities for Amplitude and alysis, Coherent analysis and many more. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only sectior ecords. d be provision of event markers ideally for marking ther by patient or Attendant.	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisis stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for es Should have interesting re There should	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. hould have integrated facilities for Amplitude and alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only sectio ecords. d be provision of event markers ideally for marking ther by patient or Attendant. aration to be customizable and in MS word format.	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisis stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for es Should have interesting re There should	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. hould have integrated facilities for Amplitude and alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only sectio ecords. d be provision of event markers ideally for marking ther by patient or Attendant. aration to be customizable and in MS word format.	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/mo Complete pr Sensitivity, fi Should have and seizure Software sho Spectral ana The EEG re review and a Should have Must be run Facility for ex- interesting re There should recording eit Report gene Yes, You cal	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. a facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil P signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. nould have integrated facilities for Amplitude and alysis, Coherent analysis and many more. aport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only section ecords. d be provision of event markers ideally for marking ther by patient or Attendant. aration to be customizable and in MS word format. n save EEG as a video clip (*.avi), ASCII-text file of the software in the software on the software.	during ferent lity to ipolar itional using isition spike alysis, quick
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for ex Should have interesting re There should recording eit Report gene Yes, You can XML file (*.xi	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil orignals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. oudd have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only section ecords. d be provision of event markers ideally for marking ther by patient or Attendant. aration to be customizable and in MS word format. n save EEG as a video clip (*.avi), ASCII-text file ( ml),	during ferent lity to ipolar itional using isition spike alysis, quick PC. ons of ng the (*.txt),
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/mo Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG re review and a Should have Must be run Facility for ex Should have interesting re There should recording eit Report gene Yes, You can XML file (*.xi set of graphi	are should consist of- are. are should consist of- are. e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil orignals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. a inbuilt facility in the same software for automatic detection and artifact detection. oudd have integrated facilities for Amplitude and alysis, Coherent analysis and many more. eport should be associated with patient data for analysis in the software. a software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any a facility to prune recordings to store only section ecords. d be provision of event markers ideally for marking ther by patient or Attendant. avation to be customizable and in MS word format. n save EEG as a video clip (*.avi), ASCII-text file ( mI), ical files (*.bmp or *.emf) or in EDF+ format (*.edf).	during ferent lity to ipolar itional using isition spike alysis, quick PC. ons of ng the (*.txt),
	he System softwa EEG Softwa Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision for Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Software shi Spectral ana The EEG reference and Should have and seizure of Should have Should have Should have Must be run Facility for ex Should have interesting reference and Should have Must be run Facility for ex Should have interesting reference should recording eit Report gene Yes, You can XML file (*.xr set of graphi The Test in	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil orignals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. ould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. oport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any e facility to prune recordings to store only sectic ecords. d be provision of event markers ideally for markin ther by patient or Attendant. eration to be customizable and in MS word format. n save EEG as a video clip (*.avi), ASCII-text file ( mI), ical files (*.bmp or *.emf) or in EDF+ format (*.edf).	during ferent lity to ipolar itional using isition spike alysis, quick PC. ons of ng the (*.txt),
	he System softwar EEG Softwar Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Software sh Spectral ana The EEG re review and a Should have Must be run Facility for ex Should have interesting re There should recording eit Report gene Yes, You can XML file (*.xi set of graphi The Test in Information of	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals or ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil or signals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. iould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. oport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any e facility to prune recordings to store only sectic ecords. d be provision of event markers ideally for markin ther by patient or Attendant. or save EEG as a video clip (*.avi), ASCII-text file ( ml), ical files (*.bmp or *.emf) or in EDF+ format (*.edf).	during ferent lity to ipolar itional using isition spike alysis, quick PC. ons of ng the (*.txt),
	he System softwar EEG Softwar Should have search & sto Should have EEG acquisi stimulations analyze ERF Any electrood derivations of reference ele Provision fo Keyboard/me Complete pr Sensitivity, fi Should have and seizure of Software sh Spectral ana The EEG re review and a Should have Must be run Facility for ex Should have interesting re There should recording eit Report gene Yes, You can XML file (*.xi set of graphi The Test in Information of	are should consist of- are- e .NET based software for comprehensive data orage. e facility for up gradation to record EP signals of ition for Event Related Potential studies with dif (Audio, video & electrical) and also have facil orignals. de can be used as a reference one, and the b can be recoded without placing any other addi ectrodes, for example ear ones. or marking events along with EEG recordings ouse rogrammable control of montage selection, acqui ilter settings, photic stimulator sequences etc. e inbuilt facility in the same software for automatic detection and artifact detection. ould have integrated facilities for Amplitude ana alysis, Coherent analysis and many more. oport should be associated with patient data for analysis in the software. e software controlled notch filter. on Windows 10 operating system. xporting data using the software for review on any e facility to prune recordings to store only sectic ecords. d be provision of event markers ideally for markin ther by patient or Attendant. eration to be customizable and in MS word format. n save EEG as a video clip (*.avi), ASCII-text file ( mI), ical files (*.bmp or *.emf) or in EDF+ format (*.edf).	during ferent lity to ipolar itional using isition spike alysis, quick PC. ons of ng the (*.txt),

EEG Consist of:-
<ul> <li>EEG Cup electrodes (Gold plated) (40 nos.)</li> </ul>
Photic Stimulator with stand
<ul> <li>Ten-20 Paste (03 Nos.)</li> <li>Neuprep Skin preparing gel (03 Nos.)</li> </ul>
EMG/NCV/EP Software :
<ul> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> </ul>
Motor NCV with automatic marking
<ul> <li>Sensory NCV with automatic marking</li> <li>Motor and Sensory inching</li> </ul>
Tremor analysis should be supplied with     Assolution and the supplied with
<ul> <li>Accelerometer/Mechanomyography</li> <li>F wave with split screen display with automatic marking of F</li> </ul>
responses showing the Max F, Min F and F block values.
<ul> <li>H Reflex ,Blink Reflex, Sacral Reflex, Bulbocavernous Reflex, T Reflex, Galvanic Skin Response</li> </ul>
Repetitive Stimulation
<ul> <li>Insertional /Spontaneous EMG recording for unlimited duration on hard disk for unlimited times or unlimited buffer storage.</li> </ul>
Quantitative EMG test features must be included.     EMG replay of stored EMG data from hard disk with audio
<ul> <li>EMG replay of stored EMG data from hard disk with audio</li> <li>Multiple Motor unit Analysis</li> </ul>
<ul> <li>Single Motor unit Analysis.</li> <li>Jitter analysis for single fibre test</li> </ul>
Incremental MUNE
Mune (MUP Decomposition)
<ul> <li>Macro EMG</li> <li>Short-, middle- and long-latency auditory EP</li> </ul>
Short- and long-latency somatosensory EP
<ul> <li>Auditory stimulation with clicks and tones.</li> <li>Visual evoked potentials: Pattern reversal VEP, Flash VEP, Goggle</li> </ul>
VEP • Cognitive EP- P300,CNV,MMN
<ul> <li>Automatic rapid report generation with unlimited user templates</li> </ul>
<ul> <li>Facility of comparing patient data with normative data &amp; to flag abnormal values automatically.</li> </ul>
Automatic sentence generator.
<ul> <li>Report generation to be customizable and in MS word format &amp; PDF format.</li> </ul>
<ul> <li>Must be operating on Windows 10 operating system.</li> </ul>
<ul> <li>Facility of including the waveform &amp; numerical data as per user requirement in patient report.</li> </ul>
<ul> <li>Provision for hard copy output of recordings on a laser printer of</li> </ul>
<ul><li>600 dpi resolution.</li><li>Dedicated control panel can function via Bluetooth or USB</li></ul>
interface.
Possibility to connect a Magnetic stimulator
Set of EMG electrodes: • Surface electrode – 1 pair.
<ul> <li>Stimulating bar electrode – 1 no.</li> </ul>
<ul> <li>Ring electrode – 1 pair.</li> <li>Ground electrode with cable (paediatric) – 01nos.</li> </ul>
<ul> <li>Ground electrode with cable (adult) – 01nos.</li> </ul>
<ul> <li>Disposable concentric needle electrode (Adult) – 25 pcs.</li> <li>Holder for needle electrode connection – 1 nos.</li> </ul>
Disposable surface electrode (set of 100 pcs.)
<ul> <li>Adapter for disposable electrodes connection with Alligator clip – 2 pcs</li> </ul>
<ul> <li>Gold plated cup electrode – 10 pcs.</li> </ul>
<ul> <li>Pup-jack linker (Jumper Electrodes) – 2 pcs.</li> </ul>
Set of stimulators:-
Electrical Handheld stimulator     TT Monitor for pattern stimulator
<ul> <li>17" TFT Monitor for pattern stimulator</li> <li>Visual stimulator (LED goggles)</li> </ul>

3	RepetitiveTranscranial	
3.	RepetitiveTranscranial Magnetic Stimulator (RTMS) system	<ol> <li>It should be a highly effective non-invasive biphasic magnetic stimulator for bilateral cortical or peripheral and central nerve pathways stimulation. The maximum output power should be 100% at 41-42Hz, 50% output at 96-100 Hz and 40% output at 100Hz. The system should have         <ul> <li>Integrated color touch screen display</li> <li>Integrated 2 channel Motor Evoked Potential monitoring with acquisition software for marking of latency and amplitude cursors; 6 no's of electrodes to be provided for MEP recording.</li> <li>Should be programmable to deliver Single pulse, Repetitive or train of sessions.</li> <li>Unique Burst Protocol option, such as Theta Burst</li> </ul> </li> <li>The pulse train interval should be user selectable with 0.1s increment and pulse width of 400 µS with maximum upper frequency of 100Hz.</li> <li>The system should be supplied with following CE approved coils         <ul> <li>Minimum 70mm Air Film/hydrothermal technology Coil which delivers up to 3000 pulses at 120% motor threshold at 10Hz stimulation (Biphasic waveform) over a 30 minute period and remains within allowed temperature limits. – 1 no.</li> <li>Figure eight coil at least 25mm - 1 no</li> <li>Minimum 70mm double placebo coil - 1 no</li> <li>EMG Interface Module to integrate system with existing EMG machine g. The system should be supplied with desktop PC with the following specs : Intel i7 processor 4 GB RAM, 1TB Sata Hard Disk, 21" TFT monitor, compatible Latest OS, DVD Burner, suitable UPS</li> </ul> </li> <li>The system should be supplied with suitable Power Supply Units (PSU)</li> <li>The system and the coils should be European CE or US FDA certified</li> <li>Should also have provision to attach other coil such as 90mm high power circular coil, 110mm double cone coil with tTMS system and the rates of these coils should be quoted separately.</li> <li>The bidder must</li></ol>
		<ul> <li>motor evoked potential, input-output curve, cortical silent period, motor mapping</li> <li>10. Should be HL7 compliant system</li> <li>Warranty :- 3 (Three) years</li> </ul>
4.	TRANSCRANIAL	<ol> <li>Two Channel Transcranial Doppler capable of Intracranial, Extracranial and intraoperative use.</li> </ol>
	DOPPLER(TCD) (Battery	2. Should be supplied with 1.6 MHz PW probe (1 Nos.), 2 MHz Probesfor Probe
	Operated with Robotic Probe	Holder (2 Nos.), 4 MhzCW/PW Probe, Total Probes: 4 Probes 3. Machine should have additional probe connection available for 1.6/2 Mhz
	Frame)	(Two Nos.), 4 Mhz, 8 Mhz and 16 Mhz so that the machine can be site
		upgraded to additional probe at any time. 4. Should have an integrated Battery inside the main unit.
		5. Should display battery level status information in software status bar.
		<ol><li>Should allow the main unit to hand carry while the charger is unplugged and the system is still powered on, and to allow the main unit to run</li></ol>
		continuously on battery and unplugged from charger for at least 3 hours.
		<ol> <li>Should have a 2MHz bilateral robotic headset that allows to automatically scan the temporal window and automatically detect the optimal velocity</li> </ol>
		signal.
		<ol> <li>The robotic headset must be a closed loop control that allows accuracy and repeatability in probe movements, and not an open loop control.</li> </ol>
		9. Should have the ability to simultaneously control, scan, and move both
		sides of the bilateral robotic headset. 10. Should have the robotic headset scan to have 3 different depth scanning
		options: at the main depth, around the main depth with a configurable
		tolerance, or across all depths. 11. Should have a special Search and Locate Mode for the robotic headset, to
		allow to search and locate the temporal window probe position without

		user software interaction.
	12.	Should have a special HEAT MAP display for the robotic headset that
		updates in real time during the robotic scan to identify and display forward
		and reverse blood flows.
	13.	Should have the option for the robotic headset, to click on a location on the
		HEAT MAP to move the probe to the new target location.
	14.	Should have for the robotic headset 2 different scan options: a high
		resolution scan and a quick scan that allows shorter scan duration.
	15.	Should have for the robotic headset scan a special display of the
		accumulated m-mode, and ability to scroll the depths over this display to
	4.0	view specific depths in the HEAT MAP.
	16.	Should have ability to modify the spectrum and audio after freeze by
		changing the Depth, Gain, Scale, Sample Volume and High Pass Filterand
		recalculate the spectrum FFT and audio by using stored history of raw data and 250 gates that were acquired during measurement.
	17.	Should have ability to view spectrum in review station, and allow to modify
	17.	the spectrum by changing the Depth, Gain, Scale, Sample Volume and High
		Pass Filter, and use the stored raw data and 250 gates to recalculate the FFT
		and audio.
	18.	Should have an integrated VESA mount on the back panel of the main unit
		to allow simple and safe connection to any VESA mount on carts, trolleys,
		stands, or monitoring arms.
	19.	-
	20.	
	21.	
	22.	Should have ability to modify number of spectrum windows in offline and
		change the depth of each spectrum window in offline
	23.	Should have ability to modify Depth in offline, then replay the new depth
		with spectrum and Doppler audio
	24.	Should have main key that allows completing a protocol with just one main
		key
		Should have automatic save to ensure no measurement is lost
	26.	Should have special Power M Mode display for each detected HITS in 250
		gates and a displayed time resolution of 125 microseconds.
	27.	Should support network share of examinations, patients and other data
	20	between multiple review stations using background sync.
	28.	Should support optionally connecting to hospital SQL server to share a
	29.	database Should support loading examinations and conjouing from the Doctor's
	29.	Should support loading examinations and replaying from the Doctor's computer and office
	30.	Should have dedicated protocols for specialty tests including protocols for
	50.	Sickle Cell Disease and Vasospasm with configurable thresholds for
		automatic interpretations. Each specialty test protocol should have a
		dedicated print layout with a separate configuration.
	31.	Should support Special design layout for each test so that each type of test
		can have a separate completely customized configuration for screen layout
		and print layout
	32.	Should have patient details customizations including customized free text
		fields, patient status fields, risk factors, physical examination, additional
		fields, symptoms and indications fields, diagnosis codes.
	33.	Should have advanced free text report including configurable templates
		(save time by building advanced templates), tables, Camera pictures, quick
		text (save time by quickly inserting commonly used text), automatic TAGS
		(use special <tags> as placeholders for data to build sophisticated reports),</tags>
		RTF Import/Export(integrate with MS Word for easier template building)
	34.	Should have automatic protection of insonation power during Ophthalmic
	25	and Siphon measurements.
	35.	Should have BHI calculation for Breath Holding protocol
	36.	Should have customized screen layout including modifying screen location
	27	of M Mode, parameters, etc. for enhanced ease of use
	37.	Should have Statistics including collection of pie charts and advanced X/Y chart to review different patient related statistical data
	38.	Should have advanced statistics with Excel support to export entire
	50.	database to Excel
	39.	
	40.	Should have password protection to protect norm accidental changes Should have self-diagnostics procedure including a report
	40. 41.	
	42.	, , ,
		Should have 4D Analysis display
	44.	
	45.	Should have zoom in/out option for parameters trends using drag & drop

	46.	Should have configurable alerts for monitoring
	47.	Should have delta mean marker for monitoring
	48.	Should allow quick review of previous exams by browsing through all
		summary screens of past examinations with few clicks.
	49.	Should have customized physicians and examiners list
	50.	Should have advanced database search based on criteria including dates,
	<b>F</b> 4	protocol, risk factors, etc.
	51.	
	гa	research purposes
	52.	Should have flexible configurable storage settings so that the user can decide exactly what is saved and what is not for optimal disk space use (Raw
		Data or the processed data)
	53.	Should have a data compression algorithm to optimize use of disk space
		Frequency range: pulsed wave: 3 – 24 kHz
	55.	Should have both phasic and power M Mode feature with able to re-adjust
		the 250 gates digitally per probe and with the ability for each M Mode gate
		to be as low as 0.08mm.
		Should have multi-frequency operation.
	57.	Should have automatic emboli detection with option to detect all HITS on all
	<b>F</b> 0	gates in the entire depth range.
	58. 59.	Should have user-definable defaults for individual blood vessels
	59.	Should have interactive summary screen which displays all studies performed on a patient on a single screen, enabling immediate comparison
		between the right and left sides of the brain.
	60.	Should have long term monitoring with trending of selected parameters
		Probe Holder should have fixation device
	62.	Should have 8 Analog inputs and 4 Analog output to interface analog signal
		from ETCO2 or any other analog device to and from this system. Also should
		have 2 digital inputs and 2 digital outputs.
	63.	Should display 8 different doppler spectrum windows in 8 different depths
		simultaneously.
		Should have manual control of gain
		Should have Vasomotor Reactivity Test.
		Should have FFT size up to 512points.
	07.	Should have ability to change spectrum display size from 1 second to 3 minutes
	68.	Should be able to display Pulsed wave parameters like Peak Velocity, Mean
	00.	Velocity, Diastolic Velocity, Pulsatility index, Resistivity index, Heart rate and
		Rise Time.
	69.	Should be able to calculate automatically and display the Lindegaard Ratio,
		and include in the report.
	70.	There should be provision of up to 6 horizontal and 2 vertical cursors for
		measuring the values manually.
		The system should measure beat to beat pulse(heart rate)
	72.	
	73.	spectrums with Color M mode feature. Should have both Phasic and power M-mode display option.
		Should have both Phasic and power in-mode display option. Should be able to generate report, with option of transferring all waveforms
	<i>,</i>	into report.
	75.	Should have facility of storing the waveforms of complete spectrum with
		audio digitally and replaying with audio the complete study as recorded.
	76.	Should have facility to export the recorded study to MP4 format.
	77.	, 5
		control must have also configurable buttons.
	78.	Should be supplied with PC of Quad Core 2GHz, at least 4 GB RAM, 18" LED
		touch panel monitor, 4 USB 3.0 Ports, Genuine Windows 10 operating
		system, Mini Keyboard, Optical Mouse, 1TB SSD Enterprise Grade hard drive with Power Loss Protection (Data at Post & Data in Elight)
	79	with Power-Loss Protection (Data at Rest & Data in Flight). Should be Portable System, Single Unit.
		The data can be exported of any display in either Excel, PDF, RTF format.
		Should be supplied with Color Laser Printer.
		Should be supplied with Metallic trolley with castor.
		Should be provided with DICOM facility, ready to use.
	84.	Should be provided with DICOM facility including both online and offline
		worklist, BMP/JPEG/Encapsulated PDF report support, multiple destinations
		(each with separate configuration), Queue for reports that could not be
	05	sent, automatic retrying in background and automatic sending.
	85.	Should support HL7 Patient Registration / Orders / Report / Query Patient Details
	86.	Should support Automatic PDF Report Export with a customized file path &
		name template

		87. Should have the ability to send MP4 (H.264) Video of each measurement
		including Doppler audio to PACS. 88. Should have the ability to send a picture or MP4 (H.264) Video of each
		detected HITS including Doppler audio to PACS.
		89. Should have the ability to send a picture or MP4 (H.264) Video of each
		monitoring/PFO event including Doppler audio to PACS.
		<ol> <li>Should have the option to customize a title of the report according to the used protocol.</li> </ol>
		91. Should have the option to quickly exclude measurements from the report
		without deleting them.
		92. Should have the option to relocate predefined, manually saved, HITS or
		artifacts event makers by using drag & drop. 93. Should have up to 10 side spectrums that will be distributed equally along
		the selected M Mode range.
		94. Should have the ability to quickly review/edit/delete saved monitoring/PFO
		events directly from summary screen.
		<ol> <li>Should have the option to include M Mode display in report</li> <li>Should have the option to automatically display on screen and in report the</li> </ol>
		Basilar/Vertebral ratio.
		97. Should have the option to automatically include long-term or day by day
		trends to the report including velocities summary table, delta mean, graphs,
		lindegaard ratio trends, and trends parameters configuration. 98. Should have a dedicated screen to manually insert EtCO2 values to VMR
		test.
		99. Should have the ability to replay HITS events directly from the HITS Analysis
		screen and with a configurable replay audio length around the event.
		100. Should have the ability to add Date/Time stamp to measurements 101. Should be US FDA and CE Approved. Certificate should be attached.
		Warranty :- 3 (Three) years
5.	Polysomnography (PSG)	
	System with Video EEG	The System hardware should consist of :-
		System should have at least 32 EEG channels with 9 Polygraphic
		channels including built-in SpO2 and 2 DC channels.
		<ul> <li>The amplifier should have USB powered and enabling the amplifier to be connected to any PC// apton and transfer data through some</li> </ul>
		to be connected to any PC/Laptop and transfer data through same
		USB for portability.
		USB for portability. <ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ)</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li><u>EEG channels specification</u>:</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS) &lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification :</li> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS) &lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS) &lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :</li> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> The System EEG software should consist of :- <ul> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones. Provision for marking events along with VEEG recordings using</li></ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ilb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> <li>Complete programmable control of montage selection, acquisition</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS) &lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> <li>Complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have inbuilt facility in the same software of automatic spike</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEC channels specification : <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> <li>Complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have inbuilt facility in the same software of automatic spike and seizure detection software.</li> </ul>
		<ul> <li>Should be supplied along with Video Camera (1MP, IR, PTZ Network Camera).</li> <li>The system should consist of LED Photic stimulator for EEG.</li> <li>The equipment shall incorporate features like full optical isolation and confirm to international standards with certification authorities.</li> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red, Green &amp; Yellow) of corresponding electrode impedance.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> <li>EEG channels specification :         <ul> <li>Common mode input impedance &gt;400Mohm</li> <li>Sampling Rate - 2 KHz (2000 Hz)</li> <li>A/D convertor - 16 bits</li> <li>Band pass Filter - 0.16 to 200Hz</li> <li>CMRR &gt; 120 dB</li> <li>Noise Level (RMS)&lt; 0.25 μV</li> <li>Sensitivity 0.01 – 12000 μV/mm with user definable steps.</li> </ul> </li> <li>The System EEG software should consist of :-</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Any electrode can be used as a reference one, and the bipolar derivations can be recoded without placing any other additional reference electrodes, for example ear ones.</li> <li>Provision for marking events along with VEEG recordings using Keyboard/mouse</li> <li>Complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have inbuilt facility in the same software of automatic spike</li> </ul>

also should have multiple EEG trends windows option simultaneously
with live EEG recording.
Photic stimulator with LED's.     A high resolution 10" color manifest for display of waveforms
<ul> <li>A high resolution 19" color monitor for display of waveforms</li> <li>Provision for hard copy output of recordings on a laser printer of 600</li> </ul>
<ul> <li>Provision for hard copy output of recordings on a laser printer of 600 dpi resolution.</li> </ul>
<ul> <li>Should have facility for User definable events and preferred recording</li> </ul>
and review settings such as amplifier set-ups, event palettes and
views to be saved as specific user protocols.
Should have provision for notch filter.
<ul> <li>Must be operating on Windows 7 and Windows 8.</li> </ul>
<ul> <li>Should have option to save EEG as a video clip (*.avi), ASCII-text file</li> </ul>
(*.txt), XML file (*.xml), set of graphical files (*.bmp or *.emf) or in
EDF+ format (*.edf).
<ul> <li>Should have facility to check electrode impedance from the head box through LED indicator (Red. Croop &amp; Vallow) of corresponding</li> </ul>
through LED indicator (Red, Green & Yellow) of corresponding electrode impedance.
<ul> <li>DVD writer for recording the VEEG data.</li> </ul>
<ul> <li>Facility for exporting data using the software for review on any PC</li> </ul>
without any additional software.
Should have facility to prune recordings to store only sections of
interesting records
<ul> <li>The equipment shall incorporate features like full optical isolation,</li> </ul>
anti-cautery devices and shall confirm to international standards with certification authorities.
<ul> <li>There should be provision of event markers ideally for marking the</li> </ul>
recording either by patient or Attendant.
Report generation to be customizable and in MS word format.
The Test information database software which should include
Information of patient database, resource scheduler and it should be
customizable according to the end user.
<ul> <li>Should have synchronous Long-term recording of EEG with VIDEO (Accuracy is 1 Frame).</li> </ul>
<ul> <li>The Video Camera should have facility to record in night mode also.</li> </ul>
<ul> <li>Video camera should have autofocus zoom lens.</li> </ul>
<ul> <li>Should have facilities to control the camera position and zoom from</li> </ul>
computer (Programming of up to 6 preliminary presets of camera
position).
Delveennegrenk Seffwere eensiste of .
Polysomnograph Software consists of :-
<ul> <li>Should have Sleep Analysis hardware and software for Long-term</li> </ul>
<ul> <li>Should have Sleep Analysis hardware and software for Long-term recording of the following channels during a sleep which allow to</li> </ul>
recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered
recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow,
recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring,
recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring,</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, addominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters,</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording and also synchronous review of</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have Quick disconnection and connection of the patient and</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, addominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> <li>Should have automatic search and classification of apnea and</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, addominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and periodic limb movement events. HR, body position, SpO2 trends visualization.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and periodic limb movement events. HR, body position, SpO2 trends visualization.</li> <li>Should be calculation of indices of apnea and hypopnea, desaturations, parameters of snoring, heart rate variability, limb</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and periodic limb movement events. HR, body position, SpO2 trends visualization.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, addominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG scoring</li> <li>Should have complete programmable control of montage selection, acquisition Sensitivity, filter settings, photic stimulator sequences etc.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and periodic limb movement events. HR, body position, SpO2 trends visualization.</li> <li>Should be calculation of indices of apnea and hypopnea, desaturations, parameters of snoring, heart rate variability, limb movements and periodic limb movements, body position separately and in connection with sleep stages.</li> </ul>
<ul> <li>recording of the following channels during a sleep which allow to analyze both the sleep structure (sleep stage) and sleep-disordered breathing includes EEG, EOG, EMG from chin area, ECG, airflow, chest movements, abdominal movements, body position, snoring, SpO2, EMG from limbs (limb movements), etc.</li> <li>Should have .NET based software for comprehensive database search &amp; storage.</li> <li>Should be supplied with all the transducers required for recording the Pressure, Respiration, Body position etc.,</li> <li>Should have automatically and manual sleep stage scoring, automatic calculation of the large number of sleep parameters, displaying of bar chart of sleep stage distribution.</li> <li>Possibility of the synchronous recording of video and audio information during PSG recording and also synchronous review of video and audio during PSG scoring</li> <li>Should have Quick disconnection and connection of the patient and the recording equipment which do not result in electrodes position removal if you use the patient unit.</li> <li>Should have automatic search and classification of apnea and hypopnea events, desaturation, snoring, limb movement and periodic limb movement events. HR, body position, SpO2 trends visualization.</li> </ul>

<ul> <li>There should be provision of event marking the recording either by patient or Attendant.</li> <li>Generation of PSG checkup report including graphic data (hyproegram, trends, apnea and hypopnea, destauration, sonorgin, limb movements and periodic limb wovements and sleep, disordneed breathing parameters, bar charts of sleep staged distribution.</li> <li>The Test information database software which should include the conting bit for each of the stage distribution.</li> <li>The Test information database software which should include the conting bit for each and it should approximate the conting on the revue settings such as amplifier server, event palettes and views to be saved as specific user protocols.</li> <li>Stouid have facility for User definable events and preference information database software which should include the conting and review settings such as amplifier serve, event palettes and views to be saved as specific user protocols.</li> <li>Stouid PSG Accessories :-</li> <li>Gold plated Cup electrode - 20 pcs.</li> <li>Disposable Adhesive Surface Electrodes (2 mtr) – 100 Pcs.</li> <li>Ariflow sensor - 1</li> <li>Storing sensor -1</li> <li>Storing sensor -1</li> <li>Body position sensor -1</li> <li>Body position sensor -1</li> <li>Finger sensor for SpO2 recording -1</li> <li>Portable Paleint Unit (Ouck Disconsection Box) for Commutation to EEG and PSG Channels, ILE amandatory that paletict an oa anywhere with detachable 32 channels 10-20 electrode system without any loss of EEG data.</li> <li>Commutation to EEG and PSG Channels, ILE amandatory that paletic an oa anywhere with detachable 32 channels 10-20 electrodes (2001/Silver chloride) (35 ncs.)</li> <li>The system should be supplied with:</li> <li>(Desktop) Intel Core is processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 5000 GB hard disk to to storing Laser printer (BW), movable tolley, USP (30 Min backup) 50 (2001 reguirements Fulfilment).</li></ul>	
<ul> <li>Gold plated Cup electrode - 20 pcs.</li> <li>Disposable Adhesive Surface Electrodes (2 mtr) – 100 Pcs.</li> <li>Airflow sensor - 1</li> <li>Airflow pressure sensor - 1</li> <li>Snoring sensor - 1</li> <li>Chest respiratory efforts sensor - 1</li> <li>Body position sensor - 1</li> <li>Finger sensor for SpO2 recording - 1</li> <li>Portable Patient Unit (Quick Disconnection Box) for Commutation to EEG and PSG Channels. It is mandatory that patient can qo anwhere with detachable 32 channels 10-20 electrode system without any loss of EEG data.</li> <li>Compatible CPAP titration device - 1</li> </ul> The system should be supplied with : <ul> <li>(Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing. Laser printer (BW), movable trolley. UPS (30 Min backup)</li> <li>Operating Manuals and user manual</li> <li>EEG Cup electrodes (Cold/Silver chloride) (35 nos.)</li> <li>Ten-20 Paste (10 Nos.)</li> <li>Neuprep Skin preparing gel (5 Nos.)</li> </ul> Safety and Electromagnetic Compatibility EIC <ul> <li>Electromagnetic compatibility E</li> <li>Electromagnetic compatibility E</li> <li>Electoral equirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1 :1988+A1:1991+A2:1995, IEC 60601-1:2:007 requirements. <ul> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul></li></ul>	<ul> <li>recording either by patient or Attendant.</li> <li>Generation of PSG checkup report including graphic data (hypnogram, trends, apnea and hypopnea, desaturation, snoring, limb movements and periodic limb movements events), calculation results of sleep parameters and sleep-disordered breathing parameters, bar charts of sleep stage distribution.</li> <li>The Test information database software which should include Information of patient database, resource scheduler and it should be customizable according to the end user.</li> <li>Should have facility for User definable events and preferred recording and review settings such as amplifier set-ups, event palettes and views to be saved as specific user protocols.</li> </ul>
<ul> <li>Gold plated Cup electrode - 20 pcs.</li> <li>Disposable Adhesive Surface Electrodes (2 mtr) – 100 Pcs.</li> <li>Airflow sensor - 1</li> <li>Airflow pressure sensor - 1</li> <li>Snoring sensor - 1</li> <li>Chest respiratory efforts sensor - 1</li> <li>Body position sensor - 1</li> <li>Finger sensor for SpO2 recording - 1</li> <li>Portable Patient Unit (Quick Disconnection Box) for Commutation to EEG and PSG Channels. It is mandatory that patient can qo anwhere with detachable 32 channels 10-20 electrode system without any loss of EEG data.</li> <li>Compatible CPAP titration device - 1</li> </ul> The system should be supplied with : <ul> <li>(Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing. Laser printer (BW), movable trolley. UPS (30 Min backup)</li> <li>Operating Manuals and user manual</li> <li>EEG Cup electrodes (Cold/Silver chloride) (35 nos.)</li> <li>Ten-20 Paste (10 Nos.)</li> <li>Neuprep Skin preparing gel (5 Nos.)</li> </ul> Safety and Electromagnetic Compatibility EIC <ul> <li>Electromagnetic compatibility E</li> <li>Electromagnetic compatibility E</li> <li>Electoral equirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1 :1988+A1:1991+A2:1995, IEC 60601-1:2:007 requirements. <ul> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul></li></ul>	Set of PSG Accessories :-
The system should be supplied with :         • (Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing, Laser printer (B/W), movable trolley, UPS (30 Min backup)         • Operating Manuals and user manual         • EEG Cup electrodes (Gold/Silver chloride) (35 nos.)         • Ten-20 Paste (10 Nos.)         • Neuprep Skin preparing gel (5 Nos.)         Safety and Electromagnetic Compatibility :         • Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfilment.         • The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1:98+A1:1991-A2:1995, IEC 60601-2-2:26:2002 and IEC 60601-2-40:1998 requirements.         • The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.         • System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class Ib.	<ul> <li>Gold plated Cup electrode - 20 pcs.</li> <li>Disposable Adhesive Surface Electrodes (2 mtr) – 100 Pcs.</li> <li>Airflow sensor – 1</li> <li>Airflow pressure sensor - 1</li> <li>Snoring sensor - 1</li> <li>Chest respiratory efforts sensor - 1</li> <li>Abdominal respiratory efforts sensor - 1</li> <li>Body position sensor - 1</li> <li>Finger sensor for SpO2 recording - 1</li> <li>Portable Patient Unit (Quick Disconnection Box) for Commutation to EEG and PSG Channels. It is mandatory that patient can go anywhere with detachable 32 channels 10-20 electrode system without any loss of EEG data.</li> </ul>
<ul> <li>(Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing, Laser printer (B/W), movable trolley, UPS (30 Min backup)</li> <li>Operating Manuals and user manual</li> <li>EEG Cup electrodes (Gold/Silver chloride) (35 nos.)</li> <li>Ten-20 Paste (10 Nos.)</li> <li>Neuprep Skin preparing gel (5 Nos.)</li> </ul> Safety and Electromagnetic Compatibility : <ul> <li>Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1:988+A1:1991+A2:1995, IEC 60601-1-1:2000, IEC 60601-2:26:2002 and IEC 60601-2-40:1998 requirements. <ul> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul></li></ul>	
<ul> <li>(Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing, Laser printer (B/W), movable trolley, UPS (30 Min backup)</li> <li>Operating Manuals and user manual</li> <li>EEG Cup electrodes (Gold/Silver chloride) (35 nos.)</li> <li>Ten-20 Paste (10 Nos.)</li> <li>Neuprep Skin preparing gel (5 Nos.)</li> </ul> Safety and Electromagnetic Compatibility : <ul> <li>Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1:988+A1:1991+A2:1995, IEC 60601-1-1:2000, IEC 60601-2:26:2002 and IEC 60601-2-40:1998 requirements. <ul> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul></li></ul>	The system should be supplied with :
<ul> <li>Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1988+A1:1991+A2:1995, IEC 60601-1-1:12000, IEC 60601-2:26:2002 and IEC 60601-2-40:1998 requirements.</li> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul>	<ul> <li>(Desktop) Intel Core i5 processor with 4GB RAM, network card, multimedia speakers, mouse, minimum 500 GB hard disk for storing, Laser printer (B/W), movable trolley, UPS (30 Min backup)</li> <li>Operating Manuals and user manual</li> <li>EEG Cup electrodes (Gold/Silver chloride) (35 nos.)</li> <li>Ten-20 Paste (10 Nos.)</li> </ul>
<ul> <li>Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1988+A1:1991+A2:1995, IEC 60601-1-1:12000, IEC 60601-2:26:2002 and IEC 60601-2-40:1998 requirements.</li> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul>	Safety and Electromagnetic Compatibility :
	<ul> <li>Electromagnetic compatibility (EMC) is provided by IEC 60601-1-2:2007 requirements Fulfillment.</li> <li>The device is intended for operation in electromagnetic environment, which special features are specified in Appendix 1. As for safety, each digital system satisfies IEC 60601-1:1988+A1:1991+A2:1995, IEC 60601-1-1:2000, IEC 60601-2:26:2002 and IEC 60601-2-40:1998 requirements.</li> <li>The digital unit is supplied by regulated power supply through USB interface, it has double isolation and BF type work parts according to IEC 60601-1.</li> <li>System should be European CE with 4 digit notified body approved and meet international safety standards like IEC and Class IIb.</li> </ul>
	<u> </u>

#### For all 5 equipments mentioned above

<u>Note</u>:- Net Present Value (NPV) of Comprehensive Annual Maintenance contract charges (CAMC) quoted for 7 years after the warranty period shall be added to the bid price for evaluation and will be calculated after discounting the quoted price by a discounting factor of 10% per annum. However the payment of CAMC shall be made to the successful bidder at approved rates. The quoted bid price and CAMC prices will also be added for comparison/ranking purpose for evaluation.

# Part-3 Contract

## **Section VII. General Conditions of Contract**

### **Table of Clauses**

1.	Definitions	. 51
2.	Contract Documents	. 52
3.	Corrupt and Fraudulent Practices	. 52
4.	Interpretation	. 52
5.	Language	. 53
6.	Joint Venture, Consortium or Association	. 53
7.	Eligibility	. 53
8.	Notices	. 54
9.	Governing Law	. 54
10.	Settlement of Disputes	. 54
11.	Obligations During Arbitrations	. 55
12.	Scope of Supply	. 55
13.	Delivery and Documents	. 55
14.	Supplier's Responsibilities	. 55
15.	Contract Price	. 55
16.	Terms of Payment	. 55
17.	Taxes and Duties	. 56
18.	Performance Security	. 56
19.	Copyright	. 57
20.	Confidential Information	. 57
21.	Subcontracting	. 58
22.	Specifications and Standards	. 58
23.	Packing and Documents	. 58
24.	Insurance	. 59
25.	Transportation and Incidental Services	. 59
26.	Inspections and Tests	. 60
27.	Liquidated Damages	. 61
28.	Warranty	. 61
29.	Patent Indemnity	. 62
30.	Force Majeure	. 62
31.	Change Orders and Contract Amendments	. 63
32.	Extensions of Time	. 63
33.	Termination	. 64
34.	Assignment	. 65

### Section VII. General Conditions of Contract

# **Definitions** 1.1 The following words and expressions shall have the meanings

1

- hereby assigned to them:
  - (a) "University" means Banaras Hindu University established under Banaras Hindu University Act 1915.
  - (b) "Contract" means the Contract Agreement entered into between the Purchaser and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
  - (c) "Contract Documents" means the documents listed in the Contract Agreement, including any amendments thereto.
  - (d) "Contract Price" means the price payable to the Supplier as specified in the Contract Agreement, subject to such additions and adjustments thereto or deductions therefrom, as may be made pursuant to the Contract.
  - (e) "Day" means calendar day.
  - (f) "Completion" means the fulfillment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
  - (g) "GCC" means the General Conditions of Contract.
  - (h) "Goods" means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Purchaser under the Contract.
  - (i) "The Project Site," term where applicable, means the place of work named in the Special Conditions of Contract (SCC).
  - (j) "Purchaser" means Banaras Hindu University, institute, faculty, department and other entities of the University competent for purchasing Goods and Services, as specified in the SCC.
  - (k) "Related Services" means the services incidental to the supply of the goods, such as insurance, installation, training and initial maintenance and other such obligations of the Supplier under the Contract.
  - (1) "SCC" means the Special Conditions of Contract.
  - (m) "Subcontractor" means any person, private or government entity, or a combination of the above, to whom any part of the Goods to be supplied or execution

of any part of the Related Services is subcontracted by the Supplier.

- (n) "Supplier" means the person, private or government entity, or a combination of the above, whose bid to perform the Contract has been accepted by the Purchaser and is named as such in the Contract Agreement.
- (o) "The Project Site," where applicable, means the place named in the SCC.
- 2 Contract
   Documents
   2.1 Subject to the order of precedence set forth in the Contract Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory. The Contract Agreement shall be read as a whole.
  - upt and3.1The University requires compliance with its policy against<br/>the corrupt and fraudulent practices as set forth Section- V.
    - 3.2 The Purchaser requires the Supplier to disclose any commissions or fees that may have been paid or are to be paid to agents or any other party with respect to the bidding process or execution of the Contract. The information disclosed must include the name and address of the agent or other party, the amount and currency, and the purpose of the commission, gratuity or fee.
- **4 Interpretation** 4.1 Unless the context requires otherwise, singular means plural and vice versa.
  - 4.2 The Contract constitutes the entire agreement between the Purchaser and the Supplier and supersedes all communications, negotiations and agreements (whether written or oral) of the parties with respect thereto made prior to the date of Contract.
  - 4.3 Amendment

No amendment or other variation of the Contract shall be valid unless it is reduced to writing, dated, expressly refers to the Contract, and is signed by the duly authorized representative of each party thereto.

- 4.4 Nonwaiver
  - a. Subject to GCC Sub-Clause 4.5(b) below, no relaxation, forbearance, delay, or indulgence by either party in enforcing any of the terms and conditions of the Contract or the granting of time by either party to the other shall prejudice, affect, or restrict the rights of that party under

3 Corrupt and Fraudulent Practices the Contract, neither shall any waiver by either party of any breach of Contract operate as waiver of any subsequent or continuing breach of Contract.

- b. Any waiver of a party's rights, powers, or remedies under the Contract must be in writing, dated, and signed by an authorized representative of the party granting such waiver, and must specify the right and the extent to which it is being waived.
- 4.5 Severability

If any provision or condition of the Contract is prohibited or rendered invalid or unenforceable, such prohibition, invalidity or unenforceability shall not affect the validity or enforceability of any other provisions and conditions of the Contract.

- **5 Language** 5.1 The Contract as well as all correspondence and documents relating to the Contract exchanged by the Supplier and the Purchaser, shall be written in the language specified in the SCC. Supporting documents and printed literature that are part of the Contract may be in any language provided they are accompanied by an accurate translation of the relevant passages in the language specified, in which case, for purposes of interpretation of the Contract, such translation shall govern.
  - 5.2 The Supplier shall bear all costs of translation to the governing language and all risks of the accuracy of such translation, for documents provided by the Supplier.
- 6 Joint Venture, Consortium or Association
  6.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly and severally liable to the Purchaser for the fulfilment of the provisions of the Contract and shall designate one party to act as a leader with authority to bind the joint venture, consortium, or association. The composition or the constitution of the joint venture, consortium, or association the prior consent of the Purchaser.
- **7 Eligibility** 7.1 The Bidder should not have been declared insolvent by the competent court.
  - 7.2 The Bidder should not be disqualified for contract under the law of the land.
  - 7.3 The Bidder should not be adjudged defaulter of Tax Payment under Income Tax Law or any other Law for the time being in-force.

- 7.4 The Supplier and its Subcontractors shall have the nationality of an eligible country. A Supplier or Subcontractor shall be deemed to have the nationality of a country if it is a citizen or constituted, incorporated, or registered, and operates in conformity with the provisions of the laws of that country. Nationality must be disclosed by the supplier
- 7.5 All Goods and Related Services to be supplied under the Contract shall have their origin in Eligible Countries. For the purpose of this Clause, origin means the country where the goods have been grown, mined, cultivated, produced, manufactured, or processed; or through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.
- 8 Notices 8.1 Any notice given by one party to the other pursuant to the Contract shall be in writing to the address specified in the SCC. The term "in writing" means communicated in written form with proof of receipt.
  - 8.2 A notice shall be effective when delivered or on the notice's effective date, whichever is later.
- 9 **Governing Law** 9.1 The Contract shall be governed by and interpreted in accordance with the laws of India, unless otherwise specified in the SCC.
  - 9.2 Throughout the execution of the Contract, the Contractor shall comply with the import of goods and services prohibitions in India when
    - (a) as a matter of law or official regulations, India prohibits commercial relations with that country; or
    - (b) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations, India prohibits any import of goods from that country or any payments to any country, person, or entity in that country.
  - 10.1 The Purchaser and the Supplier shall make every effort to resolve amicably by direct informal negotiation any disagreement or dispute arising between them under or in connection with the Contract.
    - 10.2 If the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Purchaser or the Supplier may give notice to the other party of its intention to settle the issue by arbitration, as hereinafter provided, as to the matter in dispute, no arbitration in respect of the matter be

10 Settlement of **Disputes** 

		10.3	commenced unless such notice is given in accordance with this Clause for the final settlement of the matter. Arbitration may be commenced prior to or after delivery of the Goods under the Contract. All questions, disputes and differences arising shall be referred by the Vice-Chancellor, Banaras Hindu University to the sole arbitrator for arbitration under the provision of the Arbitrations and Conciliation Act, 1996.
11	Obligations During Arbitrations	11.1	<ul><li>Notwithstanding any reference to arbitration in Clause 10,</li><li>(a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and</li><li>(b) the Purchaser shall pay any amount due to the Supplier.</li></ul>
12	Scope of Supply	12.1	The Goods and Related Services to be supplied shall be as specified in the Schedule of Requirements.
13	Delivery and Documents	13.1	Subject to GCC Sub-Clause 33.1, the Delivery of the Goods and Completion of the Related Services shall be in accordance with the Delivery and Completion Schedule specified in the Schedule of Requirements. The details of shipping and other documents to be furnished by the Supplier are specified in the SCC.
14	Supplier's Responsibilities	14.1	The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with GCC Clause 12, and the Delivery and Completion Schedule, as per GCC Clause 13.
15	Contract Price	15.1	Prices charged by the Supplier for the Goods supplied and the Related Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid, with the exception of any price adjustments authorized in the SCC.
16	Terms of Payment	16.1	<ul> <li>Ordinarily, payments for services rendered or supplies made shall be released only after the services have been rendered or supplies appropriate to the requirement made. However in following cases advance payments may be made if specified in SCC:</li> <li>(i) Advance payment demanded by firms holding maintenance contracts for servicing of Air-conditioners, computers, other costly equipment, etc.</li> </ul>

(ii) Advance payment demanded by firms against fabrication contracts, turnkey contracts etc.

Such advance payment should not exceed the following limits :-

- (i) Thirty percent of the contract value to private firms;
- (ii) Forty percent of the contract value to a State or central Government agency or a Public Sector Undertaking; or
- (iii) In case of maintenance contract, the amount should not exceed the amount payable for six months under the contract.
- 16.2 The Supplier's request for payment shall be made to the Purchaser in writing, accompanied by invoices describing, as appropriate, the Goods delivered and Related Services performed, and by the documents submitted pursuant to GCC Clause 13 and upon fulfilment of all other obligations stipulated in the Contract.
- 16.3 Payments shall be made promptly by the Purchaser, within ninety (90) days after submission of an invoice or request for payment by the Supplier, and after the Purchaser has accepted it.
- 16.4 The currencies in which payment shall be made to the supplier under this contract shall be Indian currency unless otherwise agreed.
- **17 Taxes and Duties** 17.1 For goods manufactured outside India, the Supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside India.
  - 17.2 For goods Manufactured within India, the Supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted Goods to the Purchaser.
  - 17.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in India, the Purchaser shall use its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent or country or origin, the supplies shall provide benefit from any such tax sowing to the purchaser.
- 18 Performance 18.1 If required as specified in the SCC, the Supplier shall, within twenty-one (21) days of the notification of contract award, provide a performance security for the performance of the Contract in the amount specified in the SCC.
  - 18.2 The proceeds of the Performance Security shall be payable to the Purchaser as compensation for any loss resulting from the

Supplier's failure to complete its obligations under the Contract.

- 18.3 As specified in the SCC, the Performance Security, if required, shall be denominated in the currency(ies) of the Contract or in a freely convertible currency acceptable to the Purchaser; and shall be in one of the format stipulated by the Purchaser in the SCC, or in another format acceptable to the Purchaser.
- 18.4 Performance security should remain valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including warranty obligation.
- 18.5 Bid security shall be refunded to the successful bidder within 30 days of receipt of performance security.
- 19 Copyright 19.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Purchaser by the Supplier herein shall remain vested in the Supplier, or, if they are furnished to the Purchaser directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.
- Confidential The Purchaser and the Supplier shall keep confidential and 20 20.1 shall not, without the written consent of the other party Information hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto in connection with the Contract, whether such information has been furnished prior to, during or following completion or termination of the Contract. Notwithstanding the above, the Supplier may furnish to its Subcontractor such documents, data, and other information it receives from the Purchaser to the extent required for the Subcontractor to perform its work under the Contract, in which event the Supplier shall obtain from such Subcontractor an undertaking of confidentiality similar to that imposed on the Supplier under GCC Clause 20.
  - 20.2 The Purchaser shall not use such documents, data, and other information received from the Supplier for any purposes unrelated to the contract. Similarly, the Supplier shall not use such documents, data, and other information received from the Purchaser for any purpose other than the performance of the Contract.
  - 20.3 The obligation of a party under GCC Sub-Clauses 20.1 and 20.2 above, however, shall not apply to information that:
    - a. now or hereafter enters the public domain through no

fault of that party;

- b. can be proven to have been possessed by that party at the time of disclosure and which was not previously obtained, directly or indirectly, from the other party; or
- c. otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.
- 20.4 The above provisions of GCC Clause 20 shall not in any way modify any undertaking of confidentiality given by either of the parties hereto prior to the date of the Contract in respect of the Supply or any part thereof.
- 20.5 The provisions of GCC Clause 20 shall survive completion or termination, for whatever reason, of the Contract.
- **21 Subcontracting** 21.1 The Supplier shall notify the Purchaser in writing of all subcontracts awarded under the Contract if not already specified in the bid. Such notification, in the original bid or later on shall not relieve the Supplier from any of its obligations, duties, responsibilities, or liability under the Contract.
  - 21.2 Subcontracts shall comply with the provisions of GCC Clauses 3 and 7.
  - 22.1 Technical Specifications and Drawings
    - (a) The Goods and Related Services supplied under this Contract shall conform to the technical specifications and standards mentioned in Section-VI, Schedule of Requirements and, when no applicable standard is mentioned, the standard shall be equivalent or superior to the official standards whose application is appropriate to the Goods' country of origin or India.
    - (b) Wherever references are made in the Contract to codes and standards in accordance with which it shall be executed, the edition or the revised version of such codes and standards shall be those specified in the Schedule of Requirements. During Contract execution, any changes in any such codes and standards shall be applied only after approval by the Purchaser and shall be treated in accordance with GCC Clause 33.
- 23 Packing and Documents23.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme

22 Specifications and Standards temperatures, salt and precipitation, and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.

- 23.2 The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified in the SCC, and in any other instructions ordered by the Purchaser.
- **24 Insurance** 24.1 Unless otherwise specified in the SCC, the Goods supplied under the Contract shall be fully insured—in a freely convertible currency from an eligible country—against loss or damage incidental to manufacture or acquisition, transportation, storage, and delivery, in accordance with the applicable Incoterms or in the manner specified in the SCC.
- 25 Transportation and Incidental Services
   25.1 The Supplier is required under the Contract to transport the Goods to a specified place of final destination within India, defined as the Project Site, transport to such place of destination in India, including insurance and storage, as shall be specified in the Contract, shall be arranged by the Supplier, and related costs shall be included in the Contract Price"; or any other agreed upon trade terms (specify the respective responsibilities of the Purchaser and the Supplier)
  - 25.2 The Supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
    - (a) performance or supervision of on-site assembly and/or start up of the supplied Goods;
    - (b) furnishing of tools required for assembly and/or maintenance of the supplied Goods;
    - (c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;
    - (d) performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
    - (e) training of the Purchaser's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
  - 25.3 Prices charged by the Supplier for incidental services, if not

included in the Contract Price for the Goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services otherwise shall be at the cost of suppliers.

- 26 Inspections and Tests26.1 The Supplier shall at its own expense and at no cost to the Purchaser carry out all such tests and/or inspections of the Goods and Related Services as are specified in the SCC.
  - 26.2 The inspections and tests may be conducted on the premises of the Supplier or its Subcontractor, at point of delivery, and/or at the Goods' final destination, or in another place in India as specified in the SCC. Subject to GCC Sub-Clause 26.3, if conducted on the premises of the Supplier or its Subcontractor, all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to the Purchaser.
  - 26.3 The Purchaser or its designated representative shall be entitled to attend the tests and/or inspections referred to in GCC Sub-Clause 26.2, provided that the Purchaser bear all of its own costs and expenses incurred in connection with such attendance including, but not limited to, all traveling and board and lodging expenses.
  - 26.4 Whenever the Supplier is ready to carry out any such test and inspection, it shall give a reasonable advance notice, including the place and time, to the Purchaser. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Purchaser or its designated representative to attend the test and/or inspection.
  - 26.5 The Purchaser may require the Supplier to carry out any test and/or inspection not required by the Contract but deemed necessary to verify that the characteristics and performance of the Goods comply with the technical specifications codes and standards under the Contract, provided that the Supplier's reasonable costs and expenses incurred in the carrying out of such test and/or inspection shall be added to the Contract Price. Further, if such test and/or inspection impedes the progress of manufacturing and/or the Supplier's performance of its other obligations under the Contract, due allowance will be made in respect of the Delivery Dates and Completion Dates and the other obligations so affected.
  - 26.6 The Supplier shall provide the Purchaser with a report of the results of any such test and/or inspection.
  - 26.7 The Purchaser may reject any Goods or any part thereof that fail to pass any test and/or inspection or do not conform to the specifications. The Supplier, if permitted by the purchaser, shall either rectify or replace such rejected Goods or parts thereof or make alterations necessary to meet the

specifications at no cost to the Purchaser, and shall repeat the test and/or inspection, at no cost to the Purchaser, upon giving a notice pursuant to GCC Sub-Clause 26.4.

- 26.8 The Supplier agrees that neither the execution of a test and/or inspection of the Goods or any part thereof, nor the attendance by the Purchaser or its representative, nor the issue of any report pursuant to GCC Sub-Clause 26.6, shall absolve the Supplier from any warranties or other obligations under the Contract.
- 27 Liquidated Damages
  27.1 Except as provided under GCC Clause 32, if the Supplier fails to deliver any or all of the Goods by the Date(s) of delivery or perform the Related Services within the period specified in the Contract, the Purchaser may without prejudice to all its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in the SCC of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the percentage specified in those SCC. Once the maximum is reached, the Purchaser may terminate the Contract pursuant to GCC Clause 33.
- **28 Warranty** 28.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
  - 28.2 Subject to Sub-Clause 22.1(b) of GCC, the Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in India.
  - 28.3 Unless otherwise specified in the SCC, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the SCC, or warranty period mentioned by supplier whichever period concludes later unless mutually agreed.
  - 28.4 The Purchaser shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof. The Purchaser shall afford all reasonable opportunity for the Supplier to inspect such defects.
  - 28.5 Upon receipt of such notice, the Supplier shall, within the period specified in the SCC, expeditiously repair or replace

the defective Goods or parts thereof, at no cost to the Purchaser.

- 28.6 If having been notified, the Supplier fails to remedy the defect within the period specified in the SCC; the Purchaser may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.
- 29 Patent Indemnity
   29.1 The Supplier shall, subject to the Purchaser's compliance with GCC Sub-Clause 29.2, indemnify and hold harmless the Purchaser and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Purchaser may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered.
  - 29.2 If any proceedings are brought or any claim is made against the Purchaser arising out of the matters referred to in GCC Sub-Clause 29.1, the Purchaser shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Purchaser's name conduct such proceedings or claim and any negotiations for the settlement of any such proceedings or claim.
  - 29.3 The Purchaser shall, at the Supplier's request, afford all available assistance to the Supplier in conducting such proceedings or claim, and shall be reimbursed by the Supplier for all reasonable expenses incurred in so doing.
- **30 Force Majeure** 30.1 For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of the Supplier. Such events may include, but not be limited to, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.
  - 30.2 If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably possible, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
  - 30.3 The Supplier shall not be liable for forfeiture of its

Performance Security, liquidated damages, or termination for default if and to the extent that it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

- 31 Change Orders and Contract Amendments
- 31.1 The Purchaser may at any time order the Supplier through notice in accordance GCC Clause 8, to make changes within the general scope of the Contract in any one or more of the following:
  - (a) drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Purchaser;
  - (b) the method of shipment or packing;
  - (c) the place of delivery; and
  - (d) the Related Services to be provided by the Supplier.
- 31.2 If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or in the Delivery/Completion Schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this Clause must be asserted within twenty-eight (28) days from the date of the Supplier's receipt of the Purchaser's change order.
- 31.3 Prices to be charged by the Supplier for any Related Services that might be needed but which were not included in the Contract shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.
- 31.4 Subject to the above, no variation in or modification of the terms of the Contract shall be made except by written amendment signed by the parties.
- 32 Extensions of Time
   32.1 If at any time during performance of the Contract, the Supplier or its subcontractors should encounter conditions impeding timely delivery of the Goods or completion of Related Services pursuant to GCC Clause 13, the Supplier shall promptly notify the Purchaser in writing of the delay, its likely duration, and its cause. As soon as practicable after receipt of the Supplier's notice, the Purchaser shall evaluate the situation and may at its discretion extend the Supplier's time for performance, in which case the extension shall be ratified by the parties by amendment of the Contract.
  - 32.2 Except in case of Force Majeure, as provided under GCC Clause 32, a delay by the Supplier in the performance of its

Delivery and Completion obligations shall render the Supplier liable to the imposition of liquidated damages pursuant to GCC Clause 26, unless an extension of time is agreed upon, pursuant to GCC Sub-Clause 34.1.

#### **33 Termination** 33.1 Termination for Default

- (a) The Purchaser, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the Supplier, may terminate the Contract in whole or in part:
  - (i) if the Supplier fails to deliver any or all of the Goods within the period specified in the Contract, or within any extension thereof granted by the Purchaser pursuant to GCC Clause 32;
  - (ii) if the Supplier fails to perform any other obligation under the Contract; or
  - (iii) if the Supplier, in the judgment of the Purchaser has engaged in fraud and corruption, as defined in GCC Clause 3, in competing for or in executing the Contract.
- (b) In the event the Purchaser terminates the Contract in whole or in part, pursuant to GCC Clause 33.1(a), the Purchaser may procure, upon such terms and in such manner as it deems appropriate, Goods or Related Services similar to those undelivered or not performed, and the Supplier shall be liable to the Purchaser for any additional costs for such similar Goods or Related Services. However, the Supplier shall continue performance of the Contract to the extent not terminated.
- 33.2 Termination for Insolvency.
  - (a) The Purchaser may at any time terminate the Contract by giving notice to the Supplier if the Supplier becomes bankrupt or otherwise insolvent. In such event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any right of action or remedy that has accrued or will accrue thereafter to the Purchaser.
- 33.3 Termination for Convenience.
  - (a) The Purchaser, by notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time for its convenience. The notice of termination shall specify that termination is for the Purchaser's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date

upon which such termination becomes effective.

- (b) The Goods that are complete and ready for shipment within twenty-eight (28) days after the Supplier's receipt of notice of termination shall be accepted by the Purchaser at the Contract terms and prices. For the remaining Goods, the Purchaser may elect:
  - (i) to have any portion completed and delivered at the Contract terms and prices; and/or
  - (ii) to cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Related Services and for materials and parts previously procured by the Supplier.
- **34** Assignment 34.1 Neither the Purchaser nor the Supplier shall assign, in whole or in part, their obligations under this Contract, except with prior written consent of the other party.

## Section VIII. Special Conditions of Contract

The following Special Conditions of Contract (SCC) shall supplement and / or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

GCC Clause Ref. No.	SCC
GCC 1.1 (i)	The Purchaser's country is: India.
GCC 1.1 (k)	The Purchaser is: The Head, Department of Neurology, Institute of Medical Sciences, BHU, Varanasi -221005
GCC 1.1 (j)	The Project Site(s)/Final Destination(s) is/are: <b>Department of Neurology</b> , <b>Institute of Medical Sciences</b> , <b>BHU</b> , <b>Varanasi -221005</b>
GCC 5.1	The language shall be: Hindi or English
GCC 8.1	For notices, the Purchaser's address shall be:
	Attention : Prof. Rameshwar Nath Chaurasia
	Street Address : Department of Neurology, Institute of
	Medical Sciences, BHU, Varanasi
	Floor/ Room number : Department of Neurology
	City : Varanasi
	ZIP Code : 221005
	Country : India
	Telephone : 91-542-230-9373
	91-9415353255
	Facsimile number :
	Electronic mail address : <u>email@gmail.com</u>
GCC 9.1	The governing law shall be the <b>law of India</b> .
GCC 10.2	The rules of procedure for arbitration proceedings pursuant to GCC Clause 10.2 shall be as follows:
	Contracts with Supplier for arbitral proceeding
	In the case of a dispute between the Purchaser and a Supplier, the dispute shall be referred to adjudication or arbitration in accordance with the laws of India by the arbitrator appointed by vice-chancellor, BHU, unless otherwise agreed.
GCC 13.1	Details of Shipping and other Documents to be furnished by the Supplier are:
	(i) An airway bill
	(ii) Insurance Certificate,

	(iii) Manufacturer's or Supplier's Warranty Certificate,
	(iv) Inspection Certificate issued by nominated inspection agency,
	(v) Supplier's factory shipping details etc.
	() STF S STF S ST
	The above documents shall be received by the Purchaser before arrival of the Goods and, if not received, the Supplier will be responsible for any consequent expenses.
GCC 15.1	The prices charged for the Goods supplied and the related Services performed <b>shall not</b> be adjustable.
GCC 16.1	GCC 16.1—The method and conditions of payment to be made to the Supplier under this Contract shall be as follows:
	Payment for Goods supplied from abroad:
	Payment of foreign currency portion shall be made in Indian Rupees in the following manner:
	<ul> <li>(i) On Shipment: Ninety (90%) percent of the Contract Price of the Goods shipped shall be paid through irrevocable confirmed letter of credit opened in favor of the Supplier in a bank in its country, upon receiving equipment in good condition and documents specified in GCC Clause 13.1 under Section-VII.</li> </ul>
	<ul> <li>(ii) On Acceptance: Remaining (10%) percent of the Contract Price of Equipment received shall be paid within thirty (30) days of receipt of the equipment upon submission of claim supported by the acceptance certificate issued by the Purchaser by bank draft/wire transfer.</li> </ul>
	Payment of local currency portion shall be made in <b>Indian Rupees</b> within <b>thirty (30) days</b> of presentation of claim supported by a certificate from the Purchaser declaring that the Goods have been delivered and that all other contracted Services have been performed.
	Payment for Goods and Services supplied from within India:
	Payment for Goods and Services supplied from within India shall be made in <b>Indian Rupees</b> , as follows:
	<ul> <li>(i) On Delivery &amp; Acceptance: Hundred (100%) percent of the Contract Price shall be paid on receipt of the Goods in good conditions and acceptance certificate for satisfactory installation and functioning.</li> </ul>
GCC 18.1	A Performance Security shall be required @ 5% of contract price.
GCC 18.3	If required, the Performance Security shall be in the form of : Bank Guarantee
	If required, the Performance security shall be denominated in Indian Rupees.

IT	7
GCC 23.2	The packing, marking and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract.
GCC 24.1	The insurance coverage shall be as specified in the <b>Incoterms</b> .
	If not in accordance with <b>Incoterms</b> , insurance shall be as follows: <b>NA</b>
GCC 25.1	Responsibility for transportation of the Goods shall be as specified in the <b>Incoterms</b> .
	If not in accordance with <b>Incoterms</b> , responsibility for transportations shall be as follows: <b>NA</b>
GCC 25.2	Incidental services to be provided are:
	Selected services covered under GCC Clause 25.2 and/or other should be specified with the desired features. The price quoted in the bid price or agreed with the selected Supplier shall be included in the Contract Price.
GCC 26.1	The inspections and tests shall be: Equipment should be installed at site by designated engineer of the firm. Demonstration should be made to the satisfaction of the P.I. At least four free visits should be made by the engineer during the first year after installation to clarify and rectify any doubts or problems as may be faced by the user.
GCC 26.2	The Inspections and tests shall be conducted at the Department where the equipment is installed. For the rest please refer to GCC 26.1.
GCC 27.1	The liquidated damage shall be: 1% per week
GCC 27.1	The maximum amount of liquidated damages shall be: 10%
GCC 28.3	<ul> <li>The period of validity of the Warranty shall be 36 months from date of acceptance/ satisfactory installation of the equipment.</li> <li>For purposes of the Warranty, the place(s) of final destination(s) shall be the Department where the equipment is installed.</li> <li>The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the Contract (if any). If, for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall, at its discretion, either:</li> <li>(a) make such changes, modifications, and/or additions to the Goods</li> </ul>
	(a) make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and

	expense and to carry out further performance tests in accordance with SCC 4, or
	(b) pay liquidated damages to the Purchaser with respect to the failure to meet the contractual guarantees. The rate of these liquidated damages shall be 1% per week of actual value of the equipment (maximum 10%).
GCC 28.5	The period for repair or replacement shall be: 60 days

# Part-4 Bidding Forms & Contract Forms

# Section IX: Bidding Forms

## **Table of Forms**

1.	Tender Form (Techno commercial un-priced Bid)	.72
2.	Tender Form (Priced Bid)	.75
3.	Bidder Information Form	.77
4.	Manufacturer's Authorization	.78

### (i) Tender Form

#### (Techno commercial un-priced Bid)

(On the letter head of the firm submitting the bid)

Гender No
Го
The

Dear Sir,

- 1. I/We have examined and have no reservations to the Bidding Documents, including Addenda issued in accordance with Instructions to Bidders;
- 2. I/We meet the eligibility requirements and have no conflict of interest ;
- 3. I/We have not been suspended nor declared ineligible in India;
- 4. I/We offer to supply in conformity with the Bidding Documents and in accordance with the Delivery Schedules specified in the Schedule of Requirements the following Goods: *[insert a brief description of the Goods and Related Services]*;
- 5. I/We offer to supply the items as listed in the schedule to this tender hereto/portion thereof as you may specify in the acceptance of Tender at the price given in the said Schedule and agree to hold this offer open for a period of 90 days from the date of opening of the tender.
- 6. I/we shall be bound by a communication of acceptance issued by you.
- 7. I/We have understood the Instruction to bidders and Conditions of Contract in the form as enclosed with the invitation to the tender and have thoroughly examined the specifications quoted in the Schedule hereto and am/are fully aware of the nature of the goods required and my/our offer is to supply the goods strictly in accordance with the specifications and requirements.
- 8. A crossed Bank Draft in favor of the Registrar, BHU for Rs. (Rupees.....only) as Earnest Money is enclosed. The Draft is drawn on ......Bank payable at Varanasi.

- 9. The following have been added to form part of this tender.
  - (a) Samples of items quoted for, as per instructions provided in the schedule of requirement.
  - (b) Schedule of requirements, quoting the make only duly signed and stamped.(without indicating price)
  - (c) Income Tax clearance certificate.
  - (d) Copy of last audited balance sheet.
  - (e) Copy of Valid Central/State sales tax registration certificate.
  - (f) Copy of relevant major purchase orders valuing mare than Rs. 20000/- executed during last two years for Govt. Depts., PSUs & Central Autonomous bodies..
  - (g) Proof of manufacturing Unit, dealership certificate/general order suppliers.
  - (h) Statement of deviations from financial terms & conditions, if any.
  - (i) Any other enclosure. (Please give details)
- 10. We undertake to execute all orders which have been placed to meet emergent requirements on priority basis.
- 11. Certified that the bidder is:
  - (a) A sole proprietorship firm and the person signing the bid document is the sole proprietor/constituted attorney of the sole proprietor,

#### Or

(b) A partnership firm, and the person signing the bid document is a partner of the firm and he has authority to refer to arbitration disputes concerning the business of the partnership by virtue of the partnership agreement/by virtue of general power of attorney.

#### Or

(c) A company and the person signing the document is the constituted attorney.

# (NOTE: Delete whatever is not applicable. All corrections/deletions should invariable be duly attested by the person authorized to sign the bid document).

- 12. We do hereby undertake that, until a formal notification of award, this bid, together with your written acceptance thereof shall constitute a binding contract between us.
- 13. If our bid is accepted, we commit to obtain a performance security in accordance with the Bidding Documents;
- 14. We are not participating, as a Bidder or as a subcontractor, in more than one bid in this bidding process, other than alternative bids submitted;
- 15. We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in any type of fraud and corruption

Name of the Bidder\* [insert complete name of person signing the Bid]

Name of the person duly authorized to sign the Bid on behalf of the Bidder\*\* [insert complete name of person duly authorized to sign the Bid]

Title of the person signing the Bid [insert complete title of the person signing the Bid]

Signature of the person named above [insert signature of person whose name and capacity are shown above]

Date signed \_ [insert date of signing] day of [insert month], [insert year]

\*: In the case of the Bid submitted by joint venture specify the name of the Joint Venture as Bidder

\*\*: Person signing the Bid shall have the power of attorney given by the Bidder to be attached with the Bid Schedules.

Yours faithfully,

(Signature of bidder)
Dated this day of \_\_\_\_\_\_
Address: .....
Telephone No. :\_\_\_\_\_
FAX\_\_\_\_\_
E-mail\_\_\_\_\_

Company seal

## (ii) Tender Form

#### (Priced Bid)

(On the letter head of the firm submitting the bid document)

To,

The \_\_\_\_\_

\_\_\_\_\_

Ref: Tender No......Dated:

Sir,

Having examined the bidding documents and having submitted the techno commercial unpriced bid for the same, we, the undersigned, hereby submit the priced bid for supply of goods and services as per the schedule of requirements and in conformity with the said bidding documents.

- 1. We hereby offer to supply the Goods/Services at the prices and rates mentioned in the enclosed schedule of requirement.
- 2. We do hereby undertake that, in the event of acceptance of our bid, the supply of Goods/Services shall be made as stipulated in the schedule of requirement and that we shall perform all the incidental services.
- 3. The prices quoted are inclusive of all charges net F.O.R University. We enclose herewith the complete Financial Bid as required by you. This includes:
  - a. Price Schedule (Bill of Quantity-BOQ).
  - b. Statement of deviations from financial terms and conditions.
- 4. We agree to abide by our offer for a period of 90 days from the date fixed for opening of the bid documents and that we shall remain bound by a communication of acceptance within that time.
- 5. We have carefully read and understood the terms and conditions of the bid document and we do hereby undertake to supply as per these terms and conditions. The Financial Deviations are only those mentioned in the statement of deviations from financial terms and conditions.
- 6. We have paid, or will pay the following commissions, gratuities, or fees with respect to the bidding process or execution of the Contract: **[insert complete name of each**

# Recipient, its full address, the reason for which each commission or gratuity was paid and the amount and currency of each such commission or gratuity]

Name of Recipient	Address	Reason	Amount

(If none has been paid or is to be paid, indicate "none.")

- 7. We understand that this bid, together with your written acceptance thereof included in your notification of award, shall constitute a binding contract between us, until a formal contract is prepared and executed; and
- 8. We understand that you are not bound to accept the lowest evaluated bid or any other bid that you may receive.

## Certified that the bidder is:

A sole proprietorship firm and the person signing the bid document is the sole proprietor/ constituted attorney of sole proprietor,

Or

A partnership firm, and the person signing the bid document is a partner of the firm and he has authority to refer to arbitration disputes concerning the business of the partnership by virtue of the partnership agreement/by virtue of general power of attorney,

## Or

A company and the person signing the bid document is the constituted attorney.

# (NOTE: Delete whatever is not applicable. All corrections/deletions should invariably be duly attested by the person authorized to sign the bid document.)

We do hereby undertake that, until a formal notification of award, this bid, together with your written acceptance thereof, shall constitute a binding contract between us.

Dated this day of
Signature of Bidder
Details of enclosures
Full Address:
Telephone No
Mobile No. :
Fax No. :
E-mail:

Company Seal

# **Bidder Information Form**

[The Bidder shall fill in this Form in accordance with the instructions indicated below. No alterations to its format shall be permitted and no substitutions shall be accepted.]

Date: [insert date (as day, month and year) of Bid Submission] ADVT. No.: [insert number of bidding process] Alternative No.: [insert identification No if this is a Bid for an alternative] Page \_\_\_\_\_ of\_ \_\_\_\_ pages

1. Bidder's Name [insert Bidder's legal name]

2. In case of JV, legal name of each member : [insert legal name of each member in JV]

3. Bidder's actual or intended country of registration: [insert actual or intended country of registration]

4. Bidder's year of registration: [insert Bidder's year of registration]

5. Bidder's Address in country of registration: [insert Bidder's legal address in country of registration]

6. Bidder's Authorized Representative Information

Name: [insert Authorized Representative's name]

Address: [insert Authorized Representative's Address]

Telephone/Fax numbers: [insert Authorized Representative's telephone/fax numbers]

Email Address: [insert Authorized Representative's email address]

7. Attached are copies of original documents of [check the box(es) of the attached original documents]

□ Articles of Incorporation (or equivalent documents of constitution or association), and/or documents of registration of the legal entity named above.

□ In case of JV, letter of intent to form JV or JV agreement.

□ In case of Government-owned enterprise or institution, documents establishing:

- Legal and financial autonomy
- Operation under commercial law
- Establishing that the Bidder is not dependent agency of the Purchaser
- 2. Included are the organizational chart, a list of Board of Directors, and the beneficial ownership.

## **Manufacturer's Authorization**

[The Bidder shall require the Manufacturer to fill in this Form in accordance with the instructions indicated. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. The Bidder shall include it in its bid, if so indicated in the **BDS**.]

Date: [insert date (as day, month and year) of Bid Submission] ADVT. No.: [insert number of bidding process] Alternative No.: [insert identification No if this is a Bid for an alternative]

To: [insert complete name of Purchaser]

## WHEREAS

We [insert complete name of Manufacturer], who are official manufacturers of [insert type of goods manufactured], having factories at [insert full address of Manufacturer's factories], do hereby authorize [insert complete name of Bidder] to submit a bid the purpose of which is to provide the following Goods, manufactured by us [insert name and or brief description of the Goods], and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with Clause 28 of the General Conditions of Contract, with respect to the Goods offered by the above firm.

Signed: [insert signature(s) of authorized representative(s) of the Manufacturer]

Name: [insert complete name(s) of authorized representative(s) of the Manufacturer]

Title: [insert title]

Dated on \_\_\_\_\_\_ day of \_\_\_\_\_\_, \_\_\_\_ [insert date of signing]

## Section X. Contract Forms

This Section contains forms which, once completed, will form part of the Contract. The forms for Performance Security and Advance Payment Security, when required, shall only be completed by the successful Bidder after contract award.

## **Table of Forms**

1.	Tender Acceptance Letter	. 80
2.	Contract Agreement	. 81
3.	Performance Security	. 83
4.	Advance Payment Security	. 87

## **Tender Acceptance Letter**

(To be given on Company Letter Head)

Date:

Sub: Acceptance of Terms & Conditions of Tender.

Tender Reference No.\_\_\_\_\_

Name of Tender/ Work: -

Dear Sir,

1. I/We have downloaded/ obtained the tender document(s) for the above mentioned 'Tender/Work' from the web site(s) namely:

as per your advertisement, given in the above mentioned website(s).

- 2. I/We hereby certify that I/We have read the entire terms and conditions of the tender documents from Page No. \_\_\_\_\_\_ to \_\_\_\_\_ (including all documents like section(s), schedules(s) etc.), which form part of the contract agreement and I/we shall abide hereby by the terms/conditions/ clauses contained therein.
- 3. The corrigendum(s) issued from time to time by your department/ organisation too have also been taken into consideration, while submitting this acceptance letter.
- 4. I/We hereby unconditionally accept the tender conditions of above mentioned tender document(s)/ corrigendum(s) in its totality/entirety.
- 5. In case any provisions of this tender are found violated, then your department/organisation shall without prejudice to any other right or remedy be at liberty to reject this tender/bid including the forfeiture of the full said earnest money deposit absolutely.

## Yours Faithfully,

(Signature of the Bidder, with Official Seal)

## **Contract Agreement**

[The successful Bidder shall fill in this form in accordance with the instructions indicated]

#### THIS AGREEMENT made

the [ insert: number ] day of [ insert: month ], [ insert: year ].

#### BETWEEN

- (1) [ insert complete name of Purchaser ], a [ insert description of type of legal entity, for example, an agency of the Ministry of .... of the Government of { insert name of Country of Purchaser }, or corporation incorporated under the laws of { insert name of Country of Purchaser } ] and having its principal place of business at [ insert address of Purchaser ] (hereinafter called "the Purchaser"), of the one part, and
- (2) [ insert name of Supplier ], a corporation incorporated under the laws of [ insert: country of Supplier ] and having its principal place of business at [ insert: address of Supplier ] (hereinafter called "the Supplier"), of the other part :

WHEREAS the Purchaser invited bids for certain Goods and ancillary services, viz., *[insert brief description of Goods and Services]* and has accepted a Bid by the Supplier for the supply of those Goods and Services

The Purchaser and the Supplier agree as follows:

- 1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Contract documents referred to.
- 2. The following documents shall be deemed to form and be read and construed as part of this Agreement. This Agreement shall prevail over all other contract documents.
  - (a) the Letter of Acceptance
  - (b) the Tender Forms
  - (c) the Addenda Nos.\_\_\_\_ (if any)
  - (d) Special Conditions of Contract
  - (e) General Conditions of Contract
  - (f) the Specification (including Schedule of Requirements and Technical Specifications)
  - (g) the completed Schedules (including Price Schedules(BOQ))
  - (h) any other document listed in GCC as forming part of the Contract

- 3. In consideration of the payments to be made by the Purchaser to the Supplier as specified in this Agreement, the Supplier hereby covenants with the Purchaser to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
- 4. The Purchaser hereby covenants to pay the Supplier in consideration of the provision of the Goods and Services and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of *[insert the name of the Contract governing law country]* on the day, month and year indicated above.

For and on behalf of the Purchaser

Signed:[insert signature]in the capacity of [ insert title or other appropriate designation ]in the presence of [insert identification of official witness]

For and on behalf of the Supplier

Signed: [insert signature of authorized representative(s) of the Supplier] in the capacity of [ insert title or other appropriate designation ] in the presence of [ insert identification of official witness]

## **Performance Security**

## **Option 1: (Bank Guarantee)**

[The bank, as requested by the successful Bidder, shall fill in this form in accordance with the instructions indicated]

[Guarantor letterhead or SWIFT identifier code]

Beneficiary: [insert name and Address of Purchaser]

**Date:** \_ [Insert date of issue]

## **PERFORMANCE GUARANTEE No.:** [Insert guarantee reference number]

Guarantor: [Insert name and address of place of issue, unless indicated in the letterhead]

We have been informed that \_ [insert name of Supplier, which in the case of a joint venture shall be the name of the joint venture] (hereinafter called "the Applicant") has entered into Contract No. [insert reference number of the contract] dated [insert date] with the Beneficiary, for the supply of \_ [insert name of contract and brief description of Goods and related Services] (hereinafter called "the Contract").

Furthermore, we understand that, according to the conditions of the Contract, a performance guarantee is required.

At the request of the Applicant, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of [insert amount in figures] (\_\_\_\_\_\_) [insert amount in words]<sup>5</sup>, such sum being payable in the types and proportions of currencies in which the Contract Price is payable, upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating that the Applicant is in breach of its obligation(s) under the Contract, without the Beneficiary needing to prove or to show grounds for your demand or the sum specified therein.

This guarantee shall expire, no later than the .... Day of .....,  $2...^6$ , and any demand for payment under it must be received by us at this office indicated above on or before that date.

<sup>&</sup>lt;sup>5</sup> The Guarantor shall insert an amount representing the percentage of the Accepted Contract Amount specified in the Letter of Acceptance, and denominated either in the currency(ies) of the Contract or a freely convertible currency acceptable to the Beneficiary.

<sup>&</sup>lt;sup>6</sup> Insert the date twenty-eight days after the expected completion date as described in GC Clause 18.4. The Purchaser should note that in the event of an extension of this date for completion of the Contract, the Purchaser would need to request an extension of this guarantee from the Guarantor. Such request must be in writing and must be made prior to the expiration date established in the guarantee. In preparing this guarantee, the Purchaser might consider adding the following text to the form, at the end of the penultimate paragraph: "The Guarantor

This guarantee is subject to the Uniform Rules for Demand Guarantees (URDG) 2010 Revision, ICC Publication No. 758, except that the supporting statement under Article 15(a) is hereby excluded.

[signature(s)]

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

agrees to a one-time extension of this guarantee for a period not to exceed [six months][one year], in response to the Beneficiary's written request for such extension, such request to be presented to the Guarantor before the expiry of the guarantee."

## **Option 2: Performance Bond**

By this Bond *[insert name of Principal]* as Principal (hereinafter called "the Supplier") and *[insert name of Surety]* as Surety (hereinafter called "the Surety"), are held and firmly bound unto *[insert name of Purchaser]* as Obligee (hereinafter called "the Supplier") in the amount of *[insert amount in words and figures]*, for the payment of which sum well and truly to be made in the types and proportions of currencies in which the Contract Price is payable, the Supplier and the Surety bind themselves, their heirs, executors, administrators, successors and assigns, jointly and severally, firmly by these presents.

WHEREAS the Contractor has entered into a written Agreement with the Purchaser dated the \_\_\_\_\_\_ day of \_\_\_\_\_\_, 20 \_\_\_\_\_, for [name of contract and brief description of Goods and related Services] in accordance with the documents, plans, specifications, and amendments thereto, which to the extent herein provided for, are by reference made part hereof and are hereinafter referred to as the Contract.

NOW, THEREFORE, the Condition of this Obligation is such that, if the Supplier shall promptly and faithfully perform the said Contract (including any amendments thereto), then this obligation shall be null and void; otherwise, it shall remain in full force and effect. Whenever the Supplier shall be, and declared by the Purchaser to be, in default under the Contract, the Purchaser having performed the Purchaser's obligations thereunder, the Surety may promptly remedy the default, or shall promptly:

- (1) complete the Contract in accordance with its terms and conditions; or
- (2) obtain a Bid or bids from qualified Bidders for submission to the Purchaser for completing the Contract in accordance with its terms and conditions, and upon determination by the Purchaser and the Surety of the lowest responsive Bidder, arrange for a Contract between such Bidder and Purchaser and make available as work progresses (even though there should be a default or a succession of defaults under the Contract or Contracts of completion arranged under this paragraph) sufficient funds to pay the cost of completion less the Balance of the Contract Price; but not exceeding, including other costs and damages for which the Surety may be liable hereunder, the amount set forth in the first paragraph hereof. The term "Balance of the Contract Price," as used in this paragraph, shall mean the total amount payable by Purchaser to Supplier under the Contract, less the amount properly paid by Purchaser to Contractor; or
- (3) pay the Purchaser the amount required by Purchaser to complete the Contract in accordance with its terms and conditions up to a total not exceeding the amount of this Bond.

The Surety shall not be liable for a greater sum than the specified penalty of this Bond.

Any suit under this Bond must be instituted before the expiration of one year from the date of the issuing of the Taking-Over Certificate.

No right of action shall accrue on this Bond to or for the use of any person or corporation other than the Purchaser named herein or the heirs, executors, administrators, successors, and assigns of the Purchaser.

In testimony whereof, the Supplier has hereunto set his hand and affixed his seal, and the Surety has caused these presents to be sealed with his corporate seal duly attested by the signature of his legal representative, this \_\_\_\_\_ day of \_\_\_\_\_20\_\_\_.

SIGNED ON	on behalf of
	in the capacity of
In the presence of	
SIGNED ON	on behalf of
By	in the capacity of

## **Advance Payment Security**

[Guarantor letterhead or SWIFT identifier code]

Beneficiary: [Insert name and Address of Purchaser]

**Date:** [Insert date of issue]

## ADVANCE PAYMENT GUARANTEE No.: [Insert guarantee reference number]

Guarantor: [Insert name and address of place of issue, unless indicated in the letterhead]

We have been informed that *[insert name of Supplier, which in the case of a joint venture shall be the name of the joint venture]* (hereinafter called "the Applicant") has entered into Contract No. *[insert reference number of the contract]* dated *[insert date]* with the Beneficiary, for the execution of *[insert name of contract and brief description of Goods and related Services]* (hereinafter called "the Contract").

Furthermore, we understand that, according to the conditions of the Contract, an advance payment in the sum *[insert amount in figures]* () *[insert amount in words]* is to be made against an advance payment guarantee.

At the request of the Applicant, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of *[insert amount in figures]* (\_\_\_\_\_\_) *[insert amount in words]<sup>1</sup>* upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating either that the Applicant:

- (a) has used the advance payment for purposes other than toward delivery of Goods;
  - or
- (b) has failed to repay the advance payment in accordance with the Contract conditions, specifying the amount which the Applicant has failed to repay.

A demand under this guarantee may be presented as from the presentation to the Guarantor of a certificate from the Beneficiary's bank stating that the advance payment referred to

<sup>&</sup>lt;sup>1</sup> The Guarantor shall insert an amount representing the amount of the advance payment and denominated either in the currency(ies) of the advance payment as specified in the Contract, or in a freely convertible currency acceptable to the Purchaser.

above has been credited to the Applicant on its account number [insert number] at [insert name and address of Applicant's bank].

The maximum amount of this guarantee shall be progressively reduced by the amount of the advance payment repaid by the Applicant as specified in copies of interim statements or payment certificates which shall be presented to us. This guarantee shall expire, at the latest, upon our receipt of a copy of the interim payment certificate indicating that ninety (90) percent of the Accepted Contract Amount, has been certified for payment, or on the *[insert day]* day of *[insert month]*, *[insert year]*, whichever is earlier. Consequently, any demand for payment under this guarantee must be received by us at this office on or before that date.

This guarantee is subject to the Uniform Rules for Demand Guarantees (URDG) 2010 Revision, ICC Publication No.758, except that the supporting statement under Article 15(a) is hereby excluded.

[signature(s)]

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

## **INVITATION FOR BIDS** <u>Notice Inviting Tender (NIT)</u>

## **BANARAS HINDU UNIVERSITY** FACULTY OF MEDICINE INSTITUTE OF MEDICAL SCIENCES VARANASI-221005

## e-Procurement Notice

## Ref: BHU/IMS/Neuro/2020-21/01

Online tenders are hereby invited in two cover system from reputed manufacturer/ authorized representative of a manufacture/whole sale dealer/importer for supply of :-

• Laboratory Equipments

Bidders can download complete set of bidding documents from e- procurement Platform <u>http://eprocure.gov.in/eprocure/app</u> from 17-02-2021 onwards. Bidders need to submit the bids online for the interested items by uploading all the required documents through <u>http://eprocure.gov.in/eprocure/app</u>.

**Last Date/ Time for receipt of bids through e-procurement is:** 20-03-2021 upto 04:00PM. (Server time). Late bids shall not be accepted.

For further details regarding Tender Notification & Specifications please visit website: <u>http://eprocure.gov.in/eprocure/app</u> and <u>www.bhu.ac.in</u>.

Published Date	17-02-2021 (04:00 PM)
Bid Document Download Start Date	17-02-2021 (04:00 PM)
Clarification Start Date	17-02-2021 (04:00 PM)
Clarification End Date	26-02-2021 (04:00 PM)
Pre bid meeting	NA
Bid Submission Start Date	01-03-2021 (04:00 PM)
Bid Document Download End Date	20-03-2021 (05:00 PM)
Bid Submission End Date	20-03-2021 (05:00 PM)
Bid Opening Date	22-03-2021 (03:00 PM)

## **CRITICAL DATE SHEET**

sd/-

Registrar

Dated: 11-02-2021