Date : 15.06.2021

BID DOCUMENT (e - Procurement)

Project Supply of Equipment for COVID-19 and Black Fungus Patients, suitable for hospital Sir Sunderlal Hospital, BHU, Varanasi



Issued On: 15-06-2021

SIR SUNDERLAL HOSPITAL INSTITUTE OF MEDICAL SCIENCES BANARAS HINDU UNIVERSITY VARANASI-221005, INDIA This document is subject to copyright.

This document may be used and reproduced for non-commercial purposes only. Any commercial use, including reselling, charging to access, redistribution, or for derivative works such as unofficial translations based on these documents is not allowed.

Checklist for Bid/Tender Submission

(The following check-list must be filled in and submitted with the bid documents)

Pre- Qualification Bid

Sr. No.	Particulars	Yes/No
1	Have you attached the techno commercial unpriced bid form duly filled in appropriately?	
2	Have you attached a copy of the last audited balance sheet of your firm	
3	Have you attached the details of the income tax clearance certificate, proof of manufacturing unit/ dealership letter/ general order suppliers and copy of Central / State sales tax registration certificate?	
4	EMD: Have you submitted EMD asked for (as specified in BDS).	
5	Have you submitted samples of all items indicated in the respective schedule of requirements at the address of tender inviting authority within due date.	
6	Have you enclosed the schedule of requirement indicating the make offered without indicating the pricing components along with the techno commercial unpriced bid?	
7	Have you submitted the bids both techno commercial unpriced and priced bid separately for each tender?	
8	Have you enclosed the statement of deviations from financial terms and conditions, if any?	
Price	Bid	
1	Have you signed and attached the priced bid form?	
2	Have you attached the schedule of requirements duly priced?	

Table of Contents

Part 1: Bidd	ling Procedures	5
Section I.	Instructions for Online Bid Submission	6
Section II.	Instructions to Bidders	9
Section III.	Bid Data Sheet (BDS)	27
Section IV.	Prequalification	31
Section V.	Policy of University against the Corrupt and Fraudulent Practices	32
Part 2: Supp	ply Requirements	34
Section VI.	Schedule of Requirements	35
	(1) List of Goods and Delivery Schedule	36
	(2) List of Related Services and Completion Schedule	37
	(3) Error! Reference source not found Error! Bookmark not de	fined.
Part 3: Con	tract	52
Section VII.	General Conditions of Contract	53
Section VIII.	Special Conditions of Contract	68

Part 4: Bidd	ing Forms & Contract Forms	72
Section IX.	Bidding Forms	73
Section X.	Contract Forms	81
NIT		91

Part-1 Bidding Procedures

Section I: Instructions for Online Bid Submission

Instructions to the Bidders to submit the bids online through the Central Public Procurement Portal for e Procurement at <u>https://eprocure.gov.in/eprocure/app</u>.

- 1. Possession of valid Digital Signature Certificate (DSC) and enrollment/registration of the contractors/bidders on the e-Procurement/e-tender portal are prerequisite for e-tendering.
- 2. Bidder should register for the enrollment in the e-Procurement site using the "<u>Online Bidder Enrollment</u>" option available on the home page. Portal enrollment is generally free of charge. During enrollment/registration, the bidders should provide only valid and true information including valid email id. All the correspondence shall be made directly with the contractors/bidders through email id as registered.
- 3. Bidder need to login to the site through their user ID/ password chosen during enrollment/registration.
- 4. Then the Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by SIFY/TCS/nCode/eMudra or any other Certifying Authority recognized by Controller of Certifying Authorities (CCA) India on eToken/SmartCard, should be registered.
- 5. The registered DSC only should be used by the bidder in the transactions and should ensure safety of the same.
- 6. Contractor/Bidder may go through the tenders published on the site and download the tender documents/schedules for the tenders.
- 7. After downloading / getting the tender document/schedules, the Bidder should go through them carefully and then submit the documents as required, otherwise bid will be rejected.
- 8. Any clarifications may be sought online through the tender site, through the contact details or during pre-bid meeting if any. Bidder should take into account the corrigendum if any published before submitting the bids online.
- 9. Bidder may log in to the site through the secured login by the user id/ password chosen during enrolment/registration and then by submitting the password of the e-Token/Smartcard to access DSC.
- 10. Bidder may select the tender in which he/she is interested in by using the search option and then move it to the 'my tenders' folder.
- 11. From my tender folder, he may select the tender to view all the details uploaded there.

- 12. It shall be deemed that the bidder has read and understood all the terms and conditions before submitting the offer. Bidder should go through the tender schedules carefully and upload the documents as asked; otherwise, the incomplete bid shall stand rejected.
- 13. Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document/schedule and ordinarily it shall be in PDF/xls/rar/jpg/dwf formats. If there is more than one document, all may be clubbed together and provided in the requested format. Bidders Bid documents may be scanned with 100 dpi with black and white option. It is advisable that each document to be uploaded through online for the tenders should be less than 2 MB. If any document is more than 2MB, it can be reduced through zip/rar and the same if permitted may be uploaded. The file size being less than 1 MB the transaction uploading time will be very fast.
- 14. The Bidders can update well in advance, the documents such as certificates, annual report details etc., under "My Space option" and these can be selected as per tender requirements and then send along with bid documents during bid submission. This will facilitate the bid submission process faster by reducing upload time of bids.
- 15. Bidder should submit the Tender Fee/ EMD as specified in the tender. The hard copy should be posted/couriered/given in person to the Tender Inviting Authority, within bid submission due date and time as indicated in the tender. Scanned copy of the instrument should be uploaded as part of the offer.
- 16. While submitting the bids online, the bidder shall read the terms and conditions and may accept the same to proceed further to submit the bid packets.
- 17. The bidder has to select the payment option as offline to pay the Tender FEE/ EMD as applicable and enter details of the instruments.
- 18. The details of the DD/any other accepted instrument, physically delivered, should tally with the details available in the scanned copy and the data entered during bid submission time, otherwise submitted bid shall not be acceptable or liable for rejection.
- 19. The bidder has to digitally sign and upload the required bid documents one by one as indicated. Very act of using DSC for downloading the bids and uploading their offers shall be deemed to be a confirmation that they have read, understood and agreed with all clauses of the bid document including General conditions of contract without any exception.
- 20. The bidder has to upload the relevant files required as indicated in the cover content. In case of any irrelevant files, the bid may be rejected.
- 21. If the price bid format is provided in a spread sheet file like BoQ_XXXX.xls, the rates offered should be entered in the allotted space only and uploaded after filling the relevant columns. The Priced-bid/BOQ template shall not be modified / replaced by the bidder; else the bid submitted is liable to be rejected for the tender.

- 22. The bidders are advised to submit the bids through online e-tendering system to the Tender Inviting Authority (TIA) well before the bid submission due date and time (as per Server System Clock). The TIA shall not be held responsible for any delay or the difficulties faced during the submission of bids online by the bidders.
- 23. After the bid submission (i.e. after Clicking "Freeze Bid Submission" in the portal), the acknowledgement number indicated by the system should be printed by the bidder and kept as a record of evidence for online submission of bid for the particular tender and also be used as entry pass to participate in the bid opening.
- 24. The time settings fixed in the server side and displayed at the top of the tender site, shall remain valid for all actions of requesting, bid submission, bid opening etc., in the e-Tender system. The bidders should follow such time during bid submission.
- 25. All the data being entered by the bidders would be encrypted using Public Key Infrastructure (PKI) encryption techniques to ensure the secrecy of the data. The data entered is not retrievable by unauthorized persons during the bid submission and until the time of bid opening by any person.
- 26. Any bid document that is uploaded to the server is subjected to symmetric encryption using a system generated symmetric key. Further this key is subjected to asymmetric encryption using buyers/bid openers' public keys. Overall, the uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 27. The confidentiality of the bids is maintained with the use of Secured Socket Layer (SSL) 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 28. The bidder should logout of the tendering system using the normal logout option available at the top right hand corner and not by selecting the (X) exit option in the browser.
- 29. For any queries regarding e-Tendering process, the bidders may contact at address as provided in the tender document. Parallely for any further queries, the bidders are advised to contact over phone: **1-800-233-7315** or send an e-mail to <u>cppp-nic@nic.in</u>.

Section II: Instructions to Bidders

Table of Clauses

A.	General	10
1.	Scope of Bid	
2.	Corrupt and Fraudulent Practices	
2. 3.	Eligible Bidders	
5.		
B.	Contents of Bidding Document	12
4.	Sections of Bidding Document	
5.	Clarification of Bidding Documents, Site Visit, Pre-Bid Meeting	
6.	Amendment of Bidding Document.	
C.	Preparation of Bids	14
7.	Cost of Bidding	14
8.	Language of Bid	14
9.	Documents Comprising the Bid	14
10.	Tender Forms(Technical and Price) and Price Schedule(BOQ)	16
	Alternative Bids	
	Bid Prices and Discounts	
	Currencies of Bid and Payment	
	Documents Establishing the Eligibility and Qualifications of the Bidder	
	Period of Validity of Bids	
	Bid Security	
10.		
D.	Submission and Opening of Bids	20
17.		
18.	Deadline for Submission of Bids	
	Late Bids	
	Withdrawal, Substitution, and Modification of Bids	
	Bid Opening	
	1 0	
E.	Evaluation and Comparison of Bids	21
22.	Confidentiality	21
23.	Clarification of Bids	21
24.	Determination of Responsiveness	22
25.	Conversion to Single Currency	22
	Margin of Preference	
	Evaluation of Bids	
28.	Comparison of Bids	23
29.	Qualification of the Bidder	
	Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids	
F.	Award of Contract	24
	Award Criteria	
32.	Purchaser's Right to Vary Quantities at Time of Award	24
33.	Notification of Award	24
34.	Signing of Contract	24
	Performance Security	

Section II. Instructions to Bidders

A. General

1.	Scope of Bid	1.1	Scope of Work :
			a. Supply of Equipment for COVID-19 & Black Fungus Patients as per Specifications.
		1.2	Throughout these Bidding Documents unless the context otherwise requires:
			a. "in writing" means communicated in written form (e.g. by mail, e-mail, fax, telex) with proof of receipt;
			b. "singular" means "plural" and vice versa; and
			c. "day" means calendar day.
2.	Corrupt and Fraudulent	2.1	The Purchaser requires compliance with its policy in regard to corrupt and fraudulent practices as set forth in Section V.
	Practices	2.2	Further in pursuance of this policy, Bidder shall permit and cause its agents (whether declared or not), sub-contractors, sub-consultants, service providers or suppliers to provide access to purchaser to all the accounts, records and other documents relating to submission of the applicant, bid submission (in case prequalified), and contract performance (in case of award), to inspect and to have them audited by auditors appointed by the purchaser.
3.	Eligible Bidders	3.1	A Bidder may be a firm, a company, a limited liability partnership (LLP), a government-owned entity or any combination of such entities in the form of a joint venture (JV) under an existing agreement or with the intent to enter into such an agreement supported by a letter of intent.
		3.2	In the case of a joint venture, all members shall be jointly and severally liable for the execution of the contract in accordance with the Contract terms. The JV shall nominate a Representative who shall have the authority to conduct all business for and on behalf of any and all the members of the JV during the bidding process and during the contract execution in the event the JV is awarded the contract.
		3.3	A Bidder shall not have a conflict of interest. Any Bidder found to have a conflict of interest shall be disqualified. A Bidder may be considered to have a conflict of interest for the purpose of this bidding process, if the Bidder:
			a. directly or indirectly controls, is controlled by or is under common control with another Bidder; or

b. receives or has received any direct or indirect subsidy

from another Bidder; or

- c. has the same legal representative as another Bidder; or
- d. has a relationship with another Bidder, directly or through common third parties, that puts it in a position to influence the bid of another Bidder, or influence the decisions of the Purchaser regarding this bidding process; or
- e. participates in more than one bid in this bidding process. Participation by a Bidder in more than one Bid shall result in the disqualification of all Bids in which such Bidder is involved. This, however does not limit the inclusion of the same subcontractor in more than one bid; or
- f. any of its affiliates participated as a consultant in the preparation of the design or technical specifications of the works that are the subject of the bid; or
- g. any of its affiliates has been hired (or is proposed to be hired) by the Purchaser for the contract implementation; or
- h. would be providing goods, works, or non-consulting services resulting from or directly related to consulting services for the preparation or implementation of the project that it provided or was provided by any affiliate that directly or indirectly controls, is controlled by, or is under common control with that firm; or
- i. has a close business or family relationship with a professional staff of the Purchaser (or of the project implementing agency, or of a recipient of a part of the loan) who:
 - (i) are directly or indirectly involved in the preparation of the bidding documents or specifications of the contract, and/or the bid evaluation process of such contract; or
 - (ii) would be involved in the implementation or supervision of such contract.

3.4 A foreign firm and individual may be ineligible if as a matter of law or regulations, India prohibits commercial relations with the country of bidder.

3.5 A Bidder shall provide such evidence of eligibility satisfactory to the Purchaser, as the Purchaser shall reasonably request.

B. Contents of Bidding Document

 4 Sections of Bidding Document
 4.1 The Bidding Documents consist of Parts 1, 2, 3 and 4, which include all the Sections indicated below, and should be read in conjunction with any Addenda if any, issued.

Part 1: Bidding Procedures

Section I.	Instructions for Online Bid Submission
Section II.	Instructions to Bidders
Section III.	Bid Data Sheet (BDS)
Section IV.	Prequalification
Section V.	Policy of University against the Corrupt and fraudulent Practices

Part 2: Supply Requirements

Section VI. Schedule of Requirements

Part 3: Contract

Section VII.	General Conditions of Contract
Section VIII.	Special Conditions of Contract

Part 4: Bidding and Contract Forms

Section IX.Bidding FormsSection X.Contract Forms

- 4.2 The Invitation for Bids issued by the Purchaser is not part of the Bidding Document.
- 4.3 Unless obtained directly from the Purchaser, the Purchaser is not responsible for the completeness of the document, responses to requests for clarification, the Minutes of the pre-Bid meeting (if any), or Addenda to the Bidding Document. In case of any contradiction, documents obtained directly from the Purchaser shall prevail.
- 4.4 The Bidder is expected to examine all instructions, forms, terms, and specifications in the Bidding Documents and to furnish with its Bid all information or documentation as required by the Bidding Documents.
- 5 Clarification of 5.1 A Bidder requiring any clarification of the Bidding Document shall contact the Purchaser in writing at the Purchaser's

Documents, Site Visit, Pre-Bid Meeting address specified in the BDS or raise its enquiries during the pre-bid meeting if provided. The Purchaser will respond in writing to any request for clarification, provided that such request is received prior to the deadline for submission of bids within a period specified in the BDS. The Purchaser shall forward copies of its response to all Bidders who have acquired the Bidding Documents, including a description of the inquiry but without identifying its source. If so specified in the BDS, the Purchaser shall also promptly publish its response at the web page identified in the BDS. Should the clarification results in changes to the essential elements of the Bidding Documents, the Purchaser shall amend the Bidding Documents following the due procedure.

- 5.2 If so specified in the BDS, the Bidder is advised to visit and examine the project site and obtain for itself on its own responsibility all information that may be necessary for preparing the bid and entering into a contract for procurement of Goods. The costs of visiting shall be at the Bidder's own expense.
- 5.3 The Bidder and any of its personnel or agents shall be granted permission by the Purchaser to enter upon its premises and lands upon the express condition that the Bidder, its personnel, and agents shall indemnify the Purchaser against all liability in respect thereof.
- 5.4 If so specified in the BDS, the Bidder's designated representative is invited to attend a pre-bid meeting. The purpose of the meeting will be to clarify issues and to answer questions on any matter that may be raised at that stage.
- 5.5 The Bidder is advised to submit any questions in writing to reach the Purchaser not beyond one week preceding the meeting.
- 5.6 Minutes of the pre-bid meeting, if applicable, including the text of the questions asked by Bidders, without identifying the source, and the responses given, together with any responses prepared after the meeting, will be transmitted promptly to all Bidders who have acquired the Bidding Documents. Any modification to the Bidding Documents that may become necessary as a result of the pre-bid meeting shall be made by the Purchaser exclusively through the issue of an addendum and not through the minutes of the pre-bid meeting. Absence in the pre-bid meeting shall not be a cause for disqualification of a Bidder.
- 6 Amendment of Bidding Document
 6.1 At any time prior to the deadline for submission of bids, the Purchaser may, for any reason, whether at its own initiative or in response to a clarification requested by a prospective bidder, modify the bidding documents by corrigendum. In case

of e-procurement, corrigendum / amendment shall be published on <u>http://eprocure.gov.in/eprocure/app</u>.

- 6.2 Any addendum issued shall be part of the Bidding Documents and shall be communicated in writing to all who have obtained the Bidding Documents from the Purchaser. The Purchaser shall also promptly publish the addendum on the Purchaser's web page.
- 6.3 The Purchaser may, at its discretion to give prospective Bidders reasonable time in which to take an addendum into account in preparing their bids, extend the deadline for the submission of bids.

C. Preparation of Bids

- 7 **Cost of Bidding** 7.1 The Bidder shall bear all costs associated with the preparation and submission of its bid, and the Purchaser shall not be responsible or liable for those costs, regardless of the conduct or outcome of the bidding process.
- 8 Language of Bid 8.1 The Bid, as well as all correspondence and documents relating to the bid exchanged by the Bidder and the Purchaser, shall be written in the language specified in the BDS. Supporting documents and printed literature that are part of the Bid may be in another language provided they are accompanied by an accurate translation of the relevant passages into the language specified in the BDS, in which case, for purposes of interpretation of the Bid, such translation shall govern.
- 9 Documents 9.1 The tender/Bid shall be submitted online in two part, viz., Technical Bid and Commercial Bid.
 9 Bid

9.1.1 TECHNICAL BID

The following documents are to be scanned and uploaded as part of the Technical Bid as per the tender document:

- (a) Scanned copy of Tender Forms (Techno Commercial Un-Priced Bid) and Tender Acceptance Letter);
- (b) Scanned copy of the completed schedules,
- (c) Scanned copy of Bid Security or copy of proof for submission of Tender Document Fee/ Earnest Money Deposit etc.;
- (d) Scanned copy of written confirmation authorizing the signatory of the Bid to commit the Bidder;
 - 14

- (e) Scanned copy of documentary evidence
 - (i) establishing the Bidder's qualifications to perform the contract if its bid is accepted and
 - (ii) the Bidder's eligibility to bid;
- (f) Scanned copy of
 - (i) documentary evidence, that the Goods and Related Services to be supplied by the Bidder are of eligible origin and
 - (ii) conform to the Bidding Documents, and
 - (iii) any other document required in the BDS;
- (g) Scanned copy of Pre-qualification Details as per Section-IV like PAN/TIN/Sales Tax / Service Tax etc.
- (h) Technical Bid.

All the original documents as well as the original payment instrument like Demand Draft/Bank Guarantee /Pay order or banker cheque of any scheduled bank against Tender Fee/EMD, samples as specified in this tender document have to be sent to the address of the Purchaser mentioned in Bid Data Sheet (BDS) by post/speed post/courier/by hand on or before bid Submission closing date & time. Beyond that the tender shall be summarily rejected without assigning any reason.

9.1.2 COMMERCIAL BID

The commercial bid comprises of :

- (i) Scanned copy of Tender Form (Price Bid)
- (ii) Price bid in the form of BoQ_XXXX.xls.
- (iii) Scanned copy of item wise breakup of price bid.

The Price bid format is provided as BoQ_XXXX.xls along with this Tender Document at <u>http://eprocure.gov.in/eprocure/app</u>. Bidders are advised to download this BoQ_XXXX.xls and quote their offer/rates in the prescribed column. Bidders can quote Basic Price in INR or CURRENCY (for other than INR) but it is mandatory to quote taxes/levies in INR only, in the prescribed column and upload the same in the commercial bid.

9.2 In addition to the above requirements, bids submitted by a JV shall include a copy of the Joint Venture Agreement entered into by all members. Alternatively, a letter of intent to execute a Joint Venture Agreement in the event of a successful bid shall be signed by all members and submitted with the bid, together with a copy of the proposed Agreement, there to.

- 9.3 The Bidder shall furnish in the Tender Forms information on commissions and gratuities, if any, paid or to be paid to agents or any other party relating to this Bid.
- 10 Tender Forms(Technical and Price) and Price Schedule(BOQ)
- 11 Alternative Bids
- 12 Bid Prices and Discounts
- 10.1 Tender Forms and Price Schedules (Bill of Quantity-BOQ) shall be prepared using the relevant forms furnished in Section IX, Bidding Forms and BOQ provided. The forms must be completed without any alterations to the text, and no substitutes shall be accepted. All blank spaces shall be filled in with the information requested.
- ids 11.1 Unless otherwise specified in the BDS, alternative bids shall not be considered.
- 1 12.1 The prices and discounts quoted by the Bidder in the Tender Forms and in the Price Schedules (BOQ) shall conform to the requirements specified as under.
 - (a) All lots (contracts) and items must be listed and priced separately in the Price Schedules (BOQ).
 - (b) The price to be quoted in the Tender Forms shall be the total price of the bid, excluding any discounts offered.
 - (c) The Bidder shall quote any discount and indicate the methodology for their application in the Tender Forms.
 - (d) Prices quoted by the Bidder shall be fixed during the Bidder's performance of the Contract and not subject to variation on any account, unless otherwise specified in the BDS A bid submitted with an adjustable price quotation shall be treated as nonresponsive and shall be rejected. However, if in accordance with the BDS, prices quoted by the Bidder shall be subject to adjustment during the performance of the Contract, a bid submitted with a fixed price quotation shall not be rejected, but the price adjustment shall be treated as zero.
 - 12.2 If so bids are being invited for individual lots (contracts) or for any combination of lots (packages). Unless otherwise specified in the BDS, prices quoted shall correspond to 100 % of the items specified for each lot and to 100% of the quantities specified for each item of a lot. Bidders wishing to offer discounts for the award of more than one Contract shall specify in their bid the price reductions applicable to each package, or alternatively, to individual Contracts within the package. Discounts shall be submitted provided the bids for all lots (contracts) are opened at the same time.
 - 12.3 Prices shall be quoted as specified in each Price Schedule (BOQ) as provided. The dis-aggregation of price components is required solely for the purpose of facilitating

the comparison of bids by the Purchaser. This shall not in any way limit the Purchaser's right to contract on any of the terms offered. In quoting prices, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible country. Prices shall be entered in the following manner:

- (a) For Goods manufactured in India:
 - (i) the price of the Goods quoted EXW (ex-works, ex-factory, ex warehouse, ex showroom, or offthe-shelf, as applicable), including all customs duties and sales and other taxes already paid or payable on the components and raw material used in the manufacture or assembly of the Goods;
 - (ii) any sales tax/VAT and other taxes payable on the Goods, if the contract is awarded to the Bidder; and
 - (iii) the price for inland transportation, insurance, and other local services required to convey the Goods to their final destination (Project Site) as specified in the BDS.
- (b) For Goods manufactured outside India, to be imported:
 - (i) the price of the Goods quoted under Carriage and Insurance Paid (CIP) Model upto named place of destination in India as specified in the BDS;
 - (ii) the price for inland transportation, insurance, and other local services required to convey the Goods from the named place of destination to their final destination (Project Site) specified in the BDS;
- (c) For Goods manufactured outside India, already imported:
 - (i) the price of the Goods, including the original import value of the Goods; plus any mark-up (or rebate); plus any other related local cost, and custom duties and other import taxes already paid or to be paid on the Goods already imported.
 - (ii) the custom duties and other import taxes already paid (need to be supported with documentary evidence) or to be paid on the Goods already imported;
 - (iii) the price of the Goods, obtained as the difference between (i) and (ii) above;
 - (iv) any sales and other taxes which will be payable on the Goods if the contract is awarded to the Bidder; and
 - (v) the price for inland transportation, insurance, and other local services required to convey the Goods

from the named place of destination to their final destination (Project Site) specified in the BDS.

- (d) for Related Services, other than inland transportation and other services required to convey the Goods to their final destination, whenever such Related Services are specified in the Schedule of Requirements:
 - (i) the price of each item comprising the Related Services (inclusive of any applicable taxes).
- 13 Currencies of Bid and Payment
 13.1 The currency(ies) of the bid and the currency(ies) of payments shall be as specified in the BDS. The Bidder shall quote in Indian Rupees the portion of the bid price that corresponds to expenditures incurred in Indian Rupees, unless otherwise specified in the BDS.
 - 14.1 To establish Bidder's their eligibility, Bidders shall complete the Tender Form (Techno Commercial Un-Priced Bid & Priced Bid), included in Section-IX, Bidding Forms.
 - 14.2 The documentary evidence of the Bidder's qualifications to perform the contract if its bid is accepted shall establish to the Purchaser's satisfaction:
 - (a) that, if required in the BDS, a Bidder that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Section IX, Bidding Forms to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods in India;
 - (b) that, if required in the BDS, in case of a Bidder not doing business within India, the Bidder is or will be (if awarded the contract) represented by an Agent in the country equipped and able to carry out the Supplier's maintenance, repair and spare parts-stocking obligations prescribed in the Conditions of Contract and/or Technical Specifications; and
 - (c) that, the Bidder meets each of the qualification criterion.
 - 15.1 Bids shall remain valid for the period specified in the BDS after the bid submission deadline date prescribed by the Purchaser. A bid valid for a shorter period shall be rejected by the Purchaser as non-responsive.
 - 15.2 In exceptional circumstances, the Purchaser may solicit the Bidder's consent to an extension of the period of validity. The request and the responses thereto shall be made in writing. A Bidder may refuse the request without forfeiting its Earnest Money Deposit (EMD). A Bidder acceding to the request will neither be required nor permitted to modify

14 Documents Establishing the Eligibility and Qualifications of the Bidder

15 Period of Validity 1 of Bids the bid.

- 15.3 Bid evaluation will be based on the bid prices without taking into consideration the above modifications.
- 16 Bid Security
- 16.1 The Bidder shall furnish as part of its bid, a bid security, as specified in the BDS, in original form the amount and currency as specified in the BDS.
 - 16.2 If a bid security is specified, the bid security shall be a demand guarantee in any of the following forms at the Bidder's option:
 - (a) an unconditional guarantee issued by a bank or financial institution (such as an insurance, bonding or surety company);
 - (b) an irrevocable letter of credit;
 - (c) a banker's cheque or Demand Draft; or
 - (d) any other security as specified in the BDS,

of a reputed source from an eligible country. If the unconditional guarantee is issued by a financial institution located outside India, the issuing financial institution shall have a correspondent financial institution located in India to make it enforceable The bid security shall be valid for forty five (45) days beyond the original validity period of the bid, or beyond the extended period.

- 16.3 If a Bid Security is specified, any bid not accompanied by a substantially responsive Bid Security, shall be rejected by the Purchaser as non-responsive.
- 16.4 If a Bid Security is specified, the Bid Security of unsuccessful Bidders shall be returned as promptly as possible upon the successful Bidder's signing the contract and furnishing the Performance Security.
- 16.5 The Bid Security of the successful Bidder shall be returned as promptly as possible once the successful Bidder has signed the contract and furnished the required performance security.
- 16.6 The Bid Security of the bidder may be forfeited or the Bid Securing Declaration executed:
 - (a) if he withdraws from the bid during the period of bid validity specified by the Bidder on the Tender Forms, or any extension thereto provided by the Bidder ; or
 - (b) if he being successful Bidder fails to:
 - (i) sign the Contract; or
 - (ii) furnish a performance security.
- 16.7 The bid security of a JV must be in the name of the JV that submits the bid. If the JV has not been legally constituted into a legally enforceable entity at the time of bidding, the

bid security shall be in the names of all members as named in the letter of intent.

D. Submission and Opening of Bids

- 17 Sealing and Marking of Bids
 17.1 The Bidder shall submit the bids electronically, through the e-procurement system (<u>http://eprocure.gov.in/eprocure/app</u>). Any document submitted through any other means will not be considered as part of the Bid except for the Originals as asked for in this tender.
 18 Deadline for
 18.1 The Purchaser may, at its discretion, extend the deadline for
- 18 Deadline for Submission of Bids
 18.1
 The Purchaser may, at its discretion, extend the deadline for the submission of bids by amending the Bidding Documents, in which case all rights and obligations of the Purchaser and Bidders previously subject to the deadline shall thereafter be subject to the deadline as extended.
- 19 Late Bids
 19.1 The e-Procurement system would not allow any late submission of bids after due date and time as per server system. After electronic online proposal submission, the system generates a unique identification number which is time stamped. This shall be treated as acknowledgement of the proposal submission
- 20 Withdrawal, Substitution, and Modification of Bids
 20.1 A Bidder may withdraw, substitute, or modify its bid on the e-procurement system before the date and time specified but not beyond.
 - 20.2 No bid may be withdrawn, substituted, or modified in the interval between the deadline for submission of bids and the expiration of the period of bid validity specified by the Bidder on the Tender Forms or any extension thereof Modification/Withdrawal of the Bid sent through any other means shall not be considered by the Purchaser.
- 21 Bid Opening 21.1 The Purchaser shall open the bids as per electronic bid opening procedures specified in Central Public Procurement Portal (CPPP) at the date and time specified. Bidders can also view the bid opening by logging on to the e-procurement system. Specific bid opening procedures are laid down at http://eprocure.gov.in/eprocure/app under the head "Bidders Manual Kit". The tenderer/bidder will be at liberty to be present either in person or through an authorized representative at the time of opening of the Bid or they can view the bid opening event online at their remote end. Price Bids of only those tenderers shall be opened whose technical bids qualify.
 - 21.2 The withdrawn bid will be available in the system therefore will be considered, If bidder once withdraws the bid then he will not be able to participate in the respective tender again.

Modification to the bid shall be opened and read out with the corresponding bid. Only bids that are opened and read out at bid opening shall be considered further.

21.3 The Purchaser shall prepare a record of the bid opening that shall include; the name of the Bidder; whether there is a withdrawal, substitution, or modification; the Bid Price including any discounts and alternative bids; and the presence or absence of a bid security, if one was required. The Bidders' representatives who are present in the office of the Purchaser to witness the bid opening shall be requested to sign the record. The omission/refusal of a Bidder's signature on the record shall not invalidate the contents and effect of the record. A copy of the record shall be made available on the e-procurement system.

E. Evaluation and Comparison of Bids

- 22 Confidentiality 22.1 Information relating to the evaluation of bids and recommendation of contract award shall not be disclosed to bidders or any other persons not officially concerned with the bidding process until information on Contract Award is communication to all Bidders.
 - 22.2 No Bidder shall contact the purchaser on any matter relating to its bid from the time of the bid opening to the time the contract is awarded. If the Bidder wishes to bring additional information to the notice of the Purchaser it should be done in writing.
 - 22.3 Any effort by a Bidder to influence the purchaser in its decisions on bid evaluation, bid comparison or contract award decisions may result in rejection of the Bidder's bid.
- 23 Clarification of Bids
 23.1 To assist in the examination, evaluation, comparison of the bids, and qualification of the Bidders, the Purchaser may, at its discretion, ask any Bidder for a clarification of its Bid. Any clarification submitted by a Bidder in respect to its Bid and that is not in response to a request by the Purchaser shall not be considered. The Purchaser's request for clarification and the response shall be in writing. No change, including any voluntary increase or decrease, in the prices or substance of the Bid shall be sought, offered, or permitted, except to confirm the correction of arithmetic errors discovered by the Purchaser in the Evaluation of the bids.
 - 23.2 If a Bidder does not provide clarifications of its bid by the date and time set in the Purchaser's request for clarification, its bid may be rejected.

- 24 Determination of Responsiveness
- 24.1 The Purchaser's determination of a bid's responsiveness is to be based on the contents of the bid itself.
- 24.2 A substantially responsive Bid is one that meets the requirements of the Bidding Documents without material deviation, reservation, or omission. A material deviation, reservation, or omission is one that:
 - (a) if accepted, would
 - (i) affect in any substantial way the scope, quality, or performance of the Goods and Related Services specified in the Contract; or
 - (ii) limit in any substantial way, inconsistent with the Bidding Documents, the Purchaser's rights or the Bidder's obligations under the Contract; or
 - (b) if rectified, would unfairly affect the competitive position of other bidders presenting substantially responsive bids.
- 24.3 The Purchaser shall examine the technical aspects of the bid submitted in accordance with instructions specified in tender document, in particular, to confirm that all requirements enumerated in the 'Schedule of Requirements' Section-VI have been complied with, without any material deviation or reservation or omission.
- 24.4 If a bid is not responsive to the requirements of Bidding Documents, it shall be rejected by the Purchaser and may not subsequently be made responsive by correction of the material deviation, reservation or omission.
- 25 Conversion to Single Currency25.1 For evaluation and comparison purposes, the currency(ies) of the Bid shall be converted in a single currency as specified in the BDS.
- 26 Margin of Preference
- 26.1 Unless otherwise specified in the BDS, a margin of preference shall not apply.
- 27 Evaluation of Bids 27.1 The Purchaser shall use the criteria and methodologies listed in this Clause. No other evaluation criteria or methodologies shall be permitted.
 - 27.2 To evaluate a Bid, the Purchaser shall consider the following:
 - (a) evaluation will be done for Items or Lots (contracts), as specified in the BDS; and the Bid Price
 - (b) price adjustment due to discounts offered;
 - (c) converting the amount resulting from above, if relevant, to a single currency;
 - (d) price adjustment due to quantifiable nonmaterial nonconformities in;

- 27.3 The estimated effect of the price adjustment provisions of the Conditions of Contract, applied over the period of execution of the Contract, shall not be taken into account in bid evaluation.
- 27.4 The Purchaser's evaluation of a bid shall exclude and not take into account:
 - (a) in the case of Goods manufactured in the India, sales and other similar taxes, which will be payable on the goods if a contract is awarded to the Bidder;
 - (b) in the case of Goods manufactured outside India, already imported or to be imported, customs duties and other import taxes levied on the imported Good, sales and other similar taxes, which will be payable on the Goods if the contract is awarded to the Bidder;
 - (c) any allowance for price adjustment during the period of execution of the contract, if provided in the bid.
- 27.5 The Purchaser's evaluation of a bid may require the consideration of other factors, in addition to the Bid Price quoted. These factors may be related to the characteristics, performance, and terms and conditions of purchase of the Goods and Related Services. The effect of the factors selected, if any, shall be expressed in monetary terms to facilitate comparison of bids.
- 28 Comparison of Bids
 28.1 The Purchaser shall compare the evaluated prices of all substantially responsive bids established to determine the lowest evaluated bid. The comparison shall be on the basis of CIP-Carriage and Insurance Paid to (place of destination) prices for imported goods and EXW Ex Works (named place of delivery) prices, plus cost of inland transportation and insurance to place of destination, for goods manufactured within India, together with prices for any required installation, training, commissioning and other services. The evaluation of prices shall not take into account custom duties and other taxes levied on imported goods quoted CIP and sales and similar taxes levied in connection with the sale or delivery of goods.
- 29 Qualification of the Bidder29.1 The Purchaser shall determine to its satisfaction whether the Bidder that is selected as having submitted the lowest evaluated and substantially responsive bid meets the qualifying criteria.
 - 29.2 The determination shall be based upon an examination of the documentary evidence of the Bidder's qualifications submitted by the Bidder.
 - 29.3 An affirmative determination shall be a prerequisite for award of the Contract to the Bidder. A negative determination shall result in disqualification of the bid, in

which event the Purchaser shall proceed to the next lowest evaluated bid to make a similar determination of that Bidder's qualifications to perform satisfactorily.

30 Purchaser's Right to Accept Any Bid, and to Reject Any or All Bids
 30.1 The Purchaser reserves the right to accept or reject any bid, and to annul the bidding process and reject all bids at any time prior to contract award, without thereby incurring any liability to Bidders. In case of annulment, all bids submitted and specifically, bid securities, shall be promptly returned to the Bidders.

F. Award of Contract

- **31** Award Criteria 31.1 The Purchaser shall award the Contract to the Bidder whose bid has been determined to be the lowest evaluated bid and is substantially responsive to the Bidding Documents, provided the Bidder is determined to be qualified to perform the Contract satisfactorily.
- 32 Purchaser's Right to Vary Quantities at Time of Award
 32.1 At the time the Contract is awarded, the Purchaser reserves the right to increase or decrease the quantity of Goods and Related Services originally specified in Section VI, Schedule of Requirements, provided this does not exceed the percentages specified in the BDS, and without any change in the unit prices or other terms and conditions of the bid and the Bidding Documents.
- 33 Notification of Award
 33.1 Prior to the expiration of the period of bid validity, the Purchaser shall, notify the successful Bidder, in writing, that its Bid has been accepted. The notification letter (hereinafter and in the Conditions of Contract and Contract Forms called the "Letter of Acceptance") shall specify the sum that the Purchaser will pay the Supplier in consideration of the supply of Goods (hereinafter and in the Conditions of Contract Forms called "the Contract Price"). At the same time, the Purchaser shall also notify all other Bidders of the results of the bidding.
 - 33.2 Until a formal Contract is prepared and executed, the notification of award shall constitute a binding Contract.
 - 33.3 The Purchaser shall promptly respond in writing to any unsuccessful Bidder who, after notification of award, requests in writing the grounds on which its bid was not selected.
 - 34.1 Promptly after notification, the Purchaser shall send the successful Bidder the Contract Agreement.
 - 34.2 Within twenty-eight (28) days of receipt of the Contract Agreement, the successful Bidder shall sign, date, and return it to the Purchaser.

34.3 Notwithstanding anything contained in clause 34.2, in case signing of the Contract Agreement is prevented by any export restrictions attributable to the Purchaser, or to the use of the products/goods, systems or services to be supplied, where such export restrictions arise from trade regulations from a country supplying those products/goods, systems or services, the Bidder shall not be bound by its bid, always provided however, that the Bidder can demonstrate to the satisfaction of the Purchaser that signing of the Contact Agreement has not been prevented by any lack of diligence on the part of the Bidder in completing formalities, including applying for permits, any authorizations and licenses necessary for the export of the products/goods, systems or services under the terms of the Contract.

- 35 Performance Security
 35.1 Within twenty eight (28) days of the receipt of notification of award from the Purchaser, the successful Bidder, if required, shall furnish the Performance Security in accordance with the General Conditions of Contract (GCC), using for that purpose the Performance Security Form included in Section-X, Contract Forms, or another Form acceptable to the Purchaser. If the Performance Security furnished by the successful Bidder is in the form of a bond, it shall be issued by a bonding or insurance company that has been determined by the successful Bidder to be acceptable to the Purchaser. A foreign institution providing a bond shall have a correspondent financial institution located in India.
 - 35.2 Failure of the successful Bidder to submit the abovementioned Performance Security or sign the Contract shall constitute sufficient grounds for the annulment of the award and forfeiture of the Bid Security. In that event the Purchaser may award the Contract to the next lowest evaluated Bidder, whose bid is substantially responsive and is determined by the Purchaser to be qualified to perform the Contract satisfactorily.

Section III. Bid Data Sheet (BDS)

The following specific data for the goods to be procured shall complement, supplement, or amend the provisions in the Instructions to Bidders (ITB). In case of inconsistency, the provisions herein shall prevail over those in ITB.

Serial No.	A. General	
1	The reference number of the Invitation for Bids is	
	BHU/SSH/2021-2022/08	Dated: 14-06-2021
2	The Purchaser is Medical Superintendent, Sir Sunderlal Hospital,	
	Institute of Medical Sciences, BHU, Varan	asi-221005
3	Maximum number of members in the JV shall be: 4	

Serial No.	B. Contents of Bidding Documents		
4	For Clarification of bid purposes only, the Purchaser's address is Medical Superintendent, Sir Sunderlal Hospital, IMS, BHU, Varanasi		
	Attention	: Prof. Kailash Kumar	
	Address	: Medical Superintendent Office, Sir Sunder lal	
		Hospital, IMS, BHU, Varanasi	
	Floor/ Room number	: -	
	City	: Varanasi	
	ZIP Code/Postal Code	: 221005	
	Country	: India	
	Telephone	: 91-542-2309200	
		91-542-2369033	
	Electronic mail address	: ms.ssh.bhu@gmail.com	
5	Web page	: <u>http://eprocure.gov.in/eprocure/app</u>	
6	A site visit shall not be a	organized by the purchaser.	
7	A Pre-Bid meeting	: NA	

S No	C. Preparation of Bids				
8	The language of the bid is: English or Hindi.				
	All correspondence exchange shall be in	n English or Hindi language.			
	Language for translation of supporting doc	uments and printed literature is English or Hindi.			
9	The Bidder shall submit the following a	additional documents in its bid: NA			
10	Alternative Bids shall not be considere	d.			
11	The prices quoted by the Bidder sl performance of the Contract.	nall not be subject to adjustment during the			
12	Place of Destination: Sir Sunderlal He	ospital, IMS, BHU, Varanasi			
13	Final destination (Project Site): Sir Su	nderlal Hospital, IMS, BHU, Varanasi			
14	The prices shall be quoted by the bidder in : Foreign Currency of Principal's Country (Preferably in Indian Rupees) The Bidder is required to quote in Indian Rupees (INR), the portion of the bid price that corresponds to expenditures incurred in Indian Rupees (INR).				
15	Manufacturer's authorization is Require	ed			
16	After sales service is Required.				
17	The bid validity period shall be 120 day	ys.			
18	EMD/Bid security shall be paid @ 2% of the estimated value(s) of quoted items by the way of Demand Draft (DD)/Bank Guarantee (BG)/RTGS/NTFS in favor of the Registrar, Banaras Hindu University, Varanasi-221005 and should be valid for a period of 45 days beyond the BID validity period. All tenders received without EMD/Bank Security shall be rejected.				
	BANK DETAILS:- Bank Account No.	10654904225 (Revenue A/c- BHU)			
	Bank Name (Bank name where the account is open) State Bank Of India				
	Branch Address BHU BRANCH. Varanasi-221005				
	IFSC Code (IFSC Code of the SBIN0000211 branch)				
	MICR Code (MICR code of the 221002003 Branch)				
	Beneficiary Name	Registrar, Banaras Hindu University			
	Address :	Central Office B.H.U. Varanasi- 221005			
	TENDER FEE : NIL				
19	Other types of acceptable securities: NA				

Serial No.	D Submission and Opening of Rids		
20		purposes only, the Purchaser's address is Medical Sunderlal Hospital, Institute of Medical Sciences, BHU,	
	Attention	: Prof. Kailash Kumar	
	Street Address	: Medical Superintendent,	
		Sir Sunderlal Hospital, IMS, BHU, Varanasi	
	ZIP/Postal Code	: 221005	
	Country	: India	
	The deadline for bid su	bmission is:	
	Date	: 09 JUL 2021	
	Time	: 05:00 PM	
21	The bid opening shall t	ake place at: Medical Superintendent Office, Sir Sunderlal Hospital, IMS, BHU, Varanasi	
	Street Address	: Banaras Hindu University, Varanasi	
	Floor/ Room number	•	
		: Varanasi	
	Country	: India	
	Date	: 12, JUL,2021	
	Time	: 03:00 PM	
	The electronic bidding for Online Bid Submiss	opening procedures shall be as given in Section I-Instructions sion.	

Serial No.	E. Evaluation and Comparison of Bids
22	The currency that shall be used for bid evaluation and comparison purposes to convert all bid prices expressed in various currencies into a single currency is:
	Indian Rupees
	The source of exchange rate shall be: Reserve Bank of India.
	The date for the exchange rate shall be: Last day for submission of Bids.
23	A margin of domestic preference shall apply.
24	Evaluation will be done for concern equipment.
	Note: Bids will be evaluated for each item and the Contract will comprise the item(s) awarded to the successful Bidder.

Serial No.	F. Award of Contract					
25	The maximum percentage by which quantities may be increased is: NA					
	The maximum percentage by which quantities may be decreased is: NA					

Section IV. Prequalification

- 1. A notarized affidavit by the firm that it has never been black-listed must be attached along with the Bid, failing which the Bid shall be rejected.
- Profile of each Bidder and past experience in supply of the material (certificates to be enclosed), proof of manufacturing Unit/Dealership letter and general order supplier. Manufacturer's authorization certificate as prescribed in Section IX in case bidder is not manufacturer.
- **3.** List of other Govt. Departments, Public Sector units and Central Autonomous Bodies for which the bidder is supplying material or having the similar type of contracts and a certificate regarding the satisfactory performance of the contract.
- **4.** Copy of the audited balance sheet of the bidder for the previous three financial years indicating the turnover in supply of the material.
- 5. True copy of Permanent Account Number.
- 6. Details of Sales Tax / VAT/GST along with a copy of certificate to be attached.
- 7. Service Tax No. along with copy of certificate.
- 8. TIN along with copy of certificate.
- **9.** Submission of samples if required, for all items indicated in the schedule of requirements. The make of items proposed to be supplied should be indicated in the format of the schedule of requirements and submitted along with the techno commercial unpriced bid without indicating the pricing components.
- 10. Willingness to execute all orders which are placed to meet emergency requirement on priority basis. The Bidder shall note that standards for workmanship, material and equipment, and references to brand names designated by the Purchaser in the schedule of requirements are intended to be descriptive only and not restrictive. The Bidder may substitute alternative standards, brand names and/or catalogue numbers in his bid, provided that it demonstrates to the Purchaser's satisfaction that the substitutions ensure substantial equivalence to those designated in the Technical Specifications.

Section V. Policy of University against the Corrupt and Fraudulent Practices

University strictly adheres to its policy against corruption and requires that bidders and their agents, subagents, sub-contractors, suppliers etc. shall not indulge in any kind of corrupt practices, fraudulent practices, collusive practices, coercive practices, obstructive practices or other kinds of corruption declared as crimes under Indian law.

a) If bidder or their agents, subagents, sub-contractors, suppliers etc. are found, directly or indirectly, involved in such practices, bid or agreement and execution thereof at any stage may be rejected or cancelled as the case may be by the University and besides it, University may initiate legal actions including civil and criminal proceeding.

For the purpose of this provision the terms are defined as follows:

- (i) "Corrupt practice" is the offering, giving, receiving, or soliciting, directly or indirectly, of anything of value to influence improperly the actions of another party;¹
- (ii) "Fraudulent Practices" is any act or omission including a misrepresentation which knowingly or recklessly made to mislead another party to obtain financial or other benefit or to avoid an obligation;²
- (iii) "Collusive Practice" is an arrangement between two or more parties designed to achieve an improper purpose, including to influence improperly the actions of another party;³
- (iv) "Coercive Practice" is impairing or harming, or threatening to impair or harm, directly or indirectly, any party or the property of the party to influence improperly the actions of a party;⁴
- b) "Obstructive Practice" is deliberately destroying, falsifying, altering, or concealing of evidence material to the investigation or making false statements to investigators in order to materially impede an investigation into allegations of a corrupt, fraudulent, coercive or collusive practice; and/or threatening, harassing or intimidating any party to prevent it from disclosing its knowledge of matters relevant to the investigation or from pursuing the investigation.

¹ For the purpose of this sub-paragraph, "*another party*" refers to a public official acting in relation to the procurement process or contract execution. In this context, "*public official*" includes university staff and employees of other organizations taking or reviewing procurement decisions.

² For the purpose of this sub-paragraph, "party" refers to a public official; the terms "benefit" and "obligation" relate to the procurement process or contract execution; and the "act or omission" is intended to influence the procurement process or contract execution.

³ For the purpose of this sub-paragraph, "party" refers to participants in the procurement process (including public officials) attempting either themselves, or through another person or entity not participating in the procurement or selection process, to simulate competition or to establish bid prices at artificial, non-competitive levels, or are privy to each other's bid prices or other conditions.

⁴ For the purpose of this sub-paragraph, "party" refers to a participant in the procurement process or contract execution.

c) Besides actions under clause (a) University may also take action to blacklist such bidder either indefinitely or for a specified period.

Part-2 Supply Requirements

Section VI. Schedule of Requirements

Contents

3.	Error! Reference source not foundError! Bookmark not define	d.
2.	List of Related Services and Completion Schedule	37
1.	List of Goods and Delivery Schedule	36

Line Item N°	Description of Goods	Quantity	Physical unit	Final (Project Site) Destination as specified in BDS	Delivery (as per Incoterms) Date		
					Earliest Delivery Date	Latest Delivery Date	Bidder's offered Delivery date [to be provided by the bidder]
[insert item No]	[insert description of Goods]	[insert quantity of item to be supplied]	[insert physical unit for the quantity]	[insert place of Delivery]	[insert the number of days following the date of effectiveness the Contract]	[insert the number of days following the date of effectiveness the Contract]	[insert the number of days following the date of effectiveness the Contract]
1.	Defibrillator	01.00	Nos.	As per tender documents	15	30	,,,
2.	High End ECG machine	01.00	Nos.	As per tender documents	15	30	
3.	Mechanical cardiopulmonary resuscitation compressor device	01.00	Nos.	As per tender documents	15	30	
4.	Micro Processor Controlled Syringe Infusion Pump	200.00	Nos.	As per tender documents	15	30	
5.	Multipara Monitor without ETCO2	200.00	Nos.	As per tender documents	15	30	
6.	Portable Ultrasound Machine	1.00	Nos.	As per tender documents	15	30	
7.	Ventilator	50.00	Nos.	As per tender documents	15	30	

1. List of Goods and Delivery Schedule

2. List of Related Services and Completion Schedule

Service	Description of Service	Quantity ¹	Physical Unit	Place where Services shall be performed	Final Completion Date(s) of Services
[insert Service No]	[insert description of Related Services]	[insert quantity of items to be supplied]	[insert physical unit for the items]	[insert name of the Place]	[insert required Completion Date(s)]

1. If applicable

3. TECHNICAL SPECIFICATIONS

"Summary of Technical Specifications"

Item	Name of Goods or			
No	Related Service	Technical Specifications and Standards		
1.	Defibrillator	 Machine should have defibrillator facility for adult and paediatrics. 		
		2. Machine should have AED Mode Facility.		
		3. Machine should be manual and automatic.		
		4. Machine should be portable and low in weight.		
		5. Portable defibrillator should have efficiency of minimum 20 shocks at 200 joule in one full charge.		
		6. Defibrillator with synchronizer circuit.		
		 Machine should provide ECG monitoring through ECG leads. 		
		8. Should have display for heart rate.		
		 Should have alarm for high and low HR. Physical characteristics. 		
		10. Dimension-NA		
		11. Weights-MAX 12kg		
		12. Configuration-Should have audio and visual alarm for battery low.		
		13. Noise<65db		
		 Head dissipation-Low and should be disbursed through cooling machine. 		
		15. Mobility-Portable Energy sources.		
		16. Power requirements-220v.		
		17. Battery should be powered.		
		18. Silenceable alarm for power failure.		
		Accessories spare parts and consumables		
		1. Paddles for ECG monitoring (Adult and Paediatric).		
		 50 gel sheets or pads for monitoring and defibrillation. 		
		WARRANTY + CMC:		
		 5 years from the date of installation AND annual comprehensive maintenance contract to be quoted in price bid for next 5 years after expiration of warranty. 		
		2. Total cost of tender including total cost of CMC value will be considered for finalization of Financial Bid as L1.		
2.	High End ECG machine	1. Fully automatic, microprocessor controlled main and battery operated		
		2. Built-in 8 inch or more colour LCD		
		3. Built-in Ethernet port for data transfer		

	<u>п</u>
	4. Printout on internet thermal printer.
	5. ECG Unit, including: Resting ECG software with 12 simultaneous leads ,Preview of 12 lead ECG prior to
	printout
	6. ECG measurement software, incl. average
	complexes, measurement markers and detail measurement results table
	7. Unit should have Pace maker detection
	8. Unit should have sampling rate of 30,000 Hz
	9. Unit should have frequency range 0 Hz - 255 Hz
	10. Unit should be able to print right precordial lead
	with notation
	11. Should be able to print left posterior leads
	12. Unit should have high resolution A4 size thermal printer
	13. Alphanumeric keyboard with rubber keys
	14. Rechargeable Battery backup for at least 7hrs
	15. Built in Memory of 250 patient ECG or more
	16. Unit should store at least 10 minute or more continuous ECG
	17. Should have QT correction by Bazett, Fredericia, Framingham or Hodges
	18. Should be upgradable to Culpit Artery criteria to
	detect probable site of occlusion
	Accessories to be supplied with each machine:
	1. 10 lead patient cable with banana plugs - 1 no.
	2. Extremity electrodes - 1 set
	3. Valid IEC certificate should be submitted .Machine should be CE or US FDA approved
	WARRANTY + CMC:
	1. 5 years from the date of installation AND annual
	Comprehensive maintenance contract to be quoted in
	price bid for next 5 years after expiration of warranty.
	2. Total cost of tender including total cost of CMC value will
	be considered for finalization of Financial Bid as L1.
Mechanical	Easy Pulse
cardiopulmonary	1. Automatic Chest compression device
resuscitation	Features:
compressor device	2. Portable, battery powered device for high quality
	chest compression
	3. Provides consistent rate & depth or long span
	time
	4. Multidirectional Chest compression
	5. 3D compression (cardiac and thoracic pump
	approach)6. Performs cheat compression on patient suffering
	from Acute cardiac arrest at a consistent ate and
	depth, which is impossible to achieve with
	uepui, which is impossible to achieve with

	1 •
	manual compressions.
	7. Simple operating protocol, single key to perform
	30 compressions and 2 rescue breathes (30:2) in
	continuous manner
	8. Adapted for patient transportation & suitable for
	Large Patients
	9. Autonomous operation for 45 minutes
	10. Additional battery with charge level indicator
	11. Slider with buckle straps helps for safe patient
	transportation.
	12. Achieves approx. 50mmHg of CPP
	13. Connections: external DC input, USB
	···· ·································
	Accessories include:
	(a) Rechargeable Lithium Polymer Battery 18.5V
	4.0Ah
	(b) Slider with Strap Buckle
	(c) Battery charging station
	(d) charging unit Adapter
	(e) DC Input adapter
	(f) User Manual
	(g) Carrying Bag
	WARRANTY + CMC:
	1. 5 years from the date of installation AND annual comprehensive maintenance contract to be quoted in price bid for next 5 years after expiration of warranty.
	2. Total cost of tender including total cost of CMC value
	will be considered for finalization of Financial Bid as
	L1.
Micro Processor	
Controlled Syringe	Physical Characteristics:
Infusion Pump	•
	1. Light weight and should not be more than 2.5 kg
	2. Capability for working on AC power (220V 50 Hz)
	without external adapter or transformer and internal
	rechargeable battery with later backup of 5 hrs.
	3. Backlite clearly legible LCD display Information
	Functional Characteristics:
	1. Flow rate ranges from 0.1ml/hr to 1200 ml/h with
	1. From rate ranges from 0.1111/11 to 1200 mi/if with accuracy of $\pm 1\%$ on mechanism.
	2. Works with syringe capacities of 5, 10, 20, 30/35,
	50/60 CC and is compatible with 100 types of syringe
	3. Operates on ml/hr mode, Volume / time: 0.1 - 99.9 mL;

 3 modes of Priming (Mandatory, not mandatory, or advised) with max flow rate: 1200 ml/hr. Have direct bolus option with flow rate 50ml/hr to 1200ml/hr with increment of 50ml/hr. Have KVO option ranging from 0.1 to 5ml/hr Fast start option with provision of programmable pause settable from lmin to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg). Can be enables / disabled and adjusted. Dynamic Pressure System with facility to set maximum and minimum threshold. Anti-bolus feature Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, discharged battery alarm, buttery capacity pre-alarm, discharged battery alarm, buttery capacity display in hours and minutes, unconfirmed programming technical mafunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection. Huse following later Occle/lock code (on Keypad). Push guard for syringe protection. Rusing lock clamp for versatile clamp and horizontal clamp data lock in a fination on a rail or on a pole Stacking accessory to fit in 4 or 6 pumps available. RS232 Communication port with HL7 compatibility. Save 1500 data log events in real time Li-ion Smart battery, remaining battery life (when fully charged): > 13 at 5 mL/h. Electroma	l	
 advised) with max flow rate: 1200 ml/hr. Have direct bolus option with flow rate 50ml/hr to 1200ml/hr with increment of 50ml/hr. Have KVO option ranging from 0.1 to 5ml/hr Fast start option with provision of programmable pause settable from 1min to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg		00h01 - 96h00, Volume Limit:0.1-999ml.
 Have direct bolus option with flow rate 50ml/hr to 1200ml/hr with increment of 50ml/hr. Have KVO option ranging from 0.1 to 5ml/hr Fast start option with provision of programmable pause settable from limin to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 50 momHg. (25 mmHg increment from 50 to 50 mmHg). Can be enables / disabled and adjusted. Dynamic Pressure System with facility to set maximum and minimum threshold. Anti-bolus feature Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit pre-alarm, ded fiftion alarm, volume limit pre-alarm, discharged driving mechanism alarm, low battery pre-alarm, discharged battery alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). Push guard for syringe protection. Display size of 66 mm x 33 mm Sving lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole Self stackability of tyto 3 bat 5 mL/h. Electromagnetic compatibility EMC - IEC 60601-2-2, IEC 60601-2-24 Medical Device Directive - CE 0123 marking in 		
 1200ml/hr with increment of 50ml/hr Have KVO option ranging from 0.1 to 5ml/hr Fast start option with provision of programmable pause settable from lmin to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 900 mmHg.) Can be enables / disabled and adjusted. Dynamic Pressure System with facility to set maximum and minimum threshold. Anti-bolus feature Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement Facility for Syringe barrel clasp check, plunger head detection. Have following alarm. Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, declosin flow rate flows and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation connection failure, plug-head disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). Push guard for syringe protection. Display size of 66 mm x 33 mm Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 33 mm Rus Suing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 34 minibility. Zave 1500 data log events in real time Li-ion Smart battery, remaining battery life and battery charge level available on the display. Batt		
 6. Have KVO option ranging from 0.1 to 5ml/hr 7. Fast start option with provision of programmable pause settable from limit to 24hrs 8. The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. 9. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg (25 mmHg increment from 250 to 900 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume/ dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit quarm, wolume and minimus, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarn (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Ecommunication compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life (when fully charged): > 13 h at 5 mL/h. 24. Electomagnetic compatibility EMC - IEC 60601-1		5. Have direct bolus option with flow rate 50ml/hr to
 Fast start option with provision of programmable pause settable from 1min to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 250 mmHg / 26 mmHg increment from 50 to 250 mmHg / 50 mmHg increment from 50 to 900 mmHg). Can be enables / disabled and adjusted. Dynamic Pressure System with facility to set maximum and minimum threshold. Anti-bolus feature Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement Facility for Syringe barrel class check, plunger head detection, anti-siphon system check, flange detection. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). Push guard for syringe protection. Display size of 66 mm x 33 mm Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole Stacking accessory to ft in 4 or 6 pumps available. RS232 Communication port with HL7 compatibility. Save 1500 data log events in real time Li-ion Smart battery, remaining battery Life and battery charge level available on the display. Battery Life (when fully charged) > 13 h at 5 mL/h.<td></td><td>1200ml/hr with increment of 50ml/hr.</td>		1200ml/hr with increment of 50ml/hr.
 Fast start option with provision of programmable pause settable from 1min to 24hrs The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg / 50 mm		6. Have KVO option ranging from 0.1 to 5ml/hr
 settable from 1min to 24hrs 8. The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. 9. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg (25 mmHg increment from 50 to 250 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, ned of infusion pre-alarm, discharged battery alarm, buttery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Stacking accessory to fit in 4 or 6 pumps available. 21. Ecicomagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 23. Medical Device Directive - CE 0123 marking in 		1 00
 The night mode can be programmed manually or automatically in a variable time range to decrease the brightness of the screen. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg). Can be enables. / disabled and adjusted. Dynamic Pressure System with facility to set maximum and minimum threshold. Anti-bolus feature Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement Facility for Syringe barrel clasp check, plunger head detection, anti-siphon cystem occlusion pressure alarm, olume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pase end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). Pushg uard for syringe protection. Display size of 66 mm x 33 mm Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole Self stackability of upto 3pumps Stacking accessory to fit in 4 or 6 pumps available. Li-ion Smart battery, remaining battery life (when fully charged): > 13 h at 5 mL/h. Elec 60601-2:24 Medical Device Directive - CE 0123 marking in 		
 automatically in a variable time range to decrease the brightness of the screen. 9. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 250 to 250 mmHg. Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged):>13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 brightness of the screen. 9. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion presale and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 9. Variable and 3 pre-set levels pressure mode. Range from 50 to 900 mmHg. (25 mmHg increment from 50 to 250 mmHg) / 50 mmHg increment from 250 to 900 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Stacking accessory to fit in 4 or 6 pumps available. 21. Res232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-1-24 25. Medical Device Directive - CE 0123 marking in 		· ·
 from 50 to 900 mmHg. (25 mmHg increment from 50 to 900 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit raterm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 of 6 pumps available. 21. Li-ion Smart battery, remaining battery Life (when fully charged) > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-24 25. Medical Device Directive - CE 0123 marking in 		e e
 to 250 mmHg / 50 mmHg increment from 250 to 900 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Li-ion Smart battery, remaining battery Life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		· · · ·
 mmHg). Can be enables / disabled and adjusted. 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion aream, occlusion pressure alarm, end of infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical maflunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): >13 hat 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 10. Dynamic Pressure System with facility to set maximum and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a role 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		• •
 and minimum threshold. 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged):> 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		e,
 11. Anti-bolus feature 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, buttery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 12. Onscreen graphical display of the following history "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 "Volume / dose infused, pressure, flow rate" is present. 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Rs232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 13. Drug library of 3600 drugs distributed in 19 profile, customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 		
 customisable as per user requirement 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 14. Facility for Syringe barrel clasp check, plunger head detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. Re322 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		• • • •
 detection, anti-siphon system check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion area, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		customisable as per user requirement
 check, flange detection. 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		14. Facility for Syringe barrel clasp check, plunger head
 15. Have following alarm Occlusion pressure pre-alarm, occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		detection, anti-siphon system
 occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		check, flange detection.
 occlusion pressure alarm, end of infusion pre-alarm, end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		15. Have following alarm Occlusion pressure pre-alarm,
 end of infusion alarm, volume limit pre-alarm, volume limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 limit alarm, keypad manual locking or keypad lock, hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		-
 hard and soft flow rate limits, start infusion at pause end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		· · · ·
 end, Disengaged driving mechanism alarm, low battery pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 pre-alarm, discharged battery alarm, battery capacity display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		-
 display in hours and minutes, unconfirmed programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 programming, technical malfunction alarm (auto-test, rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 rotation), drive system advance check, watchdog check, communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 communication connection failure, plug-head disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 disengagement, auto-lock/lock code (on Keypad). 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 16. Push guard for syringe protection. 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		· 1 U
 17. Display size of 66 mm x 33 mm 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 18. Swing lock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 clamp that allows the fixation on a rail or on a pole 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		1
 19. Self stackability of upto 3pumps 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		• •
 20. Stacking accessory to fit in 4 or 6 pumps available. 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 21. RS232 Communication port with HL7 compatibility. 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 22. Save 1500 data log events in real time 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		• • • • • •
 23. Li-ion Smart battery, remaining battery life and battery charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 charge level available on the display. Battery Life (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		e e
 (when fully charged): > 13 h at 5 mL/h. 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
 24. Electromagnetic compatibility EMC - IEC 60601-1-2, IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in 		
IEC 60601-2-24 25. Medical Device Directive - CE 0123 marking in		
25. Medical Device Directive - CE 0123 marking in		• • •
		25. Medical Device Directive - CE 0123 marking in
compliance with the Council Directive 93/42/EEC		

	 Safety+ Certification European CE and /or US FDA certified product Electrical Compliance - Protection against leakage current: Defibrillation-proof type CF and Protection against electric shocks: class II in accordance with IEC 60601-1 Alarm system - IEC 60601-1-8 Usability Engineering - IEC 60601-1-6 and IEC 62366 Protection against Fluid spillage
	 Warranty + CMC Warranty 5 years from the date of installation. Comprehensive Annual Maintenance to be quoted in price bid for next 5 years after expieration of warranty. Demonstration of Equipment in SSH. Total cost of tender including total cost of CMC value will be considered for finalization of Financial Bid as L1.
Multipara Monitor without ETCO2	1. Power supply AC 100-240 V 50Hz/ 60Hz/ 12 V DC input
	2. Defibrillator & Cautery protected
	3. Safety Glass Category I
	4. Patient monitor should have a minimum 12.1 "high resolution color TFT LCD display & touch screen.
	5. Should be capable of displaying 96 hours trends for the monitored parameters along with 100 NIBP readings and event recording facility of 100 events, each event to be of more than 8 secs.
	 Battery backup >240 minutes (2 battery 480 minutes) (2nd battery optional), battery charging time < 6 hours (2 batteries 12 hours)
	7. Recorder (optional) method : Thermal Dot Array, Paper width 50 mm, Paper speed 12.5/25/50 num/sec, Trace max. 3
	8. User should be able to set alarms for all parameters
	 Environment : operation temperature 5 – 40 degree C, Storage temperature 20 – 55 degree C.
	10. Should be able to display at least 5 waveforms.
	11. Should be capable of numerically displaying information of all parameters for which waveforms are being displayed, as well as displaying numeric values of additional parameters.
	12. Should be able to store the data on external memory stick
	13. should have Central monitor connectivity through Ethernet

14.	Should have numeric mode to display numeric's in large font size		
ECC 15.			
16	· · ·		
17			
18	C C		
	height)		
20.	5		
21.	In built defibrillator & cautery protection		
NII			
22.			
23.	1		
24.			
25.	Inflation pressures should be automatically released if there are errors		
26.	Measurement range 20-20mmHg.		
27.	Measurement interval 2 minutes to 2 hours (user can select interval)		
28.	Reusable blood pressure cuff for adults, pediatric, neonatal use should be available.		
	Pulse Oximeter		
29	Should display plethy smograph and S_pO_2 values.		
30	O_2 saturation range 1 to 100%		
31	Accuracy ±2% between 50 and 99%		
32	Multipara monitor suitable for use with new born		
	(wrap around / y probe) child and adult based on Nellcor/Masimo Technology for SpO2 monitoring		
33	Motion artifacts should not affect readings		
34	Reusable adult probe should be provided along with		
	Wrap around probe for newborns.		
35	Should also display perfusion index		
	Respiration		
36.	Should be part of ECG module		
37.	Range 0 to 150 breaths per minute		
38.	Resolution – one breath		
39.	Accuracy :± 2 breath		
40.	Apnea alarm : Selectable		
	Temperature		
41.	e		
42.	Resolution : 0.1° C		

		43.	Accuracy $\pm 0.2^{\circ}$ C
		44.	Skin probe: 1 number to be provided with each module
		0	thers
		45.	Import & Export Configuration option
		46.	Full disclosures facility
		47.	HRV and NIBP Analyzee option
		48.	Total 10 waveform display
		49.	ST & Arrhythmia analysis
		50.	Drug dosage, Hemodynamic, Nephridium, Ventilation & Oxygenation Calculation software.
		51.	OXYCRG Software
		52.	Standard configuration ECG, HR, RESP, NIBP, Digital Spo2, PR, Dual TEMP.
		53.	Central Station connectivity Wired
		54.	Pacemaker detection Facility.
		55.	Video output (VGA)
		G	eneral Specification
		56.	Power supply 230V \pm 15%, 5 Hz \pm 3% with inbuilt in battery charger and chargeable battery
		57.	Tropicalization: Operation temperature up to 40° C, Storage temperature up to 60° C, Relative temperature up to 90% non condensing
		58.	Trends 168 hours graphical tend and tabular trend with full disclosure
		59.	Alarms for all parameters manual or default alarm setting. Display of cautionary/ alarm messages.
		60. Certification European CE/USFDA	
		W	/ARRANTY + CMC
		61.	5 years from the date of installation AND annual comprehensive maintenance.
		62.	contract to be quoted in price bid for next 5 years after expiration of warranty. Total cost of tender
			including total cost of CMC value will be
			considered for finalization of Financial Bid as L1.
Portable	Ultrasound	A stat	te of art fully digital, compact Color Doppler Ultrasound
Machine		machi	ine is required for neonatal and paediatric use with
		follow	ving technical features:
		1.	Unit should be able to give very high image quality
			with advance technologies like compound imaging for
			better contrast resolution, tissue differentiation and

edge detection, equivalent to high end cart based
systems. Please specify the technology.
2. Unit should be compact, durable & less than 5 kg in
weight (including battery).
3. Imaging modes of Real time 2D, Color Doppler, Power
Doppler, Pulsed wave Doppler, Continuous wave
Doppler must be available.
4. System must have fast start up to scanning in less than
30 seconds from off condition, for use in ICU and
emergency conditions.
5. System should support transducer technologies like
phased array, convex, linear & TEE format.
6. The system should have a broadband architecture with
an operating frequency of at least 1 to 15 MHz
7. Cine memory of atleast 250 frames should be available
on all operating modes.
8. The system shall process a dynamic range that is at
least 165db. The system must display at a maximum
depth of 35 cm.
9. The system must have dedicated calculation packages
for cardiac, Vascular measurements.
10. The offered unit must have flat LCD/ TFT monitor of
less than 12 inches with Anti-glare coating and wide
viewing angle.
11. Alphanumeric soft keys backlit and splash resistant
keypad with easy access scans controls, facility to
sanitize the system keyboard to avoid cross
contamination.
12. System should possess Needle Visualization software to track the needle clearly at steep angles during procedural guidance while maintaining image quality of the target structures and the surrounding anatomy with simple On/Off functionality on both linear & curvilinear transducers.
13. The system must have the ability to function by AC/DC
or battery power with the same degree of functionality,
the battery life (run time) shall be al least 2 (Two)
hours, this need to demonstrate.

14. The system must have archive capability for storage &
retrieval of images and clips. It should have atleast 2
USB slots, which allow for direct sharing of images
(JPEG) and clips (AVI) to a PC.
15. Unit & transducers must be rugged, drop safe on
accidental fall/hit on hard surface so as to withstand
busy hospital situations.
16. The system must have in-built memory of at least 16
GB for storing Patient data & studies.
17. The system should be capable of supporting all DICOM
functionality (Storage, Print, and Work List), also shall
be compatible to connect to PACS.
18. System should have both European CE and US
FDA quality certification
Transducers & other accessories to be supplied as
standard:
 3-8MHz (+/-1 MHz) curved array multi-frequency, broadband transducer for pediatric abdominal, MSk, nerve & spine exam applications.
2. 6-13 (+/-1) MHz multi-frequency, broadband, 25 mm
linear array transducer for pediatric vascular,
musculoskeletal, nerve and superficial imaging.
3. 4-8 MHz (±1 MHz) multi-frequency broadband
Pediatric Phased array transducer for cardiac
examination.
4. 6-13 MHz (±1 MHz) multi-frequency broadband 25
mm Linear Hockey stick shape transducer for vascular,
nerve and superficial images.
5. Detachable, imported, molded, OEM Trolley/cart to
mount transducers and machine.
6. When the machine is mounted on trolley, facility for
connecting three transducers simultaneously with easy
selection of active transducer should be possible.
Optional Transducers/Software/Items:
WARRANTY:
1. Main unit and transducers should be covered with
comprehensive onsite warranty for five years

	 commencing from the date of issue of installation certificate. Annual Comprehensive maintenance contract to be quoted in price bid for next 5 years after expiration of warranty. 2. Total cost of tender including total cost of CMC value will be considered for finalization of Financial Bid as L1. 		
	ESSENTIAL REQUIREMENT: 1. Onsite Product training and access to online education		
	material must be provided post installation of the system.		
Ventilator	Intended use all type patients – Adult, Paediatric and Neonatal (including premature) 1. TECHNICAL CHARACTERISTICS		
	 A. A Should have facility for Invasive and Non-Invasive ventilation. B. Microprocessor Control suitable for Neonatal and Paediatric ventilation and adult ventilation. 		
	C. Should have operative modes of ventilation equipped with newer modes of ventilation.		
	a) Assist/Control, Volume control & Pressure control		
	b) Pressure support		
	c) SIMV with pressure support (Pressure and volume control)		
	d) PEEP		
	e) Inverse ration Ventilation/Restrictive and pertains-APRV/BIPAP/Equivalent		
	f) Non invasive ventilation-BIPAP, CPAP		
	 g) HFNC with Active servo controlled humidifier preferably Fishered Paykel 850 make or its equivalent. 		
	h) Apnea ventilation, user selectable, volume & pressure control		
	D. Should have built in color screen TFT/LCD display of minimum 12 ["] for display of waveforms and monitored value with standard accuracy with alarms.		
	E. Should have inbuilt facility with ETcO2.		
	F. Should have facility to measure and display of the following parameters :		
	a) Airway Pressure (Peak & Mean)		
	b) Tidal volume (Inspired & Expired).		
	c) Minute volume (Inspired & Expired).		
	d) Respiratory mechanics.		

	e)	Spontaneous Minute Volume.
	c) f)	Total Frequency.
	<i></i>	FiO2 dynamic.
	g) b)	Intrinsic PEEP
	h)	Plateau Pressure.
	i)	
	j)	Resistance & Compliance.
	k)	Use selector Alarms for all measures & monitored parameters.
	1)	Occlusion Pressure.
	m)	Pressure Flow & Volume curves.
	n)	Leak.
G.		matic compliance and leakage compensation for it and ET tube.
H.		ld have facility of log book, for event and alarms date & time.
I.	Shou	ld have following setting:
	a)	Tidal volume (Minimum 2ml, Maximum up to 2000ml); pre-set range for both neo-natal & paediatric modes to be provide.
	b)	Inspiratory pressure (upto 60cm of H2O).
	c)	Respiratory rate 1 to 150 bpm.
	d)	Apnoea back up rate.
	e)	CPAP/PEEP.
	f)	Pressure support.
	g)	FiO2 setting range between 21% and 100%.
	h)	Pause time.
	i)	Pressure/flow Trigger.
	j)	Inspiratory flow up to 120 Lpm.
	k)	Oxygen cylinder/central pipeline connector/(to be supplied along with the machines) should be compatible with ventilator
	1)	Disposable Heat Moisture Exchanger, qty. 100 to be supplied with unit.
	m)	Imported stand alone Medical Air compressor from the same manufacturer (of ventilator) M. Medical Air Compressor should automatically activate in the event of wall air supply loss.
1.1	User	's interface: Manual and Automatic
1.2	Softv	vare
	a.	Inbuilt software.
	b.	convenient and quick USB/RJ45/RS232/Ethernet interfaces, for
	communications.	
2.	PHYSICAL CHARACTERISTICS	
2.1	Conf	iguration

a	a. Compatible hinged arm for holding the circuit
ł	b. Should have caster with braking system
C	e. Noise of device operation max 50dbA
Ċ	 Should have audio visual alarm for battery low, source gas low and high/low pressure in the breathing circuit or souce gas inlet.
e	e. Should maintain nominal Temp of the control unit and the heat should be disbursed through a cooling mechanism.
f	Alarm volume – min. 65dB
2.2 N	Mobility , portability : Yes
3. I	ENERGY SOURCE
3.1 I	Power requirements: Input voltage 220 VAC, 50Hz
3.2 I	Sattery operated :
8	a. Battery powered, silenceable alarm for power failure
ł	b. Battery charger to be integral to mains power supply and to charge battery during mains power operation of unit.
C	c. Internal, replaceable, rechargeable battery allows operation for at least 1 hour in the event of power failure with three years replacement warratny.
C	Folerance : Voltage corrector/ stabilizer to allow operation at $\pm 10\%$ of 220V AC Use of SMPS to correct voltage
3.4 I	Protection:
8	a. Electrical protection, resettable over current breakers or replaceable fuses (fitted in both live and neutral lines).
ł	b. Leakage.
3.5 I	Power consumption: To be declared by the supplier.
	ACCESSORIES, SPARE PARTS, CONSUMABLES
4.1 A	Accessories & Spares:
а	a. Full face mask 5 Nos. each of 0, 1 and 3
ł	b. Nasal cannulae for neonates -5 nos.
C	c. Reusable breathing circuit of silicone material 2 no. each with embedded heater wire for neonatal and 3 no. with embedded heater wire for paediatric patient.
Ċ	 Air & oxygen hose – 1 each.
e	e. Electronic humidifier
f	 Bipap Mask (for non invasive ventilation) all sizes.
Ę	g. Nebulisation (in line) facility.

h. External UPS with capacity to run the compressor for at least 1 hour.
5. ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS
5.1 Atmosphere/Ambiance:
 a. Operating condition : Capable of operating continuously in ambient temperature of 10 to 40 deg C and relative humidity of 15to 90% in ideal circumstances.
 b. Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.
5.2 User's care, Cleaning, Disinfection & Sterility issues: Complete unit to be easily wipeable and sterilisable using alcohol and other chemical agents.
6. STANDARDS AND SAFETY
6.1 Certifications:
a. FDA (US)/CE (EU) from authorized third party and BIS/ISO 13485
 b. Relevant IEC-60601-Part 1 & 2, certificates by a notified agency or reputed testing laboratory.
6.2 Manufacturer/Supplier should have ISO 13485 certificate for quality standard
7. TRAINING AND ISNTALLATION
7.1 Availability of 5 amp/15Amp. electrical sockets.
7.2 Oxygen supply.
7.3 Medical air supply.
7.4 Supplier to perform installation, safety and operation checks before handover.
7.5 Local clinical staff to affirm completion of installation.
7.6 Training of users in operation and basic maintenance shall be provided and
ventilator calibration training and troubleshooting training.
7.7 Advanced maintenance tasks required shall be documented.
7.8 Helpline support with maximum response time less than 24 hours.
8. WARRANTY + CMC
8.1 5 years from the date of installation AND annual comprehensive maintenance contract to be quoted in price bid for next 5 years after expiration of warranty.
8.2 Total cost of tender including total cost of CMC value will be considered for finalization of Financial Bid as L1.

Note:-

- 1. The quantity shown in the schedules can be 50% increased or decreased depending upon the actual requirement at the time of placing of orders. However, the University reserves its right to take delivery as per the actual requirement.
- 2. The shelf life of the above materials mentioned in the schedules of the requirement should be not less than 80% at the time of supply. The supplier should supply fresh stock.
- 3. The bidders should submit the samples of the quoted items in sealed original of the same batch on or before the bid submission end date and time.

Part-3 Contract

Section VII. General Conditions of Contract

Table of Clauses

1.	Definitions	. 54
2.	Contract Documents	. 55
3.	Corrupt and Fraudulent Practices	. 55
4.	Interpretation	. 55
5.	Language	. 56
6.	Joint Venture, Consortium or Association	. 56
7.	Eligibility	. 56
8.	Notices	. 57
9.	Governing Law	. 57
10.	Settlement of Disputes	. 57
11.	Obligations During Arbitrations	. 58
12.	Scope of Supply	. 58
13.	Delivery and Documents	. 58
14.	Supplier's Responsibilities	. 58
15.	Contract Price	. 58
16.	Terms of Payment	. 58
17.	Taxes and Duties	. 59
18.	Performance Security	. 59
19.	Copyright	. 60
20.	Confidential Information	. 60
21.	Subcontracting	. 60
22.	Specifications and Standards	.61
23.	Packing and Documents	. 61
24.	Insurance	.61
25.	Transportation and Incidental Services	. 62
26.	Inspections and Tests	. 62
27.	Liquidated Damages	. 63
28.	Warranty	. 64
29.	Patent Indemnity	. 64
30.	Force Majeure	. 65
31.	Change Orders and Contract Amendments	. 65
32.	Extensions of Time	. 66
33.	Termination	. 66
34.	Assignment	. 67

Section VII. General Conditions of Contract

- 1 Definitions
- 1.1 The following words and expressions shall have the meanings hereby assigned to them:
 - (a) "University" means Banaras Hindu University established under Banaras Hindu University Act 1915.
 - (b) "Contract" means the Contract Agreement entered into between the Purchaser and the Supplier, together with the Contract Documents referred to therein, including all attachments, appendices, and all documents incorporated by reference therein.
 - (c) "Contract Documents" means the documents listed in the Contract Agreement, including any amendments thereto.
 - (d) "Contract Price" means the price payable to the Supplier as specified in the Contract Agreement, subject to such additions and adjustments thereto or deductions therefrom, as may be made pursuant to the Contract.
 - (e) "Day" means calendar day.
 - (f) "Completion" means the fulfillment of the Related Services by the Supplier in accordance with the terms and conditions set forth in the Contract.
 - (g) "GCC" means the General Conditions of Contract.
 - (h) "Goods" means all of the commodities, raw material, machinery and equipment, and/or other materials that the Supplier is required to supply to the Purchaser under the Contract.
 - (i) "The Project Site," term where applicable, means the place of work named in the Special Conditions of Contract (SCC).
 - (j) "Purchaser" means Banaras Hindu University, institute, faculty, department and other entities of the University competent for purchasing Goods and Services, as specified in the SCC.
 - (k) "Related Services" means the services incidental to the supply of the goods, such as insurance, installation, training and initial maintenance and other such obligations of the Supplier under the Contract.
 - (l) "SCC" means the Special Conditions of Contract.
 - (m) "Subcontractor" means any person, private or government entity, or a combination of the above, to whom any part of the Goods to be supplied or execution of any part of the Related Services is subcontracted by the Supplier.

- (n) "Supplier" means the person, private or government entity, or a combination of the above, whose bid to perform the Contract has been accepted by the Purchaser and is named as such in the Contract Agreement.
- (o) "The Project Site," where applicable, means the place named in the SCC.
- 2 Contract 2.1 Subject to the order of precedence set forth in the Contract Agreement, all documents forming the Contract (and all parts thereof) are intended to be correlative, complementary, and mutually explanatory. The Contract Agreement shall be read as a whole.
- 3 Corrupt and Fraudulent Practices 3.1 The University requires compliance with its policy against the corrupt and fraudulent practices as set forth Section-V.
 - 3.2 The Purchaser requires the Supplier to disclose any commissions or fees that may have been paid or are to be paid to agents or any other party with respect to the bidding process or execution of the Contract. The information disclosed must include the name and address of the agent or other party, the amount and currency, and the purpose of the commission, gratuity or fee.
 - **Interpretation** 4.1 Unless the context requires otherwise, singular means plural and vice versa.
 - 4.2 The Contract constitutes the entire agreement between the Purchaser and the Supplier and supersedes all communications, negotiations and agreements (whether written or oral) of the parties with respect thereto made prior to the date of Contract.
 - 4.3 Amendment

4

No amendment or other variation of the Contract shall be valid unless it is reduced to writing, dated, expressly refers to the Contract, and is signed by the duly authorized representative of each party thereto.

- 4.4 Nonwaiver
 - a. Subject to GCC Sub-Clause 4.5(b) below, no relaxation, forbearance, delay, or indulgence by either party in enforcing any of the terms and conditions of the Contract or the granting of time by either party to the other shall prejudice, affect, or restrict the rights of that party under the Contract, neither shall any waiver by either party of any breach of Contract operate as waiver of any subsequent or continuing breach of Contract.
 - b. Any waiver of a party's rights, powers, or remedies under
 - 55

the Contract must be in writing, dated, and signed by an authorized representative of the party granting such waiver, and must specify the right and the extent to which it is being waived.

4.5 Severability

If any provision or condition of the Contract is prohibited or rendered invalid or unenforceable, such prohibition, invalidity or unenforceability shall not affect the validity or enforceability of any other provisions and conditions of the Contract.

- 5 Language 5.1 The Contract as well as all correspondence and documents relating to the Contract exchanged by the Supplier and the Purchaser, shall be written in the language specified in the SCC. Supporting documents and printed literature that are part of the Contract may be in any language provided they are accompanied by an accurate translation of the relevant passages in the language specified, in which case, for purposes of interpretation of the Contract, such translation shall govern.
 - 5.2 The Supplier shall bear all costs of translation to the governing language and all risks of the accuracy of such translation, for documents provided by the Supplier.
- 6 Joint Venture, Consortium or Association
 6.1 If the Supplier is a joint venture, consortium, or association, all of the parties shall be jointly and severally liable to the Purchaser for the fulfilment of the provisions of the Contract and shall designate one party to act as a leader with authority to bind the joint venture, consortium, or association. The composition or the constitution of the joint venture, consortium, or association the prior consent of the Purchaser.
 - **Eligibility** 7.1 The Bidder should not have been declared insolvent by the competent court.

7

- 7.2 The Bidder should not be disqualified for contract under the law of the land.
- 7.3 The Bidder should not be adjudged defaulter of Tax Payment under Income Tax Law or any other Law for the time being in-force.
- 7.4 The Supplier and its Subcontractors shall have the nationality of an eligible country. A Supplier or Subcontractor shall be deemed to have the nationality of a country if it is a citizen or constituted, incorporated, or registered, and operates in conformity with the provisions of the laws of that country. Nationality must be disclosed by the supplier
- 7.5 All Goods and Related Services to be supplied under the

Contract shall have their origin in Eligible Countries. For the purpose of this Clause, origin means the country where the goods have been grown, mined, cultivated, produced, manufactured, or processed; or through manufacture, processing, or assembly, another commercially recognized article results that differs substantially in its basic characteristics from its components.

- 8 Notices 8.1 Any notice given by one party to the other pursuant to the Contract shall be in writing to the address specified in the SCC. The term "in writing" means communicated in written form with proof of receipt.
 - 8.2 A notice shall be effective when delivered or on the notice's effective date, whichever is later.
- **9** Governing Law 9.1 The Contract shall be governed by and interpreted in accordance with the laws of India, unless otherwise specified in the SCC.
 - 9.2 Throughout the execution of the Contract, the Contractor shall comply with the import of goods and services prohibitions in India when
 - (a) as a matter of law or official regulations, India prohibits commercial relations with that country; or
 - (b) by an act of compliance with a decision of the United Nations Security Council taken under Chapter VII of the Charter of the United Nations, India prohibits any import of goods from that country or any payments to any country, person, or entity in that country.
 - 10.1 The Purchaser and the Supplier shall make every effort to resolve amicably by direct informal negotiation any disagreement or dispute arising between them under or in connection with the Contract.
 - 10.2 If the parties have failed to resolve their dispute or difference by such mutual consultation, then either the Purchaser or the Supplier may give notice to the other party of its intention to settle the issue by arbitration, as hereinafter provided, as to the matter in dispute, no arbitration in respect of the matter be commenced unless such notice is given in accordance with this Clause for the final settlement of the matter. Arbitration may be commenced prior to or after delivery of the Goods under the Contract.
 - 10.3 All questions, disputes and differences arising shall be referred by the Vice-Chancellor, Banaras Hindu University to the sole arbitrator for arbitration under the provision of the Arbitrations and Conciliation Act, 1996.

10 Settlement of Disputes

11	Obligations During Arbitrations	11.1	 Notwithstanding any reference to arbitration in Clause 10, (a) the parties shall continue to perform their respective obligations under the Contract unless they otherwise agree; and
			(b) the Purchaser shall pay any amount due to the Supplier.
12	Scope of Supply	12.1	The Goods and Related Services to be supplied shall be as specified in the Schedule of Requirements.
13	Delivery and Documents	13.1	Subject to GCC Sub-Clause 33.1, the Delivery of the Goods and Completion of the Related Services shall be in accordance with the Delivery and Completion Schedule specified in the Schedule of Requirements. The details of shipping and other documents to be furnished by the Supplier are specified in the SCC.
14	Supplier's Responsibilities	14.1	The Supplier shall supply all the Goods and Related Services included in the Scope of Supply in accordance with GCC Clause 12, and the Delivery and Completion Schedule, as per GCC Clause 13.
15	Contract Price	15.1	Prices charged by the Supplier for the Goods supplied and the Related Services performed under the Contract shall not vary from the prices quoted by the Supplier in its bid, with the exception of any price adjustments authorized in the SCC.
16	Terms of Payment	16.1	Ordinarily, payments for services rendered or supplies made shall be released only after the services have been rendered or supplies appropriate to the requirement made. However in following cases advance payments may be made if specified in SCC:
			(i) Advance payment demanded by firms holding maintenance contracts for servicing of Air-conditioners, computers, other costly equipment, etc.
			(ii) Advance payment demanded by firms against fabrication contracts, turnkey contracts etc.
			Such advance payment should not exceed the following limits :-
			(i) Thirty percent of the contract value to private firms;
			(ii) Forty percent of the contract value to a State or central Government agency or a Public Sector Undertaking; or
			(iii) In case of maintenance contract, the amount should not exceed the amount payable for six months under the

contract.

- 16.2 The Supplier's request for payment shall be made to the Purchaser in writing, accompanied by invoices describing, as appropriate, the Goods delivered and Related Services performed, and by the documents submitted pursuant to GCC Clause 13 and upon fulfilment of all other obligations stipulated in the Contract.
- 16.3 Payments shall be made promptly by the Purchaser, within ninety (90) days after submission of an invoice or request for payment by the Supplier, and after the Purchaser has accepted it.
- 16.4 The currencies in which payment shall be made to the supplier under this contract shall be Indian currency unless otherwise agreed.
- **17 Taxes and Duties** 17.1 For goods manufactured outside India, the Supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside India.
 - 17.2 For goods Manufactured within India, the Supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted Goods to the Purchaser.
 - 17.3 If any tax exemptions, reductions, allowances or privileges may be available to the Supplier in India, the Purchaser shall use its best efforts to enable the Supplier to benefit from any such tax savings to the maximum allowable extent or country or origin, the supplies shall provide benefit from any such tax sowing to the purchaser.
- 18 Performance Security
 18.1 If required as specified in the SCC, the Supplier shall, within twenty-one (21) days of the notification of contract award, provide a performance security for the performance of the Contract in the amount specified in the SCC.
 - 18.2 The proceeds of the Performance Security shall be payable to the Purchaser as compensation for any loss resulting from the Supplier's failure to complete its obligations under the Contract.
 - 18.3 As specified in the SCC, the Performance Security, if required, shall be denominated in the currency(ies) of the Contract or in a freely convertible currency acceptable to the Purchaser; and shall be in one of the format stipulated by the Purchaser in the SCC, or in another format acceptable to the Purchaser.
 - 18.4 Performance security should remain valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including warranty obligation.
 - 18.5 Bid security shall be refunded to the successful bidder within

30 days of receipt of performance security.

19 Copyright 19.1 The copyright in all drawings, documents, and other materials containing data and information furnished to the Purchaser by the Supplier herein shall remain vested in the Supplier, or, if they are furnished to the Purchaser directly or through the Supplier by any third party, including suppliers of materials, the copyright in such materials shall remain vested in such third party.

The Purchaser and the Supplier shall keep confidential and 20 Confidential 20.1 shall not, without the written consent of the other party Information hereto, divulge to any third party any documents, data, or other information furnished directly or indirectly by the other party hereto in connection with the Contract, whether such information has been furnished prior to, during or following completion or termination of the Contract. Notwithstanding the above, the Supplier may furnish to its Subcontractor such documents, data, and other information it receives from the Purchaser to the extent required for the Subcontractor to perform its work under the Contract, in which event the Supplier shall obtain from such Subcontractor an undertaking of confidentiality similar to that imposed on the Supplier under GCC Clause 20.

- 20.2 The Purchaser shall not use such documents, data, and other information received from the Supplier for any purposes unrelated to the contract. Similarly, the Supplier shall not use such documents, data, and other information received from the Purchaser for any purpose other than the performance of the Contract.
- 20.3 The obligation of a party under GCC Sub-Clauses 20.1 and 20.2 above, however, shall not apply to information that:
 - a. now or hereafter enters the public domain through no fault of that party;
 - b. can be proven to have been possessed by that party at the time of disclosure and which was not previously obtained, directly or indirectly, from the other party; or
 - c. otherwise lawfully becomes available to that party from a third party that has no obligation of confidentiality.
- 20.4 The above provisions of GCC Clause 20 shall not in any way modify any undertaking of confidentiality given by either of the parties hereto prior to the date of the Contract in respect of the Supply or any part thereof.
- 20.5 The provisions of GCC Clause 20 shall survive completion or termination, for whatever reason, of the Contract.
- 21 Subcontracting 21.1 The Supplier shall notify the Purchaser in writing of all

subcontracts awarded under the Contract if not already specified in the bid. Such notification, in the original bid or later on shall not relieve the Supplier from any of its obligations, duties, responsibilities, or liability under the Contract.

- 21.2 Subcontracts shall comply with the provisions of GCC Clauses 3 and 7.
- 22.1 Technical Specifications and Drawings
 - (a) The Goods and Related Services supplied under this Contract shall conform to the technical specifications and standards mentioned in Section-VI, Schedule of Requirements and, when no applicable standard is mentioned, the standard shall be equivalent or superior to the official standards whose application is appropriate to the Goods' country of origin or India.
 - (b) Wherever references are made in the Contract to codes and standards in accordance with which it shall be executed, the edition or the revised version of such codes and standards shall be those specified in the Schedule of Requirements. During Contract execution, any changes in any such codes and standards shall be applied only after approval by the Purchaser and shall be treated in accordance with GCC Clause 33.
- 23 Packing and Documents
 23.1 The Supplier shall provide such packing of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the Contract. During transit, the packing shall be sufficient to withstand, without limitation, rough handling and exposure to extreme temperatures, salt and precipitation, and open storage. Packing case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.
 - 23.2 The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified in the SCC, and in any other instructions ordered by the Purchaser.
- 24.1 Unless otherwise specified in the SCC, the Goods supplied under the Contract shall be fully insured—in a freely convertible currency from an eligible country—against loss or damage incidental to manufacture or acquisition, transportation, storage, and delivery, in accordance with the applicable Incoterms or in the manner specified in the SCC.

22 Specifications and Standards

- 25 Transportation and Incidental Services
 25.1 The Supplier is required under the Contract to transport the Goods to a specified place of final destination within India, defined as the Project Site, transport to such place of destination in India, including insurance and storage, as shall be specified in the Contract, shall be arranged by the Supplier, and related costs shall be included in the Contract Price"; or any other agreed upon trade terms (specify the respective responsibilities of the Purchaser and the Supplier)
 - 25.2 The Supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
 - (a) performance or supervision of on-site assembly and/or start up of the supplied Goods;
 - (b) furnishing of tools required for assembly and/or maintenance of the supplied Goods;
 - (c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods;
 - (d) performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
 - (e) training of the Purchaser's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
 - 25.3 Prices charged by the Supplier for incidental services, if not included in the Contract Price for the Goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services otherwise shall be at the cost of suppliers.
- 26 Inspections and 26.1 Tests
- **ns and** 26.1 The Supplier shall at its own expense and at no cost to the Purchaser carry out all such tests and/or inspections of the Goods and Related Services as are specified in the SCC.
 - 26.2 The inspections and tests may be conducted on the premises of the Supplier or its Subcontractor, at point of delivery, and/or at the Goods' final destination, or in another place in India as specified in the SCC. Subject to GCC Sub-Clause 26.3, if conducted on the premises of the Supplier or its Subcontractor, all reasonable facilities and assistance, including access to drawings and production data, shall be furnished to the inspectors at no charge to the Purchaser.
 - 26.3 The Purchaser or its designated representative shall be entitled to attend the tests and/or inspections referred to in GCC Sub-Clause 26.2, provided that the Purchaser bear all of its own costs and expenses incurred in connection with such

attendance including, but not limited to, all traveling and board and lodging expenses.

- 26.4 Whenever the Supplier is ready to carry out any such test and inspection, it shall give a reasonable advance notice, including the place and time, to the Purchaser. The Supplier shall obtain from any relevant third party or manufacturer any necessary permission or consent to enable the Purchaser or its designated representative to attend the test and/or inspection.
- 26.5 The Purchaser may require the Supplier to carry out any test and/or inspection not required by the Contract but deemed necessary to verify that the characteristics and performance of the Goods comply with the technical specifications codes and standards under the Contract, provided that the Supplier's reasonable costs and expenses incurred in the carrying out of such test and/or inspection shall be added to the Contract Price. Further, if such test and/or inspection impedes the progress of manufacturing and/or the Supplier's performance of its other obligations under the Contract, due allowance will be made in respect of the Delivery Dates and Completion Dates and the other obligations so affected.
- 26.6 The Supplier shall provide the Purchaser with a report of the results of any such test and/or inspection.
- 26.7 The Purchaser may reject any Goods or any part thereof that fail to pass any test and/or inspection or do not conform to the specifications. The Supplier, if permitted by the purchaser, shall either rectify or replace such rejected Goods or parts thereof or make alterations necessary to meet the specifications at no cost to the Purchaser, and shall repeat the test and/or inspection, at no cost to the Purchaser, upon giving a notice pursuant to GCC Sub-Clause 26.4.
- 26.8 The Supplier agrees that neither the execution of a test and/or inspection of the Goods or any part thereof, nor the attendance by the Purchaser or its representative, nor the issue of any report pursuant to GCC Sub-Clause 26.6, shall absolve the Supplier from any warranties or other obligations under the Contract.
- 27 Liquidated Damages
 27.1 Except as provided under GCC Clause 32, if the Supplier fails to deliver any or all of the Goods by the Date(s) of delivery or perform the Related Services within the period specified in the Contract, the Purchaser may without prejudice to all its other remedies under the Contract, deduct from the Contract Price, as liquidated damages, a sum equivalent to the percentage specified in the SCC of the delivered price of the delayed Goods or unperformed Services for each week or part thereof of delay until actual delivery or performance, up to a maximum deduction of the percentage specified in those SCC. Once the maximum is reached, the Purchaser may terminate the Contract pursuant
 - 63

- **28 Warranty** 28.1 The Supplier warrants that all the Goods are new, unused, and of the most recent or current models, and that they incorporate all recent improvements in design and materials, unless provided otherwise in the Contract.
 - 28.2 Subject to Sub-Clause 22.1(b) of GCC, the Supplier further warrants that the Goods shall be free from defects arising from any act or omission of the Supplier or arising from design, materials, and workmanship, under normal use in the conditions prevailing in India.
 - 28.3 Unless otherwise specified in the SCC, the warranty shall remain valid for twelve (12) months after the Goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the SCC, or warranty period mentioned by supplier whichever period concludes later unless mutually agreed.
 - 28.4 The Purchaser shall give notice to the Supplier stating the nature of any such defects together with all available evidence thereof, promptly following the discovery thereof. The Purchaser shall afford all reasonable opportunity for the Supplier to inspect such defects.
 - 28.5 Upon receipt of such notice, the Supplier shall, within the period specified in the SCC, expeditiously repair or replace the defective Goods or parts thereof, at no cost to the Purchaser.
 - 28.6 If having been notified, the Supplier fails to remedy the defect within the period specified in the SCC; the Purchaser may proceed to take within a reasonable period such remedial action as may be necessary, at the Supplier's risk and expense and without prejudice to any other rights which the Purchaser may have against the Supplier under the Contract.
 - 29.1 The Supplier shall, subject to the Purchaser's compliance with GCC Sub-Clause 29.2, indemnify and hold harmless the Purchaser and its employees and officers from and against any and all suits, actions or administrative proceedings, claims, demands, losses, damages, costs, and expenses of any nature, including attorney's fees and expenses, which the Purchaser may suffer as a result of any infringement or alleged infringement of any patent, utility model, registered design, trademark, copyright, or other intellectual property right registered.
 - 29.2 If any proceedings are brought or any claim is made against the Purchaser arising out of the matters referred to in GCC Sub-Clause 29.1, the Purchaser shall promptly give the Supplier a notice thereof, and the Supplier may at its own expense and in the Purchaser's name conduct such

28.4 The Purchaser shal

29 Patent

Indemnity

proceedings or claim and any negotiations for the settlement of any such proceedings or claim.

- 29.3 The Purchaser shall, at the Supplier's request, afford all available assistance to the Supplier in conducting such proceedings or claim, and shall be reimbursed by the Supplier for all reasonable expenses incurred in so doing.
- **30 Force Majeure** 30.1 For purposes of this Clause, "Force Majeure" means an event or situation beyond the control of the Supplier that is not foreseeable, is unavoidable, and its origin is not due to negligence or lack of care on the part of the Supplier. Such events may include, but not be limited to, wars or revolutions, fires, floods, epidemics, quarantine restrictions, and freight embargoes.
 - 30.2 If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably possible, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.
 - 30.3 The Supplier shall not be liable for forfeiture of its Performance Security, liquidated damages, or termination for default if and to the extent that it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.
- 31 Change Orders and Contract Amendments
- **s** 31.1 The Purchaser may at any time order the Supplier through notice in accordance GCC Clause 8, to make changes within the general scope of the Contract in any one or more of the following:
 - (a) drawings, designs, or specifications, where Goods to be furnished under the Contract are to be specifically manufactured for the Purchaser;
 - (b) the method of shipment or packing;
 - (c) the place of delivery; and
 - (d) the Related Services to be provided by the Supplier.
 - 31.2 If any such change causes an increase or decrease in the cost of, or the time required for, the Supplier's performance of any provisions under the Contract, an equitable adjustment shall be made in the Contract Price or in the Delivery/Completion Schedule, or both, and the Contract shall accordingly be amended. Any claims by the Supplier for adjustment under this Clause must be asserted within twenty-eight (28) days from the date of the Supplier's receipt of the Purchaser's change order.
 - 31.3 Prices to be charged by the Supplier for any Related Services

that might be needed but which were not included in the Contract shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

- 31.4 Subject to the above, no variation in or modification of the terms of the Contract shall be made except by written amendment signed by the parties.
- 32 Extensions of Time
 32.1 If at any time during performance of the Contract, the Supplier or its subcontractors should encounter conditions impeding timely delivery of the Goods or completion of Related Services pursuant to GCC Clause 13, the Supplier shall promptly notify the Purchaser in writing of the delay, its likely duration, and its cause. As soon as practicable after receipt of the Supplier's notice, the Purchaser shall evaluate the situation and may at its discretion extend the Supplier's time for performance, in which case the extension shall be ratified by the parties by amendment of the Contract.
 - 32.2 Except in case of Force Majeure, as provided under GCC Clause 32, a delay by the Supplier in the performance of its Delivery and Completion obligations shall render the Supplier liable to the imposition of liquidated damages pursuant to GCC Clause 26, unless an extension of time is agreed upon, pursuant to GCC Sub-Clause 34.1.

33 Termination 33.1 Termination

- 33.1 Termination for Default
 - (a) The Purchaser, without prejudice to any other remedy for breach of Contract, by written notice of default sent to the Supplier, may terminate the Contract in whole or in part:
 - (i) if the Supplier fails to deliver any or all of the Goods within the period specified in the Contract, or within any extension thereof granted by the Purchaser pursuant to GCC Clause 32;
 - (ii) if the Supplier fails to perform any other obligation under the Contract; or
 - (iii) if the Supplier, in the judgment of the Purchaser has engaged in fraud and corruption, as defined in GCC Clause 3, in competing for or in executing the Contract.
 - (b) In the event the Purchaser terminates the Contract in whole or in part, pursuant to GCC Clause 33.1(a), the Purchaser may procure, upon such terms and in such manner as it deems appropriate, Goods or Related Services similar to those undelivered or not performed, and the Supplier shall be liable to the Purchaser for any additional costs for such similar Goods or Related Services. However, the Supplier shall continue

performance of the Contract to the extent not terminated.

- 33.2 Termination for Insolvency.
 - (a) The Purchaser may at any time terminate the Contract by giving notice to the Supplier if the Supplier becomes bankrupt or otherwise insolvent. In such event, termination will be without compensation to the Supplier, provided that such termination will not prejudice or affect any right of action or remedy that has accrued or will accrue thereafter to the Purchaser.
- 33.3 Termination for Convenience.
 - (a) The Purchaser, by notice sent to the Supplier, may terminate the Contract, in whole or in part, at any time for its convenience. The notice of termination shall specify that termination is for the Purchaser's convenience, the extent to which performance of the Supplier under the Contract is terminated, and the date upon which such termination becomes effective.
 - (b) The Goods that are complete and ready for shipment within twenty-eight (28) days after the Supplier's receipt of notice of termination shall be accepted by the Purchaser at the Contract terms and prices. For the remaining Goods, the Purchaser may elect:
 - (i) to have any portion completed and delivered at the Contract terms and prices; and/or
 - (ii) to cancel the remainder and pay to the Supplier an agreed amount for partially completed Goods and Related Services and for materials and parts previously procured by the Supplier.

- 34 Assignment
- 34.1 Neither the Purchaser nor the Supplier shall assign, in whole or in part, their obligations under this Contract, except with prior written consent of the other party.

Section VIII. Special Conditions of Contract

The following Special Conditions of Contract (SCC) shall supplement and / or amend the General Conditions of Contract (GCC). Whenever there is a conflict, the provisions herein shall prevail over those in the GCC.

GCC Clause Ref. No.	SCC	
GCC 1.1 (i)	The Purchaser's country is: India.	
GCC 1.1 (k)	The Purchaser is: Medical Superintendent, Sir Sunderlal Hospital, Institute of Medical Science, BHU, Varanasi -221005	
GCC 1.1 (j)	The Project Site(s)/Final Destination(s) is/are: Sir Sunderlal Hospital, Institute of Medical Science, BHU, Varanasi -221005	
GCC 5.1	The language shall be: Hindi or English	
GCC 8.1	For notices, the Purchaser's address shall be:	
	Attention : Prof. Kailash Kumar	
	Street Address : Medical Superintendent,	
	Sir Sunderlal Hospital	
	Institute of Medical Sciences, BHU, Varanasi	
	Floor/ Room number : Medical Superintendent Office	
	City : Varanasi	
	ZIP Code : 221005	
	Country : India	
	Telephone : 91-542-2369033	
	91-542-2309200	
	Electronic mail address : <u>ms.ssh.bhu@gmail.com</u>	
GCC 9.1	The governing law shall be the law of India .	
GCC 10.2	The rules of procedure for arbitration proceedings pursuant to GCC Clause 10.2 shall be as follows:	
	Contracts with Supplier for arbitral proceeding	
	In the case of a dispute between the Purchaser and a Supplier, the dispute shall be referred to adjudication or arbitration in accordance with the laws of India by the arbitrator appointed by vice-chancellor, BHU, unless otherwise agreed.	
GCC 13.1	Details of Shipping and other Documents to be furnished by the Supplier are:	
	(i) An airway bill	
	(ii) Insurance Certificate,	
	(iii) Manufacturer's or Supplier's Warranty Certificate,	

	
	(iv) Inspection Certificate issued by nominated inspection agency,(v) Supplier's factory shipping details etc.
	The above documents shall be received by the Purchaser before arrival of the Goods and, if not received, the Supplier will be responsible for any consequent expenses.
GCC 15.1	The prices charged for the Goods supplied and the related Services performed shall not be adjustable.
GCC 16.1	GCC 16.1—The method and conditions of payment to be made to the Supplier under this Contract shall be as follows:
	Payment for Goods supplied from abroad:
	Payment of foreign currency portion shall be made in Indian Rupees in the following manner:
	 (i) On Shipment: Ninety (90%) percent of the Contract Price of the Goods shipped shall be paid through irrevocable confirmed letter of credit opened in favor of the Supplier in a bank in its country, upon receiving equipment in good condition and documents specified in GCC Clause 13.1 under Section-VII.
	(ii) On Acceptance: Remaining (10%) percent of the Contract Price of Equipment received shall be paid within thirty (30) days of receipt of the equipment upon submission of claim supported by the acceptance certificate issued by the Purchaser by bank draft/wire transfer.
	Payment of local currency portion shall be made in Indian Rupees within thirty (30) days of presentation of claim supported by a certificate from the Purchaser declaring that the Goods have been delivered and that all other contracted Services have been performed.
	Payment for Goods and Services supplied from within India:
	Payment for Goods and Services supplied from within India shall be made in Indian Rupees , as follows:
	 (i) On Delivery & Acceptance: Hundred (100%) percent of the Contract Price shall be paid on receipt of the Goods in good conditions and acceptance certificate for satisfactory installation and functioning.
GCC 18.1	A Performance Security shall be required @ 5% of contract price.
GCC 18.3	If required, the Performance Security shall be in the form of : Bank Guarantee
	If required, the Performance security shall be denominated in Indian Rupees.
GCC 23.2	The packing, marking and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract.

GCC 24.1	The insurance coverage shall be as specified in the Incoterms .
	If not in accordance with Incoterms , insurance shall be as follows: NA
GCC 25.1	Responsibility for transportation of the Goods shall be as specified in the Incoterms .
	If not in accordance with Incoterms , responsibility for transportations shall be as follows: NA
GCC 25.2	Incidental services to be provided are:
	Selected services covered under GCC Clause 25.2 and/or other should be specified with the desired features. The price quoted in the bid price or agreed with the selected Supplier shall be included in the Contract Price.
GCC 26.1	The inspections and tests shall be: Equipment should be installed at site by designated engineer of the firm. Demonstration should be made to the satisfaction of the P.I. At least four free visits should be made by the engineer during the first year after installation to clarify and rectify any doubts or problems as may be faced by the user.
GCC 26.2	The Inspections and tests shall be conducted at the Department where the equipment is installed. For the rest please refer to GCC 26.1.
GCC 27.1	The liquidated damage shall be: 1% per week
GCC 27.1	The maximum amount of liquidated damages shall be: 10%
GCC 28.3	The period of validity of the Warranty shall be 12 months from date of acceptance/ satisfactory installation of the equipment.
	For purposes of the Warranty, the place(s) of final destination(s) shall be the Department where the equipment is installed.
	The Supplier shall, in addition, comply with the performance and/or consumption guarantees specified under the Contract (if any). If, for reasons attributable to the Supplier, these guarantees are not attained in whole or in part, the Supplier shall, at its discretion, either:
	 (a) make such changes, modifications, and/or additions to the Goods or any part thereof as may be necessary in order to attain the contractual guarantees specified in the Contract at its own cost and expense and to carry out further performance tests in accordance with SCC 4, or
	(b) pay liquidated damages to the Purchaser with respect to the failure to meet the contractual guarantees. The rate of these liquidated

	damages shall be 1% per week of actual value of the equipment (maximum 10%).
GCC 28.5	The period for repair or replacement shall be: 60 days

Part-4 Bidding Forms & Contract Forms

Section IX: Bidding Forms

Table of Forms

1.	Tender Form (Techno commercial un-priced Bid)	74
2.	Tender Form (Priced Bid)	77
3.	Bidder Information Form	79
4.	Manufacturer's Authorization	80

(i) Tender Form

(Techno commercial un-priced Bid)

(On the letter head of the firm submitting the bid)

Tender No
То
The

Dear Sir,

- 1. I/We have examined and have no reservations to the Bidding Documents, including Addenda issued in accordance with Instructions to Bidders;
- 2. I/We meet the eligibility requirements and have no conflict of interest;
- 3. I/We have not been suspended nor declared ineligible in India;
- I/We offer to supply in conformity with the Bidding Documents and in accordance with the Delivery Schedules specified in the Schedule of Requirements the following Goods: [insert a brief description of the Goods and Related Services];
- 5. I/We offer to supply the items as listed in the schedule to this tender hereto/portion thereof as you may specify in the acceptance of Tender at the price given in the said Schedule and agree to hold this offer open for a period of 90 days from the date of opening of the tender.
- 6. I/we shall be bound by a communication of acceptance issued by you.
- 7. I/We have understood the Instruction to bidders and Conditions of Contract in the form as enclosed with the invitation to the tender and have thoroughly examined the specifications quoted in the Schedule hereto and am/are fully aware of the nature of the goods required and my/our offer is to supply the goods strictly in accordance with the specifications and requirements.
- 8. A crossed Bank Draft in favor of the Registrar, BHU for Rs. (Rupees.....only) as Earnest Money is enclosed. The Draft is drawn onBank payable at Varanasi.

- 9. The following have been added to form part of this tender.
 - (a) Samples of items quoted for, as per instructions provided in the schedule of requirement.
 - (b) Schedule of requirements, quoting the make only duly signed and stamped.(without indicating price)
 - (c) Income Tax clearance certificate.
 - (d) Copy of last audited balance sheet.
 - (e) Copy of Valid Central/State sales tax registration certificate.
 - (f) Copy of relevant major purchase orders valuing mare than Rs. 20000/- executed during last two years for Govt. Depts., PSUs & Central Autonomous bodies..
 - (g) Proof of manufacturing Unit, dealership certificate/general order suppliers.
 - (h) Statement of deviations from financial terms & conditions, if any.
 - (i) Any other enclosure. (Please give details)
- 10. We undertake to execute all orders which have been placed to meet emergent requirements on priority basis.
- 11. Certified that the bidder is:
 - (a) A sole proprietorship firm and the person signing the bid document is the sole proprietor/constituted attorney of the sole proprietor,

Or

(b) A partnership firm, and the person signing the bid document is a partner of the firm and he has authority to refer to arbitration disputes concerning the business of the partnership by virtue of the partnership agreement/by virtue of general power of attorney.

Or

(c) A company and the person signing the document is the constituted attorney.

(NOTE: Delete whatever is not applicable. All corrections/deletions should invariable be duly attested by the person authorized to sign the bid document).

- 12. We do hereby undertake that, until a formal notification of award, this bid, together with your written acceptance thereof shall constitute a binding contract between us.
- 13. If our bid is accepted, we commit to obtain a performance security in accordance with the Bidding Documents;
- 14. We are not participating, as a Bidder or as a subcontractor, in more than one bid in this bidding process, other than alternative bids submitted;
- 15. We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in any type of fraud and corruption

Name of the Bidder* [insert complete name of person signing the Bid]

Name of the person duly authorized to sign the Bid on behalf of the Bidder** [insert complete name of person duly authorized to sign the Bid]

Title of the person signing the Bid [insert complete title of the person signing the Bid]

Signature of the person named above <u>[insert signature of person whose name and</u> capacity are shown above]

Date signed _ [insert date of signing] day of [insert month], [insert year]

*: In the case of the Bid submitted by joint venture specify the name of the Joint Venture as Bidder

**: Person signing the Bid shall have the power of attorney given by the Bidder to be attached with the Bid Schedules.

Yours faithfully,

(Signature of bidder)

Dated this day of _____

Address:

.....

.....

Telephone No. :_____

FAX_____

E-mail

Company seal

(ii) Tender Form

(Priced Bid)

(On the letter head of the firm submitting the bid document)

To,

 The ______

 Ref: Tender No......Dated:

Sir,

Having examined the bidding documents and having submitted the techno commercial unpriced bid for the same, we, the undersigned, hereby submit the priced bid for supply of goods and services as per the schedule of requirements and in conformity with the said bidding documents.

- 1. We hereby offer to supply the Goods/Services at the prices and rates mentioned in the enclosed schedule of requirement.
- 2. We do hereby undertake that, in the event of acceptance of our bid, the supply of Goods/Services shall be made as stipulated in the schedule of requirement and that we shall perform all the incidental services.
- 3. The prices quoted are inclusive of all charges net F.O.R University. We enclose herewith the complete Financial Bid as required by you. This includes:
 - a. Price Schedule (Bill of Quantity-BOQ).
 - b. Statement of deviations from financial terms and conditions.
- 4. We agree to abide by our offer for a period of 90 days from the date fixed for opening of the bid documents and that we shall remain bound by a communication of acceptance within that time.
- 5. We have carefully read and understood the terms and conditions of the bid document and we do hereby undertake to supply as per these terms and conditions. The Financial Deviations are only those mentioned in the statement of deviations from financial terms and conditions.
- 6. We have paid, or will pay the following commissions, gratuities, or fees with respect to the bidding process or execution of the Contract: [insert complete name of each Recipient, its full address, the reason for which each commission or gratuity was paid and the amount and currency of each such commission or gratuity]

Name of Recipient	Address	Reason	Amount

(If none has been paid or is to be paid, indicate "none.")

- 7. We understand that this bid, together with your written acceptance thereof included in your notification of award, shall constitute a binding contract between us, until a formal contract is prepared and executed; and
- 8. We understand that you are not bound to accept the lowest evaluated bid or any other bid that you may receive.

Certified that the bidder is:

A sole proprietorship firm and the person signing the bid document is the sole proprietor/ constituted attorney of sole proprietor,

Or

A partnership firm, and the person signing the bid document is a partner of the firm and he has authority to refer to arbitration disputes concerning the business of the partnership by virtue of the partnership agreement/by virtue of general power of attorney,

Or

A company and the person signing the bid document is the constituted attorney.

(NOTE: Delete whatever is not applicable. All corrections/deletions should invariably be duly attested by the person authorized to sign the bid document.)

We do hereby undertake that, until a formal notification of award, this bid, together with your written acceptance thereof, shall constitute a binding contract between us.

Dated this day of
Signature of Bidder
Details of enclosures
Full Address:
Telephone No
Mobile No. :
Fax No. :
E-mail:

Company Seal

Bidder Information Form

[The Bidder shall fill in this Form in accordance with the instructions indicated below. No alterations to its format shall be permitted and no substitutions shall be accepted.]

Date: [insert date (as day, month and year) of Bid Submission] ADVT. No.: [insert number of bidding process] Alternative No.: [insert identification No if this is a Bid for an alternative] Page ______ of _____ pages

1. Bidder's Name [insert Bidder's legal name]

2. In case of JV, legal name of each member : [insert legal name of each member in JV]

3. Bidder's actual or intended country of registration: [insert actual or intended country of registration]

4. Bidder's year of registration: [insert Bidder's year of registration]

5. Bidder's Address in country of registration: [insert Bidder's legal address in country of registration]

6. Bidder's Authorized Representative Information

Name: [insert Authorized Representative's name]

Address: [insert Authorized Representative's Address]

Telephone/Fax numbers: [insert Authorized Representative's telephone/fax numbers]

Email Address: [insert Authorized Representative's email address]

7. Attached are copies of original documents of [check the box(es) of the attached original documents]

- □ Articles of Incorporation (or equivalent documents of constitution or association), and/or documents of registration of the legal entity named above.
- $\hfill\square$ In case of JV, letter of intent to form JV or JV agreement.
- □ In case of Government-owned enterprise or institution, documents establishing:
 - Legal and financial autonomy
 - Operation under commercial law
 - Establishing that the Bidder is not dependent agency of the Purchaser
- 2. Included are the organizational chart, a list of Board of Directors, and the beneficial ownership.

Manufacturer's Authorization

[The Bidder shall require the Manufacturer to fill in this Form in accordance with the instructions indicated. This letter of authorization should be on the letterhead of the Manufacturer and should be signed by a person with the proper authority to sign documents that are binding on the Manufacturer. The Bidder shall include it in its bid, if so indicated in the **BDS**.]

Date: [insert date (as day, month and year) of Bid Submission] ADVT. No.: [insert number of bidding process] Alternative No.: [insert identification No if this is a Bid for an alternative]

To: [insert complete name of Purchaser]

WHEREAS

We [insert complete name of Manufacturer], who are official manufacturers of [insert type of goods manufactured], having factories at [insert full address of Manufacturer's factories], do hereby authorize [insert complete name of Bidder] to submit a bid the purpose of which is to provide the following Goods, manufactured by us [insert name and or brief description of the Goods], and to subsequently negotiate and sign the Contract.

We hereby extend our full guarantee and warranty in accordance with Clause 28 of the General Conditions of Contract, with respect to the Goods offered by the above firm.

Signed: [insert signature(s) of authorized representative(s) of the Manufacturer]

Name: [insert complete name(s) of authorized representative(s) of the Manufacturer]

Title: [insert title]

Dated on ______ day of ______, ____ [insert date of signing]

Section X. Contract Forms

This Section contains forms which, once completed, will form part of the Contract. The forms for Performance Security and Advance Payment Security, when required, shall only be completed by the successful Bidder after contract award.

Table of Forms

1.	Tender Acceptance Letter	. 82
2.	Contract Agreement	. 83
3.	Performance Security	. 85
4.	Advance Payment Security	. 89

Tender Acceptance Letter

(To be given on Company Letter Head)

Date:

Sub: Acceptance of Terms & Conditions of Tender.

Tender Reference No._____

Name of Tender/ Work: -

Dear Sir,

1. I/We have downloaded/ obtained the tender document(s) for the above mentioned 'Tender/Work' from the web site(s) namely:

as per your advertisement, given in the above mentioned website(s).

- 2. I/We hereby certify that I/We have read the entire terms and conditions of the tender documents from Page No. ______ to _____ (including all documents like section(s), schedules(s) etc.), which form part of the contract agreement and I/we shall abide hereby by the terms/conditions/ clauses contained therein.
- 3. The corrigendum(s) issued from time to time by your department/ organisation too have also been taken into consideration, while submitting this acceptance letter.
- 4. I/We hereby unconditionally accept the tender conditions of above mentioned tender document(s)/ corrigendum(s) in its totality/entirety.
- 5. In case any provisions of this tender are found violated, then your department/organisation shall without prejudice to any other right or remedy be at liberty to reject this tender/bid including the forfeiture of the full said earnest money deposit absolutely.

Yours Faithfully,

(Signature of the Bidder, with Official Seal)

Contract Agreement

[The successful Bidder shall fill in this form in accordance with the instructions indicated]

THIS AGREEMENT made

the [insert: number] day of [insert: month], [insert: year].

BETWEEN

- [insert complete name of Purchaser], a [insert description of type of legal entity, for example, an agency of the Ministry of of the Government of { insert name of Country of Purchaser }, or corporation incorporated under the laws of { insert name of Country of Purchaser }] and having its principal place of business at [insert address of Purchaser] (hereinafter called "the Purchaser"), of the one part, and
- (2) [insert name of Supplier], a corporation incorporated under the laws of [insert: country of Supplier] and having its principal place of business at [insert: address of Supplier] (hereinafter called "the Supplier"), of the other part :

WHEREAS the Purchaser invited bids for certain Goods and ancillary services, viz., *[insert brief description of Goods and Services]* and has accepted a Bid by the Supplier for the supply of those Goods and Services

The Purchaser and the Supplier agree as follows:

- 1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Contract documents referred to.
- 2. The following documents shall be deemed to form and be read and construed as part of this Agreement. This Agreement shall prevail over all other contract documents.
 - (a) the Letter of Acceptance
 - (b) the Tender Forms
 - (c) the Addenda Nos.____ (if any)
 - (d) Special Conditions of Contract
 - (e) General Conditions of Contract
 - (f) the Specification (including Schedule of Requirements and Technical Specifications)
 - (g) the completed Schedules (including Price Schedules(BOQ))
 - (h) any other document listed in GCC as forming part of the Contract

- 3. In consideration of the payments to be made by the Purchaser to the Supplier as specified in this Agreement, the Supplier hereby covenants with the Purchaser to provide the Goods and Services and to remedy defects therein in conformity in all respects with the provisions of the Contract.
- 4. The Purchaser hereby covenants to pay the Supplier in consideration of the provision of the Goods and Services and the remedying of defects therein, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

IN WITNESS whereof the parties hereto have caused this Agreement to be executed in accordance with the laws of *[insert the name of the Contract governing law country]* on the day, month and year indicated above.

For and on behalf of the Purchaser

Signed: *[insert signature]* in the capacity of *[insert title or other appropriate designation]* in the presence of *[insert identification of official witness]*

For and on behalf of the Supplier

Signed: *[insert signature of authorized representative(s) of the Supplier]* in the capacity of *[insert title or other appropriate designation]* in the presence of *[insert identification of official witness]*

Performance Security

Option 1: (Bank Guarantee)

[The bank, as requested by the successful Bidder, shall fill in this form in accordance with the instructions indicated]

[Guarantor letterhead or SWIFT identifier code]

Beneficiary: [insert name and Address of Purchaser]

Date: *[Insert date of issue]*

PERFORMANCE GUARANTEE No.: [Insert guarantee reference number]

Guarantor: [Insert name and address of place of issue, unless indicated in the letterhead]

We have been informed that _ [insert name of Supplier, which in the case of a joint venture shall be the name of the joint venture] (hereinafter called "the Applicant") has entered into Contract No. [insert reference number of the contract] dated [insert date] with the Beneficiary, for the supply of _ [insert name of contract and brief description of Goods and related Services] (hereinafter called "the Contract").

Furthermore, we understand that, according to the conditions of the Contract, a performance guarantee is required.

At the request of the Applicant, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of [insert amount in figures] (______) [insert amount in words]⁵, such sum being payable in the types and proportions of currencies in which the Contract Price is payable, upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating that the Applicant is in breach of its obligation(s) under the Contract, without the Beneficiary needing to prove or to show grounds for your demand or the sum specified therein.

This guarantee shall expire, no later than the Day of, $2...^6$, and any demand for payment under it must be received by us at this office indicated above on or before that date.

⁵ The Guarantor shall insert an amount representing the percentage of the Accepted Contract Amount specified in the Letter of Acceptance, and denominated either in the currency(ies) of the Contract or a freely convertible currency acceptable to the Beneficiary.

⁶ Insert the date twenty-eight days after the expected completion date as described in GC Clause 18.4. The Purchaser should note that in the event of an extension of this date for completion of the Contract, the Purchaser would need to request an extension of this guarantee from the Guarantor. Such request must be in writing and must be made prior to the expiration date established in the guarantee. In preparing this guarantee, the Purchaser might consider adding the following text to the form, at the end of the penultimate paragraph: "The Guarantor agrees to a one-time extension of this guarantee for a period not to exceed [six months][one year], in response to the Beneficiary's written request for such extension, such request to be presented to the Guarantor before the expiry of the guarantee."

This guarantee is subject to the Uniform Rules for Demand Guarantees (URDG) 2010 Revision, ICC Publication No. 758, except that the supporting statement under Article 15(a) is hereby excluded.

[signature(s)]

Note: All italicized text (including footnotes) is for use in preparing this form and shall be deleted from the final product.

Option 2: Performance Bond

By this Bond *[insert name of Principal]* as Principal (hereinafter called "the Supplier") and *[insert name of Surety]* as Surety (hereinafter called "the Surety"), are held and firmly bound unto *[insert name of Purchaser]* as Obligee (hereinafter called "the Supplier") in the amount of *[insert amount in words and figures]*, for the payment of which sum well and truly to be made in the types and proportions of currencies in which the Contract Price is payable, the Supplier and the Surety bind themselves, their heirs, executors, administrators, successors and assigns, jointly and severally, firmly by these presents.

WHEREAS the Contractor has entered into a written Agreement with the Purchaser dated the ______ day of ______, 20 _____, for [name of contract and brief description of Goods and related Services] in accordance with the documents, plans, specifications, and amendments thereto, which to the extent herein provided for, are by reference made part hereof and are hereinafter referred to as the Contract.

NOW, THEREFORE, the Condition of this Obligation is such that, if the Supplier shall promptly and faithfully perform the said Contract (including any amendments thereto), then this obligation shall be null and void; otherwise, it shall remain in full force and effect. Whenever the Supplier shall be, and declared by the Purchaser to be, in default under the Contract, the Purchaser having performed the Purchaser's obligations thereunder, the Surety may promptly remedy the default, or shall promptly:

- (1) complete the Contract in accordance with its terms and conditions; or
- (2) obtain a Bid or bids from qualified Bidders for submission to the Purchaser for completing the Contract in accordance with its terms and conditions, and upon determination by the Purchaser and the Surety of the lowest responsive Bidder, arrange for a Contract between such Bidder and Purchaser and make available as work progresses (even though there should be a default or a succession of defaults under the Contract or Contracts of completion arranged under this paragraph) sufficient funds to pay the cost of completion less the Balance of the Contract Price; but not exceeding, including other costs and damages for which the Surety may be liable hereunder, the amount set forth in the first paragraph hereof. The term "Balance of the Contract Price," as used in this paragraph, shall mean the total amount payable by Purchaser to Supplier under the Contract, less the amount properly paid by Purchaser to Contractor; or
- (3) pay the Purchaser the amount required by Purchaser to complete the Contract in accordance with its terms and conditions up to a total not exceeding the amount of this Bond.

The Surety shall not be liable for a greater sum than the specified penalty of this Bond.

Any suit under this Bond must be instituted before the expiration of one year from the date of the issuing of the Taking-Over Certificate.

No right of action shall accrue on this Bond to or for the use of any person or corporation other than the Purchaser named herein or the heirs, executors, administrators, successors, and assigns of the Purchaser.

In testimony whereof, the Supplier has hereunto set his hand and affixed his seal, and the Surety has caused these presents to be sealed with his corporate seal duly attested by the signature of his legal representative, this _____ day of _____ 20____.

SIGNED	ON		 	on	behalf	of
Ву			 in	the	capacity	of
In 		the	 presen	ce		of
SIGNED	ON		 	on	behalf	of
Ву			 in	the	capacity	of
In		the	presen	ce		of

Advance Payment Security

[Guarantor letterhead or SWIFT identifier code]

Beneficiary: [Insert name and Address of Purchaser]

Date: [Insert date of issue]

ADVANCE PAYMENT GUARANTEE No.: [Insert guarantee reference number]

Guarantor: [Insert name and address of place of issue, unless indicated in the letterhead]

We have been informed that *[insert name of Supplier, which in the case of a joint venture shall be the name of the joint venture]* (hereinafter called "the Applicant") has entered into Contract No. *[insert reference number of the contract]* dated *[insert date]* with the Beneficiary, for the execution of *[insert name of contract and brief description of Goods and related Services]* (hereinafter called "the Contract").

Furthermore, we understand that, according to the conditions of the Contract, an advance payment in the sum *[insert amount in figures]* () *[insert amount in words]* is to be made against an advance payment guarantee.

At the request of the Applicant, we as Guarantor, hereby irrevocably undertake to pay the Beneficiary any sum or sums not exceeding in total an amount of *[insert amount in figures]* (______) *[insert amount in words]*¹ upon receipt by us of the Beneficiary's complying demand supported by the Beneficiary's statement, whether in the demand itself or in a separate signed document accompanying or identifying the demand, stating either that the Applicant:

- (a) has used the advance payment for purposes other than toward delivery of Goods;
 - or
- (b) has failed to repay the advance payment in accordance with the Contract conditions, specifying the amount which the Applicant has failed to repay.

A demand under this guarantee may be presented as from the presentation to the Guarantor of a certificate from the Beneficiary's bank stating that the advance payment referred to

¹ The Guarantor shall insert an amount representing the amount of the advance payment and denominated either in the currency(ies) of the advance payment as specified in the Contract, or in a freely convertible currency acceptable to the Purchaser.

above has been credited to the Applicant on its account number [insert number] at [insert name and address of Applicant's bank].

The maximum amount of this guarantee shall be progressively reduced by the amount of the advance payment repaid by the Applicant as specified in copies of interim statements or payment certificates which shall be presented to us. This guarantee shall expire, at the latest, upon our receipt of a copy of the interim payment certificate indicating that ninety (90) percent of the Accepted Contract Amount, has been certified for payment, or on the *[insert day]* day of *[insert month]*, *[insert year]*, whichever is earlier. Consequently, any demand for payment under this guarantee must be received by us at this office on or before that date.

This guarantee is subject to the Uniform Rules for Demand Guarantees (URDG) 2010 Revision, ICC Publication No.758, except that the supporting statement under Article 15(a) is hereby excluded.

[signature(s)]

.

Note: All *italicized text (including footnotes)* is for use in preparing this form and shall be deleted from the final product.

INVITATION FOR BIDS Notice Inviting Tender (NIT)

BANARAS HINDU UNIVERSITY SIR SUNDERLAL HOSPITAL INSTITUTE OF MEDICAL SCIENCES VARANASI-221005

e-Procurement Notice

Ref: BHU/SSH/2021-2022/08

Online tenders are hereby invited in two cover system from reputed manufacturer/ authorized representative of a manufacture/whole sale dealer/importer for supply of :-Equipment for COVID-19 & Black Fungus Patients

Bidders can download complete set of bidding documents from e- procurement Platform <u>http://eprocure.gov.in/eprocure/app</u> from 18-06-2021 onwards. Bidders need to submit the bids online for the interested items by uploading all the required documents through <u>http://eprocure.gov.in/eprocure/app</u>.

Last Date/ Time for receipt of bids through e-procurement is: 07-07-2021 upto 05:00PM. (Server time). Late bids shall not be accepted.

For further details regarding Tender Notification & Specifications please visit website: <u>http://eprocure.gov.in/eprocure/app</u> and <u>www.bhu.ac.in</u>.

Published Date	18-06-2021 (04:00 PM)
Bid Document Download Start Date	18-06-2021 (04:00 PM)
Clarification Start Date	18-06-2021 (04:00 PM)
Clarification End Date	21-06-2021 (04:00 PM)
Pre bid meeting	NA
Bid Submission Start Date	22-06-2021 (04:00 PM)
Bid Document Download End Date	09-07-2021 (05:00 PM)
Bid Submission End Date	09-07-2021 (05:00 PM)
Bid Opening Date	12-07-2021 (03:00 PM)

CRITICAL DATE SHEET

sd/-Registrar

Dated: 15-06-2021