


**CORRIGENDUM**

**Date: 04.03.2022**

**Tender Ref. No.:** BHU/PED./EQUIP/2022/02  
**Tender ID:** 2022\_BHU\_673673829\_1  
**Tender Title:** Procurement of Medical Equipments.  
**Corrigendum Title:** Update in Technical Specifications.

Due to technical reasons, the specifications required in e-tender are revised and updated as mentioned below:

Specifications before Corrigendum	Specifications after Corrigendum
<p><b>Item 3: Multipara Monitor (7 para or more)</b> <b>Point 5:</b> Monitors must be able to monitor ECG (5/12 lead display with reduced lead sets), SpO2, NIBP, Respiration, dual temp, dual IBP &amp; ETCO2 with simultaneous monitoring of 2xIBP and 2x Temp for all monitors</p>	<p><b>Item 3: Multipara Monitor (7 para or more):</b> <b>Point 5:</b> Monitors must be able to monitor ECG (5/12 lead display with reduced lead sets), SpO2, NIBP, Respiration, dual temp, dual IBP and Channel EEG with amplitude EEG trend graph.</p>
<p><b>Point 6:</b> The monitor should be upgradeable to a) Multi Gas measurements including CO2, N2O, O2, and Anesthetic Agents including inspiratory &amp; expiratory values, b) BIS, c) NMT d) ScvO2 e) EEG f) Continuous Cardiac output using Non-Invasive and Minimal Invasive technologies (Price to be quoted as optional).</p>	<p><b>Point 6:</b> The monitor should be upgradeable to Modules like, SCVo2 a) ETCO2, (b)ScvO2 (c) Continuous Cardiac output using Non-Invasive and Minimal Invasive technologies (Price to be quoted as optional).</p>
<p><b>Point 9:</b></p> <ul style="list-style-type: none"> <li>• Adult SpO2 Probe</li> <li>• NIBP cuffs for Adult &amp; Pediatrics</li> </ul>	<p><b>Point 9:</b></p> <ul style="list-style-type: none"> <li>• Pediatric and Neonatal SpO2 Probe</li> <li>• NIBP cuffs for Adult, Pediatric &amp; Neonatal</li> </ul>

  
Tender Inviting Authority  
(TIA)  
Nodal Officer  
Scheme No (6036)  
Department of Pediatrics  
IMS, BHU

**Government e-Procurement System  
Corrigendum Input Form**

1	Tender Ref. No	BHU/PCD/EQUIP/2022/02
2	Tender Title	Procurement of Medical Equipment
3	Tender ID *	2022-BHU-673829-1
4	Corrigendum Title *	Update in Technical Specifications.

**5 Corrigendum Type \***

**5 a) Before Tender Opening**

**5(b) After Tender Opening**

- Date    Fees    BOQ     
 Technical Bid    Financial Bid  
 Bid Openers    Terms and Conditions  
 Cancellation of Tender  
 Technical Bid Opening Date Extension  
 Others

- Retender  
 Cancellation of Tender

**5 a1) If Corrigendum Type is Date, Enter the appropriate change in date**

DD / MM / YYYY    HH : MM

Document Sale Start Date	<input type="text"/>	<input type="text"/>
Document Sale End Date	<input type="text"/>	<input type="text"/>
Seek Clarification Start Date	<input type="text"/>	<input type="text"/>
Seek Clarification End Date	<input type="text"/>	<input type="text"/>
Pre Bid Meeting Date	<input type="text"/>	<input type="text"/>
Bid Submission Start Date	<input type="text"/>	<input type="text"/>
Bid Submission End Date	<input type="text"/>	<input type="text"/>
Bid Opening Date	<input type="text"/>	<input type="text"/>

**5 a2) If Corrigendum Type is Fee**

(1) Tender Fee	<input type="text"/>
Processing Fee	<input type="text"/>
Surcharges	<input type="text"/>
Other Charges	<input type="text"/>

**Government e-Procurement System  
Corrigendum Input Form**

	Tender Fee Payable To	
	Tender Fee Payable At	
	(2) EMD Fee	<input type="radio"/> Fixed <input type="radio"/> Percentage
	If EMD is Fixed, Enter Amount	
	If EMD is Percentage enter %	
	EMD Exemption Allowed	<input type="radio"/> Yes <input type="radio"/> No
	EMD Payable To	
	EMD Payable At	
5a3)	If Corrigendum Type is BOQ, Enter revised / New BOQ	
	File Name :	File Description :
	Type :	Size :
5a4)	If Corrigendum Type is Bid Openers, enter new openers	
	Name / Designation	Email ID
	BO1	<input type="checkbox"/>
	BO2	<input type="checkbox"/>
	BO3	<input type="checkbox"/>
5a5)	If Corrigendum Type is Tender Opening Date Postponement, etc	New Date & Time DD/MM/YYYY                      HH:MM <input type="text"/> <input type="text"/>
6	Corrigendum Document Should be Uploaded for all Corrigendum Types	
	File name	Description      Type      Size

Prepared by *Kavita Singh*  
Name/Designation *(Senior Assist.)*

Approved by *[Signature]*  
Name/Designation  
*Prof Sunil Kumar Rao*  
Nodal Officer  
Scheme No (6036)  
Department of Pediatrics  
IMS, BHU

Date of updation:

Seal of the Office of the TIA:  
*[Signature]*  
*Prof Sunil Kumar Rao*  
Nodal Officer  
Scheme No (6036)  
Department of Pediatrics  
IMS, BHU