Corrigendum in technical Specifications of Multi-mode Micro Plate reader

Tender ID	2022_BHU_686968_1
Tender Title	
Corrigendum Title	Supply of Multi-mode Microplate reader
	Corrigendum in Did D
Tender Reference	Corrigendum in Bid Documents (in Technical Specification)
	BHII/MMV/700/02-
	BHU/MMV/ZOO/GKM/EQP/2022-23/01

Item No.	
Item	01 Multi-mode Microplate reader
Previous Specification	
1 - Microplate types: 6, 12, 24	Revised Specification
plates, PCR plates, Quartz Cuvette (Biocell) and compatible with Multi-Volume Plate with 2 µL Microspots 2- Reading speed:	1- Microplate types: 6, 12, 24, 48, 96 and 384-well plates, PCR plates, 1 ml Quartz vessel / Cuvette with vertical reading capability and compatible with Multi-Volume Plate with 2 μL Microspots.
96: 11 seconds 384: 22 seconds	2- Reading speed: 96 Wells: 10 - 15 seconds 384 Wells: 20 - 25 seconds

Land for my